

MINUTES OF TECHNICAL ADVISORY COMMITTEE (TAC) MEETING REGARDING THE PURCHASE OF ELECTRO MEDICAL EQUIPMENT / MISC. / I. T. & ELECTRIC ITEMS FOR 2nd UNIT OF NEONATOLOGY FOR THE YEAR 2021-2022 MEETING HELD ON 26-10-2021 AT 11:00 A.M. IN THE CONFERENCE ROOM OF THE ADMIN BLOCK THE CHILDREN'S HOSPITAL LAHORE.

In compliance to letter No. P.C./50071-82/CH&ICH dated 22-10-2021, Technical Advisory Committee (TAC) meeting held on 26-10-2021 at 11:00 a.m. in conference room of the admin block of The Children's Hospital & Institute of Child Health, Lahore.

Meeting started with the name of Allah by recitation of the Holy Quran. Bio medical engineers/technical officers briefed the committee that the equipment to be purchased for the 2nd unit of Neonatology advertised through press as well as on PPRA website as per approved PC-I. All the committee members, end-users and co-opted members scrutinized / evaluated and discussed all the quoted bids in detail and give the following recommendations for finalization of the technical advisory committee (TAC) reports. Following decision were made unanimously by all the members of TAC committee, co-opted members & end users.

Tender bid Sr. No.	Item Description	Remarks by TAC Committee
1.	(A) Non-invasive Cardiac Monitor (B) Invasive Cardiac Monitor (As a package)	<p>Eight firms participated in the bid & four firms out of eight were declared responsive</p> <p>Detail of the responsive firms are as under:-</p> <p>M/s Bio Tech Services, M/s Radiant Medical (Pvt) Ltd., M/s Shirazi Trading Company (Pvt) Ltd & M/s Medical Equipment & Systems, as they fulfill the criteria of bid evaluation hence declared responsive.</p> <p>Details of Non Responsive Firms are as under:</p> <p>M/s Vital Care & M/s Delta Plus (Pvt) Ltd. quoted same model of the same manufacturer as per bidding criteria firm should be exclusive authorization or sole distributor, hence both firms were declared non-responsive.</p> <p>M/s Vertex Medical (Pvt) Ltd: Quoted model is not compliance to technical specifications.</p> <p>M/s Human Health Care did not attached original bid security (CDR / Bank Guarantee) with the technical proposal.</p>
2.	Ventilator With Compressor & Humidifier (Neonatal) (HFO)	Five firms participated in the bid but there is a some confusion & miss-understanding of different bidders regarding compressor in approved PVMS, the TAC committee recommended to scrap of HFO Ventilator tender & re-advertise.

MINUTES OF TECHNICAL ADVISORY COMMITTEE (TAC) MEETING REGARDING THE PURCHASE OF ELECTRO MEDICAL EQUIPMENT / MISC. / I.T. & ELECTRIC ITEMS FOR 2nd UNIT OF NEONATOLOGY FOR THE YEAR 2021-2022 MEETING HELD ON 26-10-2021 AT 11:00 A.M. IN THE CONFERENCE ROOM OF THE ADMIN BLOCK THE CHILDREN'S HOSPITAL LAHORE.

Tender bid Sr. No.	Item Description	Remarks by TAC Committee
3.	Ventilator With Compressor & Humidifier (Neonatal)	<p>Seven firms participated in the bid and five firms were declared responsive.</p> <p>Details of responsive firms are as under:- M/s Vertex Medical (Pvt) Ltd., M/s Total Technologies (Pvt) Ltd., M/s Hospicare Systems, M/s Shirazi Trading Company (Pvt) Ltd. & M/s Digionics (Pvt) Ltd. were declared responsive by fulfilling the criteria of bid evolution.</p> <p>Details of No responsive firms are as under:- M/s Eastern Medical Technology Quoted model is not compliance to technical specifications . M/s Care Worth Pakistan (Pvt) Ltd. did not attached bid security (CDR / Bank Guarantee) with the technical proposal and submitted loose bid.</p>
4.	Infusion Pumps	<p>Five firms participated in the bid and two firms were responsive</p> <p>Details of responsive firm are as under:- M/s Medical Equipment & Systems & M/s B. Braun Pakistan (Pvt) Ltd. were declared technically responsive.</p> <p>Details of responsive firm are as under:- M/s Human Health Care did not attached original bid security (CDR / Bank Guarantee) with the technical proposal. M/s T.K. Medical Instruments Co. did not attached bid security (CDR / Bank Guarantee) with the technical proposal furthermore bid is conditional on warranty clause as they providing three years warranty with parts instead of five years warranty. M/s Shirazi Trading did not attached the satisfactory past performance certificate of quoted product.</p>

MINUTES OF TECHNICAL ADVISORY COMMITTEE (TAC) MEETING REGARDING THE PURCHASE OF ELECTRO MEDICAL EQUIPMENT / MISC. / I.T. & ELECTRIC ITEMS FOR 2nd UNIT OF NEONATOLOGY FOR THE YEAR 2021-2022 MEETING HELD ON 26-10-2021 AT 11:00 A.M. IN THE CONFERENCE ROOM OF THE ADMIN BLOCK THE CHILDREN'S HOSPITAL LAHORE.

Tender bid Sr. No.	Item Description	Remarks by TAC Committee
5.	Neonatal Intensive Care Incubator	<p>Six firms participated in the bid and two firms ere responsive Details of responsive firms ar as under M/s Saarf Medical Solutions & M/s Medical Equipment & Systems were declared responsive. End-user / TAC committee is of the view that demonstration of the said item / equipment may be called for better assessment as the product quoted by M/s Saarf Medical Solutions is new for enduser and they wants to hands on product. Aletter was sent to said company to arrange demo unit by company failed to provide the demo within give period of time another 1 week is given to arrange demo but not succeed. As the product was not being evaluated so the firm was declare Non responsive. Details of Non responsive firms are as under: M/s Eastern Medical Technology Services do not attached satisfactory past performance in the bid M/s Delta Plus (Pvt) Ltd. do not attached bid security (CDR / Bank Guarantee) with the technical proposal and affidavit from bidder not found in their bid regarding the black listing. Bidding documents not attached. Principle under taking regarding after sale & service not attached. Past performance certificate of quoted products not attached M/s T.K. Medical Instruments Co. did not attached bid security (CDR / Bank Guarantee) with the technical proposal furthermore its bid is conditional on warranty clause as they providing three years warranty with parts instead of five years warranty hence M/s T.K. Medical Instruments Co. M/s Aair Medicals submit its loose bid and bid security (CDR / Bank Guarantee) is less than 2%. Original tender receipt was also not attached and delivery period was not mentioned.</p>
6.	Intensive care Radiant Warmer Resuscitative	<p>Six firms i.e. M/s Medical Equipment & Systems, M/s Eastern Medical Technology Services, M/s Saarf Medical Solutions, M/s T.K. Medical Instruments Co., M/s Delta Plus & M/s Aair Medicals participated in the bid & only one firm i.e. M/s Medical Equipment & Systems declared responsive. Detail of non-responsive firms are as under:- M/s Eastern Medical Technology Services did not attached past performance certificate. M/s Saarf Medical Solutions did not attached past performance certificate M/s T.K. Medical Instruments Co. did not attached original tender receipt & bid security (CDR / Bank Guarantee) with the technical proposal, M/s Delta Plus did not attached bid security (CDR / Bank Guarantee) and affidavit regarding the non-black listing of firm M/s Aair Medicals submitted loose bid, bid security (CDR / Bank Guarantee) and affidavit from bidder regarding the non-black listing was also not attached .</p>

MINUTES OF TECHNICAL ADVISORY COMMITTEE (TAC) MEETING REGARDING THE PURCHASE OF ELECTRO MEDICAL EQUIPMENT / MISC. / I.T. & ELECTRIC ITEMS FOR 2nd UNIT OF NEONATOLOGY FOR THE YEAR 2021-2022 MEETING HELD ON 26-10-2021 AT 11:00 A.M. IN THE CONFERENCE ROOM OF THE ADMIN BLOCK THE CHILDREN'S HOSPITAL LAHORE.

Tender bid Sr. No.	Item Description	Remarks by TAC Committee
7.	Anesthesia Machine	Four firms participated in the bid & three firms out of four were compliance / responsive on bidding evaluation criteria except M/s Shirazi Trading Company (Pvt) Ltd. which was declared non-responsive their quoted model did not comply on technical specifications. As per end user request demonstration of said equipment may be called to evaluate the product. On the basis of demo M/s Digionics were declare Non Responsive by the end user.
8.	Pulse Oximeter (Neonatal)	Four firms participated in the bid & two firms i.e. M/s Hospicare Systems & M/s Saarf Medical Solutions were declared responsive as per bidding evaluation criteria. M/s Al-Kareem Medical Technologies did not comply on defined quality standard hence declared non-responsive. M/s Radiant Medical (Pvt) Ltd. did not submit bidder past performance & did not define quality standard too, also (CE Certificate) submitted by the M/s Radiant Medical (Pvt) Ltd. not compliance with quality standards. End-user / TAC committee is of the view that demonstration of the said item / equipment may be called for better assessment. M/s Hospicare presented demonstration successfully and declare responsive by the end-user. Whereas M/s Saarf Medical failed to present the successfully demo to end-user, hence declared non Responsive .
9.	Neonatal Transport Systems	Five firms participated in the bid and two firms Details of Responsive firms are as under. M/s Vertex Medical (Pvt) Ltd. & M/s Saarf Medical Solutions were declared technically responsive Details of Non Responsive firms are as under. M/s T.K. Medical Instruments Co. did not attached bid security (CDR / Bank Guarantee) and original tender receipt was also not attached with the technical proposal. M/s Eastern Medical Technology Services did not attached bidder performance certificate M/s Aair Medicals did not attached tender receipt, bid security is also less than 2%, bid validity was not available & delivery period was also not mentioned.

[Handwritten signature]

[Handwritten signature]

[Handwritten signature]

[Handwritten signature]

MINUTES OF TECHNICAL ADVISORY COMMITTEE (TAC) MEETING REGARDING THE PURCHASE OF ELECTRO MEDICAL EQUIPMENT / MISC. / I.T. & ELECTRIC ITEMS FOR 2nd UNIT OF NEONATOLOGY FOR THE YEAR 2021-2022 MEETING HELD ON 26-10-2021 AT 11:00 A.M. IN THE CONFERENCE ROOM OF THE ADMIN BLOCK THE CHILDREN'S HOSPITAL LAHORE.

Tender bid Sr. No.	Item Description	Remarks by TAC Committee
10.	A) Bilirubinometer B) Centrifuge Machine	Three firms participated in the bid & all firms were declared non-responsive detail is as under:- M/s Al-Kareem Medical Technologies did not attached past performance certificate, M/s Vital Care did not attached past performance certificate & complete package was also not quoted M/s Human Health Care did not attached technical offer nor the bid security (CDR / Bank Guarantee)
11.	Transcutaneous Bilirubino Meter	Only one firm i.e. M/s Vertex Medical (Pvt) Ltd. participated in the bid & declared technically responsive as per bidding evaluation criteria.
12.	Low Flow Meters	Three firms i.e. M/s Total Technologies (Pvt) Ltd., M/s Al Kareem Medical Technologies & M/s Advance Systems participated in the bid End-user / TAC committee is of the view that demonstration of the said item / equipment may be called for better assessment. M/s Total Technologies presented Demonstration but the lower limit of flow was higher as per clinical application, so rejected as per end-user for Neonates the lower limit of oxygen. Alkareem Medical Technologies Do not present the Demo, hence rejected M/s Advance System do not present the Demo Hence Rejected.
13.	Neonatal Overhead Warmer / Heater (Local)	Three firms participated in the bid & only one firm i.e. M/s Al Kareem Medical Technologies declared responsive & two firms declare non-responsive detail is as under:- M/s Sci tech Enterprises did not attached affidavit regarding the nonblack listing of the firm and its bid was also found loose. M/s Techno Orbits did not attached bid security (CDR / Bank Guarantee) & also there is no affidavit regarding the nonblack listing of the firm
15.	ECG Machine (3 Channel)	Three firms i.e. M/s Shirazi Trading Company (Pvt) Ltd., M/s Radiant Medical (Pvt) Ltd. & M/s Kasbn International participated in the bid and all firms declared responsive.

[Handwritten signatures and initials]


MINUTES OF TECHNICAL ADVISORY COMMITTEE (TAC) MEETING REGARDING THE PURCHASE OF ELECTRO MEDICAL EQUIPMENT / MISC. / I.T. & ELECTRIC ITEMS FOR 2nd UNIT OF NEONATOLOGY FOR THE YEAR 2021-2022 MEETING HELD ON 26-10-2021 AT 11:00 A.M. IN THE CONFERENCE ROOM OF THE ADMIN BLOCK THE CHILDREN'S HOSPITAL LAHORE.

Tender bid Sr. No.	Item Description	Remarks by TAC Committee
16.	Miscellaneous Items (As a package) a. Baby Weighing scale b. Saturation Probes (Neonatal) c. Glucometer d. Ambu Bag with masks (Neonatal) e. Boiler (Feeder) f. PT Meter & APT Meter g. Temperature Clock (Digital) h. KMC Chairs (Kangro Mother Chair) i. Breast Pumps (Manual)	Three firms i.e. M/s 3C international, M/s Life care Technologies were declare Non Responsive. M/s Al Kareem Medical Technologies Declare Responsive .
17.	I.T. / Electric Items Computer System with UPS(APC 650VA) and Black & white Laser Printer	Only one firm i.e. M/s Computer Point participated in the bid and technically responsive.
19.	A) LED TV 42" B.) LED TV 46" (As a package)	Two firms i.e. M/s Computer Point & M/s Mian& Co. participated in the bid, M/s Computer Point technically declared responsive and M/s Mian& Co. was technically non-responsive as the said firms did not mentioned the model in the technical bid.


MINUTES OF TECHNICAL ADVISORY COMMITTEE (TAC) MEETING REGARDING THE PURCHASE OF ELECTRO MEDICAL EQUIPMENT / MISC. / I.T. & ELECTRIC ITEMS FOR 2nd UNIT OF NEONATOLOGY FOR THE YEAR 2021-2022 MEETING HELD ON 26-10-2021 AT 11:00 A.M. IN THE CONFERENCE ROOM OF THE ADMIN BLOCK THE CHILDREN'S HOSPITAL LAHORE.


Tender bid Sr. No.	Item Description	Remarks by TAC Committee
20.	Multimedia Projector	Four firms participated in the bid & three firms i.e. M/s Mian& Co., M/s ABM Data Systems (Pvt) Ltd. & M/s Computer Point declared technically responsive. M/s Nedo Corporation was declared non-responsive as affidavit from bidder & regarding the non black listing & bid security (CDR / Bank Guarantee) were not attached with the technical proposal.

Enduser Concerned / Co-opted Member



Mazhar Iqbal Noor
I.T. Incharge / Enduser


Dr. Farah Haroon
Associate Prof. of Neonatology
Department



Dr. Khawaja Ahmed Irfan Waheed
Prof. of Neonatology



Engr. Irfan Haider
Bio Medical Engineer / Purchase Officer


Technical Advisory Committee



Engr. Shujaat Ali
Bio Medical Engineer / Technical Officer



Riaz Shahid
Computer Programmer / Incharge Payroll



Dr. Zeeshan Sarfraz
AMS (Purchase)



Shoukat Mustafa
Chief Pharmacist



Dr. Syed Salman Ahmad Shah
Assistant Prof. of Cardiovascular Surgery



Dr. Nazish Saqlan
Associate Prof. of Pathology



Dr. Sarfraz Ahmad
Associate Prof. of Radiatric PM&R


Dr. Farah Naz
Associate Prof. of Peds Medicine


Dr. Iqbal Hussain Dogar
Associate Prof. Peds Radiology


Dr. Syed Najam Fayder
Prof. of Pead Cardiology


Dr. Aizza Zafar
Prof. of Microbiology /
Head of Pathology Department


Dr. Nabila Talat
Prof. of Paediatric Surgery /
Head of Surgery Unit-II


Dr. Junaid Rashid
Prof. of Paediatric Medicine /
Head of Medical Unit-I

The Children's Hospital & The Institute of Child Health Lahore

Technical Evaluation

Package No. / Bid No. :36375/01/CIH & ICH

Name of the Equipment:

(A) Non-invasive Cardiac Monitor (As a package)

(B) Invasive Cardiac Monitor

Required Quantity:

10+10 = 20 Nos.

Part - I

KNOCKED DOWN CRITERIA - (BID EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	M/s Vital Care	M/s Bio Tech Services	M/s Delta Plus (Pvt) Ltd	M/s Vertex Medical (Pvt) Ltd	M/s Radiant Medical (Pvt) Ltd	M/s Shirazi Trading Company (Pvt) Ltd	M/s Medical Equipment & Systems	M/s Human Health Care
01	Complete Package / Tender (Mandatory)	Yes	Yes	No	Yes	Yes	Yes	Yes	Yes
02	Original Receipt of Tender (Mandatory)	Yes	Yes	Yes	Yes	Yes	Yes	Yes	Yes
03	Affidavit from Bidder	Yes	Yes	Yes	Yes	Yes	Yes	Yes	Yes
04	Bid Security CDR / Bank Guarantee (Mandatory)	Yes	Yes	No	Yes	Yes	Yes	Yes	No
05	Bid Validity (Mandatory)	Yes	Yes	Yes	Yes	Yes	Yes	Yes	Yes
06	Delivery Period (Mandatory)	Yes	Yes	Yes	Yes	Yes	Yes	Yes	Yes
Remarks:		Responsive	Responsive	Non Responsive	Responsive	Responsive	Responsive	Responsive	Non Responsive

PART - II

KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	M/s Vital Care	M/s Bio Tech Services	M/s Delta Plus (Pvt) Ltd	M/s Vertex Medical (Pvt) Ltd	M/s Radiant Medical (Pvt) Ltd	M/s Shirazi Trading Company (Pvt) Ltd	M/s Medical Equipment & Systems	M/s Human Health Care
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	Both Quoted same Product model	Yes	Both Quoted same Product model	Yes	Yes	Yes	Yes	
02	Technical & Engineering capability (As defined for the specific Bid in specifications)		Yes		Yes	Yes	Yes	Yes	
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal		Yes		Yes	Yes	Yes	Yes	
04	Bidder Past Performance for Quoted Product in Teaching Hospitals		Yes		Yes		Yes	Yes	
05	Availability of relevant Tools and Testing / Calibration Equipment		Yes		Yes	Yes	Yes	Yes	
06	Warranty		Yes		Yes	Yes	Yes	Yes	
Remarks:		Non-Responsive	Responsive	Non-Responsive	Responsive	Responsive	Responsive	Responsive	Non-Responsive

Man

SAS

20

2/1/17

2/1/17

Fazh

2/1/17

PART - III

KNOCKED DOWN CRITERIA - (PRODCUT EVALUATION)

Evaluated by (TAC) Committee

I	Required Item Name	M/s Vital Care	M/s Bio Tech Services	M/s Delta Plus (Pvt) Ltd	M/s Vertex Medical (Pvt) Ltd	M/s Radiant Medical (Pvt) Ltd	M/s Shirazi Trading Company (Pvt) Ltd	M/s Medical Equipment & Systems	M/s Human Health Care
	(A) Non-invasive Cardiac Monitor (As a package) (B) Invasive Cardiac Monitor								
	Brand	Advanced Instrumentations	Nihon Kohden	Advanced Instrumentations	Dragerwerk	Huntleigh	GE-Healthcare	Space Lab	Medicinos GIGA UAB
Sr. #	Quoted Model (Non-Invasive)	PM-2000XL PRO	Life Scope BSM-3562	PM-2000XL PRO	Vista 120S	SC-1200	B125	Qube Compact 91390	Signscope 15
	Invasive	PM-2000XL PRO	Life Scope BSM-3562	PM-2000XL PRO	Vista 120S	SC-1200	B125	Qube Compact 91390	Signscope 15
	Vital Sign Simulator	FLUKE ProSim4	NK AX-400G	Missing	FLUKE Prosim 8	REGEL UNI-SIM	FLUKE Prosim 8/ Equivalent	REGEL UNI-SIM	Model not quoted
01	Country of Manufacturer	USA	Japan	USA	Germany	UK	USA	USA	Lithuania
02	Manufacturing site	USA	Japan	USA	China	UK	China	USA	Lithuania
03	Compliance with defined Quality Standards		Compliance		Compliance	Compliance	Compliance	Compliance	
04	Specifications Compliance Feature Wise		Compliance		Non Compliance (ECG Leads not as per requirement)	Compliance	Compliance	Compliance	
05	Technical Eligibility of Product		Accepted		Not Accepted	Accepted	Accepted	Accepted	
06	Technical Eligibility of Firm		Accepted		Accepted	Accepted	Accepted	Accepted	
Technical Status		Non-Responsive	Responsive	Non-Responsive	Non-Responsive	Responsive	Responsive	Responsive	Non-Responsive

Enduser Concerned / Co-opted Member

Farah
Dr. Farah Haroon
Associate Prof. of Neonatology Department
Technical Advisory Committee

Engr. Shujaat Ali
Engr. Shujaat Ali
Bio Medical Engineer / Technical Officer

Dr. Nazish Saqlan
Dr. Nazish Saqlan
Associate Prof. of Pathology

Dr. Aizza Zafar
Dr. Aizza Zafar
Prof. of Microbiology / Head of Pathology Department

Riaz Shahid
Riaz Shahid
Computer Programmer / Incharge Payroll

Dr. Sarfraz Ahmad
Dr. Sarfraz Ahmad
Associate Prof. of Paediatric PM&R

Dr. Khawaja Ahmed Irfan Waheed
Dr. Khawaja Ahmed Irfan Waheed
Prof. of Neonatology

Dr. Zeeshan Sarfraz
Dr. Zeeshan Sarfraz
AMS (Purchase)

Dr. Farah Naz
Dr. Farah Naz
Associate Prof. of Peds Medicine

Dr. Nabila Talat
Dr. Nabila Talat
Prof. of Paediatric Surgery / Head of Surgery Unit-II

Engr. Irfan Haider
Engr. Irfan Haider
Bio Medical Engineer / Purchase Officer / Coordinator

Dr. Syed Salman Ahmad Shah
Dr. Syed Salman Ahmad Shah
Assistant Prof. of Cardiovascular Surgery

Shoukat Mustafa
Shoukat Mustafa
Chief Pharmacist

Dr. Iqbal Hussain Dogar
Dr. Iqbal Hussain Dogar
Associate Prof. Peds Radiology

Dr. Syed Najam Hayder
Dr. Syed Najam Hayder
Prof. of Paed Cardiology

Dr. Junaid Rashid
Dr. Junaid Rashid
Prof. of Paediatric Medicine / Head of Medical Unit-I



The Children's Hospital & The Institute of Child Health Lahore

Technical Evaluation

Package No. / Bid No. :36375/02/CH & ICH

Name of the Equipment:

Required Quantity:

VENTILATOR WITH COMPRESSOR & HUMIDIFIER (Neonatal) (HFO)

10 Nos.

Part - I

KNOCKED DOWN CRITERIA - (BID EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	M/s Total Technologies (Pvt) Ltd	M/s Vertex Medical (Pvt) Ltd	M/s Clinical Life Inc.	M/s Meditech (Pvt) Ltd	M/s Digionics (Pvt) Ltd
01	Complete Package / Tender (Mandatory)	Yes	Yes	Yes	Yes	Yes
02	Original Receipt of Tender (Mandatory)	Yes	Yes	Yes	Yes	Yes
03	Affidavit from Bidder	Yes	Yes	Yes	Yes	Yes
04	Bid Security CDR / Bank Guarantee (Mandatory)	Yes	Yes	Yes	copy attach without amount	Yes
05	Bid Validity (Mandatory)	Yes	Yes	Yes	Yes	Yes
06	Delivery Period (Mandatory)	Yes	Yes	Yes	Yes	Yes
Remarks:						

PART - II

KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	M/s Total Technologies (Pvt) Ltd	M/s Vertex Medical (Pvt) Ltd	M/s Clinical Life Inc.	M/s Meditech (Pvt) Ltd	M/s Digionics (Pvt) Ltd
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	Yes	Yes	Yes	No	Yes
02	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes	Yes	No	Yes
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal	Yes	Yes	Yes	No	Yes
04	Bidder Past Performance for Quoted Product in Teaching Hospitals	Usman Memorial Hospital	Yes	Yes	No	Yes
05	Availability of relevant Tools and Testing / Calibration Equipment	Yes	Yes	Yes	Yes	Yes
06	Warranty	Yes	Yes	Yes		Yes
Remarks:						

PART - III

KNOCKED DOWN CRITERIA - (PRODUCT EVALUATION)

Evaluated by (IAC) Committee

Sr. #	Required Item Name	M/s Total Technologies (Pvt) Ltd	M/s Vertex Medical (Pvt) Ltd	M/s Clinical Life Inc.	M/s Meditech (Pvt) Ltd	M/s Digionics (Pvt) Ltd
	VENTILATOR WITH COMPRESSOR & HUMIDIFIER (Neonatal) (HFO)					
	Brand	Lowenstein Medical	Drägerwerk	E. Stephan	SLE	MAQUET
	Quoted Model	Leonitus plus	Babylog VN600	Sophie	6000	SERVO-N
01	Country of Manufacturer	Germany	Germany	Germany	UK	Sweden
02	Manufacturing site	Germany	Germany	Germany	UK	Sweden
03	Compliance with defined Quality Standards	Compliance	Compliance	Compliance	Compliance	Compliance
04	Specifications Compliance Feature Wise	Compliance	Compliance	Compliance	Compliance	Compliance
05	Technical Eligibility of Product	Compliance without compressor	Accepted	Compliance without compressor	Compliance without compressor	Accepted
06	Technical Eligibility of Firm		Accepted	Accepted	Not Accepted	
Technical Status						

Note: Due to confusion & miss understanding of different bidder regarding compressor in approved PVMS. Committee decided to retender in order to make healthy competition.

Enduser Concerned / Co-opted Member

Dr. Farah Haroon
Associate Prof. of Neonatology Department

Dr. Khawaja Ahmed Irfan Waheed
Prof. of Neonatology

Engr. Irfan Haider
Bio Medical Engineer / Purchase Officer

Technical Advisory Committee

Engr. Shujaat Ali
Bio Medical Engineer / Technical Officer

Riaz Shahid
Computer Programmer / Incharge Payroll

Dr. Zeeshan Sarfraz
AMS (Purchase)

Shoukat Mustafa
Chief Pharmacist

Dr. Syed Salman Ahmad Shah
Assistant Prof. of Cardiovascular Surgery

Dr. Nazish Saqlan
Associate Prof. of Pathology

Dr. Sarfraz Ahmad
Associate Prof. of Paediatric PM&R

Dr. Farah Naz
Associate Prof. of Peds Medicine

Dr. Iqbal Hussain Dogar
Associate Prof. Peds Radiology

Dr. Syed Najam Haider
Prof. of Paed Cardiology

Dr. Aizza Zafar
Prof. of Microbiology / Head of Pathology Department

Dr. Nabila Talat
Prof. of Paediatric Surgery / Head of Surgery Unit-II

Dr. Junaid Rashid
Prof. of Paediatric Medicine / Head of Medical Unit-I



The Children's Hospital & The Institute of Child Health Lahore

Technical Evaluation

Package No. / Bid No. :36375/03/CH & ICH

Name of the Equipment:

Ventilator With Compressor & Humidifier (Neonatal)

Required Quantity:

05 Nos.

Part - I

KNOCKED DOWN CRITERIA - (BID EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	M/s Vertex Medical (Pvt) Ltd	M/s Total Technologies (Pvt) Ltd	M/s Eastern Medical Technology Services	M/s Hospicare Systems	M/s Shirazi Trading Company (Pvt) Ltd	M/s Digionics (Pvt) Ltd	M/s Cares Worth Pakistan (Pvt) Ltd
01	Complete Package / Tender (Mandatory)	Yes	Yes	Yes	Yes	Yes	Yes	Yes
02	Original Receipt of Tender (Mandatory)	Yes	Yes	Yes	Yes	Yes	Yes	Yes
03	Affidavit from Bidder	Yes	Yes	Yes	Yes	Yes	Yes	Yes
04	Bid Security CDR / Bank Guarantee (Mandatory)	Yes	Yes	Yes	Yes	Yes	Yes	CDR not original
05	Bid Validity (Mandatory)	Yes	Yes	Yes	Yes	Yes	Yes	Yes
06	Delivery Period (Mandatory)	Yes	Yes	Yes	Yes	Yes	Yes	Yes
Remarks:		Responsive	Responsive	Responsive	Responsive	Responsive	Responsive	Non Responsive

PART - II

KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	M/s Vertex Medical (Pvt) Ltd	M/s Total Technologies (Pvt) Ltd	M/s Eastern Medical Technology Services	M/s Hospicare Systems	M/s Shirazi Trading Company (Pvt) Ltd	M/s Digionics (Pvt) Ltd	M/s Cares Worth Pakistan (Pvt) Ltd
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	Yes	Yes	Yes	Yes	Yes	Yes	
02	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes	Yes	Yes	Yes	Yes	
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal	Yes	Yes	Yes	Yes	Yes	Yes	
04	Bidder Past Performance for Quoted Product in Teaching Hospitals	Yes	Yes	Yes	Yes	Yes	Yes	
05	Availability of relevant Tools and Testing / Calibration Equipment	Yes	Yes	Yes	Yes	Yes	Yes	
06	Warranty	Yes	Yes	Yes	Yes	Yes	Yes	
Remarks:		Responsive	Responsive	Responsive	Responsive	Responsive	Responsive	Non Responsive

[Handwritten signature]

[Handwritten signature]

[Handwritten signature]

[Handwritten signature]

[Handwritten signature]

[Handwritten signature]

[Handwritten signature]

Fazh

[Handwritten signature]


PART - III


KNOCKED DOWN CRITERIA - (PRODCUT EVALUATION)

Evaluated by (TAC) Committee

1	Required Item Name	M/s Vertex Medical (Pvt) Ltd	M/s Total Technologies (Pvt) Ltd	M/s Eastern Medical Technology Services	M/s Hospicare Systems	M/s Shirazi Trading Company (Pvt) Ltd	M/s Digionics (Pvt) Ltd	M/s Cares Worth Pakistan (Pvt) Ltd
	Ventilator With Compressor & Humidifier (Neonatal)							
Sr. #	Brand	Dragerwerk	Lowenstein Medical	Event	Medtronic-Covidien AG	GE-Healthcare, USA	Maquet Getinge	Atlantamed
01	Quoted Model	Evita V600	Elisa 600	Inspiration 7i	PB 980	Caresscape R860	Servo-U	Diomedeo NEO
02	Country of Manufacturer	Germany	Germany	USA	USA	USA	Sweden	UK
03	Manufacturing site	Germany	Germany	USA	Ireland	USA/ China	Sweden	UK
04	Compliance with defined Quality Standards	Compliance	Compliance	Compliance	Compliance	Compliance	Compliance	
05	Specifications Compliance Feature Wise	Compliance	Compliance	Non-Compliance(Automatic compressor not truns on)	Compliance	Compliance	Compliance	
06	Technical Eligibility of Product	Accepted	Accepted	Not Accepted	Accepted	Accepted	Accepted	
	Technical Eligibility of Firm	Accepted	Accepted	Accepted	Accepted	Accepted	Accepted	
	Technical Status	Responsive	Responsive	Non-Responsive	Responsive	Responsive	Responsive	Non-Responsive


Enduser Concerned / Co-opted Member


Dr. Farah Haroon
Associate Prof. of Neonatology
Department



Dr. Khawaja Ahmed Irfan Waheed
Prof. of Neonatology


Engr. Irfan Haider
Bio Medical Engineer / Purchase Officer


Technical Advisory Committee



Ewaj Shujant Ali
Bio Medical Engineer / Technical Officer



Riaz Shahid
Computer Programmer / Incharge Payroll


Dr. Zeeshan Sarfraz
AMS (Purchase)



Shoukat Mustafa
Chief Pharmacist



Dr. Syed Salman Ahmad Shah
Assistant Prof. of Cardiovascular Surgery



Dr. Nadeem Saqlan
Associate Prof. of Pathology



Dr. Sarfraz Ahmad
Associate Prof. of Paediatric PM&R


Dr. Farah Naz
Associate Prof. of Paeds
Medicine


Dr. Iqbal Hussain Dogar
Associate Prof. Paeds Radiology


Dr. Syed Najam Hayder
Prof. of Paed Cardiology


Dr. Aizza Zafar
Prof. of Microbiology / Head of Pathology
Department


Dr. Nabila Zafar
Prof. of Paediatric Surgery / Head of
Surgery Unit-II


Dr. Junaid Rashid
Prof. of Paediatric Medicine / Head of
Medical Unit-I



The Children's Hospital & The Institute of Child Health Lahore

Technical Evaluation

Package No. / Bid No. :36375/04/CH & ICH

Name of the Equipment:

INFUSION PUMPS

Required Quantity:

40 Nos.

Part - I

KNOCKED DOWN CRITERIA - (BID EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	M/s Medical Equipment & Systems	M/s Human Health Care	M/s Shirazi Trading	M/s T.K. Medical Instruments Co	M/s B. Braun Pakistan (Pvt) Ltd
01	Complete Package / Tender (Mandatory)	Yes	Yes	Yes	Yes	Yes
02	Original Receipt of Tender (Mandatory)	Yes	Yes	Yes	Yes	Yes
03	Affidavit from Bidder	Yes	Yes	Yes	Yes	Yes
04	Bid Security CDR / Bank Guarantee (Mandatory)	Yes	No	Yes	No (Copy of CDR attached not original)	Yes
05	Bid Validity (Mandatory)	Yes	Yes	Yes	Yes	Yes
06	Delivery Period (Mandatory)	Yes	Yes	Yes	Yes	Yes
Remarks:		Responsive	Non Responsive	Responsive	Non Responsive	Responsive

PART - II

KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	M/s Medical Equipment & Systems	M/s Human Health Care	M/s Shirazi Trading	M/s T.K. Medical Instruments Co	M/s B. Braun Pakistan (Pvt) Ltd
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	Yes		Yes		Yes
02	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes		Yes		Yes
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal	Yes		Yes		Yes
04	Bidder Past Performance for Quoted Product in Teaching Hospitals	Yes		No		Yes
05	Availability of relevant Tools and Testing / Calibration Equipment	Yes		Yes		Yes
06	Warranty	Yes		Yes	No (3 years warranty quoted instead of 5 years)	Yes
Remarks:		Responsive	Non Responsive	Non Responsive	Non Responsive	Responsive


PART - III


KNOCKED DOWN CRITERIA - (PRODCUT EVALUATION)


Evaluated by (TAC) Committee

1	Required Item Name	M/s Medical Equipment & Systems	M/s Human Health Care	M/s Shirazi Trading	M/s T.K. Medical Instruments Co	M/s B. Braun Pakistan (Pvt) Ltd
	Infusion Pumps					
	Brand	Fresenius KABI's	MEDICINOS GIJA UAB	Progetti	CANAFUSION Technologies	B Braun
Sr. #	Quoted Model	Agilia VP	MEDIFUSION IP	PG807I	CA 2000	Infusomat Compact Plus
01	Country of Manufacturer	Germany	Lithuania	Italy	Canada	Germany
02	Manufacturing site	China	Lithuania	Italy	Canada	Germany
03	Compliance with defined Quality Standards	Compliance				Compliance
04	Specifications Compliance Feature Wise	Compliance				Compliance
05	Technical Eligibility of Product	Accepted				Accepted
06	Technical Eligibility of Firm	Accepted				Accepted
Technical Status		Responsive	Non Responsive	Non Responsive	Non Responsive	Responsive


Enduser Concerned / Co-opted Member



Dr. Farah Haroon
Associate Prof. of Neonatology
Department



Dr. Khawaja Ahmed Irfan Waheed
Prof. of Neonatology



Engr. Irfan Haider
Bio Medical Engineer / Purchase Officer
/ Coordinator


Technical Advisory Committee



Engr. Shujaat Ali
Bio Medical Engineer / Technical Officer



Riaz Shahid
Computer Programmer / Incharge Payroll



Dr. Zeeshan Sarfraz
AMS (Purchase)


Shoukat Mustafa
Chief Pharmacist


Dr. Syed Salman Ahmad Shah
Assistant Prof. of Cardiovascular Surgery



Dr. Nazish Saqlan
Associate Prof. of Pathology



Dr. Sarfraz Ahmad
Associate Prof. of Peadiatric PM&R



Dr. Farah Naz
Associate Prof. of Peads Medicine


Dr. Iqbal Hussain Dogar
Associate Prof. Peds Radiology


Dr. Syed Najam Hayder
Prof. of Pead Cardiology


Dr. Aizza Zafar
Prof. of Microbiology / Head of
Pathology Department


Dr. Nabila Talat
Prof. of Paediatric Surgery / Head of Surgery
Unit-II


Dr. Junaid Rashid
Prof. of Paediatric Medicine / Head of
Medical Unit-I



The Children's Hospital & The Institute of Child Health Lahore

Technical Evaluation

Package No. / Bid No. :36375/05/CH & ICH

Name of the Equipment:

NEONATAL INTENSIVE CARE INCUBATOR

Required Quantity:

28 Nos.

Part - I

KNOCKED DOWN CRITERIA - (BID EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	M/s Eastern Medical Technology Services	M/s Delta Plus (Pvt) Ltd	M/s T.K. Medical Instruments Co.	M/s Saarf Medical Solutions	M/s Medical Equipment & Systems	M/s Aair Medicals
01	Complete Package / Tender (Mandatory)	Yes	Yes	Yes	Yes	Yes	Yes
02	Original Receipt of Tender (Mandatory)	Yes	Yes	Yes	Yes	Yes	No
03	Affidavit from Bidder	Yes	No	Yes	Yes	Yes	No
04	Bid Security CDR / Bank Guarantee (Mandatory)	Yes	copy attached without amount	No	Yes	Yes	Less then 2%
05	Bid Validity (Mandatory)	Yes	Yes	Yes	Yes	Yes	Yes
06	Delivery Period (Mandatory)	Yes	Yes	Yes	Yes	Yes	No
Remarks:		Responsive	Non Responsive	Non Responsive	Responsive	Responsive	Non Responsive

PART - II

KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	M/s Eastern Medical Technology Services	M/s Delta Plus (Pvt) Ltd	M/s T.K. Medical Instruments Co.	M/s Saarf Medical Solutions	M/s Medical Equipment & Systems	M/s Aair Medicals
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	Yes			Yes	Yes	
02	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes			Yes	Yes	
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal	Yes			Yes	Yes	
04	Bidder Past Performance for Quoted Product in Teaching Hospitals	No			Yes	Yes	
05	Availability of relevant Tools and Testing / Calibration Equipment	Yes			Yes	Yes	
06	Warranty	Yes		No	Yes	Yes	
Remarks:		Non Responsive		Non Responsive	Responsive	Responsive	

PART - III

KNOCKED DOWN CRITERIA - (PRODCUT EVALUATION)

Evaluated by (TAC) Committee

1	Required Item Name:	M/s Eastern Medical Technology Services	M/s Delta Plus (Pvt) Ltd	M/s T.K. Medical Instruments Co.	M/s Saarf Medical Solutions	M/s Medical Equipment & Systems	M/s Aair Medicals
	Neonatal Intensive Care Incubator						
	Brand	TSE Medical	Advanced Instrumentations Inc.		COBAMS	ATOM	
Sr. #	Quoted Model	SI-610-1 (Shelly)	A3186+		Cristina Sch	Air Incu i	
01	Country of Manufacturer	Czech Republic	USA		Italy	Japan	
02	Manufacturing site	Czech Republic	USA		Italy	Japan	
03	Compliance with defined Quality Standards				Compliance	Compliance	
04	Specifications Compliance Feature Wise				Compliance	Compliance	
05	Technical Eligibility of Product				Compliance	Accepted	
06	Technical Eligibility of Firm		Signed Bidding Document not attached, Principal undertaking of after sales and service not attached, Past Performance not attached		Demo Not Accepted	Accepted	
Technical Status		Non Responsive	Non Responsive	Non Responsive	Non Responsive	Responsive	Non Responsive

Enduser Concerned / Co-opted Member

Dr. Farah Haroon
Associate Prof. of Neonatology Department
Technical Advisory Committee

Engr. Shujaat Ali
Bio Medical Engineer / Technical Officer

Dr. Nazish Saqlan
Associate Prof. of Pathology

Dr. Aizza Zafar
Prof. of Microbiology / Head of Pathology Department

Riaz Shahid
Computer Programmer / Incharge Payroll

Dr. Safriz Ahmad
Associate Prof. of Paediatric PM&R

Dr. Nabila Tahir
Prof. of Paediatric Surgery / Head of Surgery Unit-II

Dr. Khawaja Ahmed Irfan Waheed
Prof. of Neonatology

Dr. Zeeshan Sarfraz
AMS (Purchase)

Dr. Farah Naz
Associate Prof. of Peds Medicine

Shoukat Mustafa
Chief Pharmacist

Dr. Iqbal Hussain Dogar
Associate Prof. Peds Radiology

Engr. Irfan Haider
Bio Medical Engineer / Purchase Officer / Coordinator

Dr. Syed Salman Ahmad Shah
Assistant Prof. of Cardiovascular Surgery

Dr. Syed Najam Haider
Prof. of Paed Cardiology

Dr. Junaid Rashid
Prof. of Paediatric Medicine / Head of Medical Unit-I



The Children's Hospital & The Institute of Child Health Lahore

Technical Evaluation

Package No. / Bid No. :36375 /06/CH & ICH

Name of the Equipment:

INTENSIVE CARE RADIANT WARMER RESUSCITATIVE

Required Quantity:

20 Nos.

Part - I

KNOCKED DOWN CRITERIA - (BID EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	M/s Medical Equipment & Systems	M/s Eastern Medical Technology Services	M/s Saarf Medical Solutions	M/s T.K. Medical Instruments Co	M/s Delta Plus	M/s Aair Medicals
01	Complete Package / Tender (Mandatory)	Yes	Yes	Yes	Yes	Yes	Yes (loose bid)
02	Original Receipt of Tender (Mandatory)	Yes	Yes	Yes	No	Yes	Yes
03	Affidavit from Bidder	Yes	Yes	Yes	Yes	No	No
04	Bid Security CDR / Bank Guarantee (Mandatory)	Yes	Yes	Yes	No	No	Less then 2%
05	Bid Validity (Mandatory)	Yes	Yes	Yes	Yes	Yes	Yes
06	Delivery Period (Mandatory)	Yes	Yes	Yes	Yes	Yes	Yes
Remarks:		Responsive	Responsive	Responsive	Non Responsive	Non Responsive	Non Responsive

PART - II

KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	M/s Medical Equipment & Systems	M/s Eastern Medical Technology Services	M/s Saarf Medical Solutions	M/s T.K. Medical Instruments Co	M/s Delta Plus	M/s Aair Medicals
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	Yes	Yes	Yes			
02	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes	Yes			
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal	Yes	Yes	Yes			
04	Bidder Past Performance for Quoted Product in Teaching Hospitals	Yes	NO	No			
05	Availability of relevant Tools and Testing / Calibration Equipment	Yes	Yes	Yes			
06	Warranty	Yes	Yes	Yes	No		
Remarks:		Responsive	Non Responsive	Non Responsive	Non Responsive	Non Responsive	Non Responsive

PART - III

KNOCKED DOWN CRITERIA - (PRODCUT EVALUATION)

1	Required Item Name	Evaluated by (TAC) Committee					
	Intensive care Radiant Warmer Resuscitative	M/s Medical Equipment & Systems	M/s Eastern Medical Technology Services	M/s Saarf Medical Solutions	M/s T.K. Medical Instruments Co	M/s Delta Plus	M/s Aair Medicals
Sr. #	Brand	Atoms	TSE	COBAMS			
01	Quoted Model	Sunflower	RW610(NATALLAY Basic)	LR 90 07A			
02	Country of Manufacturer	Japan	Czech Republic	Italy			
03	Manufacturing site	Japan	Czech Republic	Italy			
04	Compliance with defined Quality Standards	Compliance					
05	Specifications Compliance Feature Wise	Compliance					
06	Technical Eligibility of Product	Accepted					
	Technical Eligibility of Firm	Accepted					
	Technical Status	Responsive	Non Responsive	Non Responsive	Non Responsive	Non Responsive	Non Responsive

Enduser Concerned / Co-opted Member

Dr. Farah Haroon
Associate Prof. of Neonatology
Department

Technical Advisory Committee

Engr. Shujaat Ali
Bio Medical Engineer / Technical Officer

Dr. Nazish Saqlan
Associate Prof. of Pathology

Dr. Aizza Zafar
Prof. of Microbiology / Head of Pathology
Department

Dr. Khawaja Ahmed Irfan Waheed
Prof. of Neonatology

Engr. Irfan Haider
Bio Medical Engineer / Purchase Officer / Coordinator

Riaz Shahid
Computer Programmer / Incharge
Payroll

Dr. Zeeshan Sarfraz
AMS (Purchase)

Shoukat Mustafa
Chief Pharmacist

Dr. Syed Salman Ahmad Shah
Assistant Prof. of Cardiovascular Surgery

Dr. Sarfraz Ahmad
Associate Prof. of
Paediatric PM&R

Dr. Farah Naz
Associate Prof. of Peds
Medicine

Dr. Iqbal Hussain Dogar
Associate Prof. Peds Radiology

Dr. Syed Najam Haider
Prof. of Paediatric Radiology

Dr. Nabila Talat
Prof. of Paediatric Surgery / Head of
Surgery Unit-II

Dr. Junaid Rashid
Prof. of Paediatric Medicine / Head of Medical
Unit-I



The Children's Hospital & The Institute of Child Health Lahore

Technical Evaluation

Package No. / Bid No. :36375 /07/CH & ICH

Name of the Equipment:

Anesthesia Machine

Required Quantity:

02 Nos.

Part - I

KNOCKED DOWN CRITERIA - (BID EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	M/s Shirazi Trading Company (Pvt) Ltd	M/s Vertex Medical (Pvt) Ltd	M/s Total Technologies (Pvt) Ltd	M/s Digionics (Pvt) Ltd
01	Complete Package / Tender (Mandatory)	Yes	Yes	Yes	Yes
02	Original Receipt of Tender (Mandatory)	Yes	Yes	Yes	Yes
03	Affidavit from Bidder	Yes	Yes	Yes	Yes
04	Bid Security CDR / Bank Guarantee (Mandatory)	Yes	Yes	Yes	Yes
05	Bid Validity (Mandatory)	Yes	Yes	Yes	Yes
06	Delivery Period (Mandatory)	Yes	Yes	Yes	Yes
Remarks:		Responsive	Responsive	Responsive	Responsive

PART - II

KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	M/s Shirazi Trading Company (Pvt) Ltd	M/s Vertex Medical (Pvt) Ltd	M/s Total Technologies (Pvt) Ltd	M/s Digionics (Pvt) Ltd
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	Yes	Yes	Yes	Yes
02	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes	Yes	Yes
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal	Yes	Yes	Yes	Yes
04	Bidder Past Performance for Quoted Product in Teaching Hospitals	Yes	Yes	Yes	Yes
05	Availability of relevant Tools and Testing / Calibration Equipment	Yes	Yes	Yes	Yes
06	Warranty	Yes	Yes	Yes	Yes
Remarks:		Responsive	Responsive	Responsive	Responsive

PART - III					
KNOCKED DOWN CRITERIA - (PRODCUT EVALUATION)					
Evaluated by (TAC) Committee					
1	Required Item Name	M/s Shirazi Trading Company (Pvt) Ltd	M/s Vertex Medical (Pvt) Ltd	M/s Total Technologies (Pvt) Ltd	M/s Diglonics (Pvt) Ltd
	Anesthesia Machine				
	Brand	GE Healthcare	Dragerwerk	Loweinstein\Nk	Geringe Sweden\Penlon
Sr. #	Quoted Model	Avacne CS2-B850	Atlan A350-IACS C500	Leon Plus-CSM1702 with BSM 1763	Flow i-Annavue 4000
01	Country of Manufacturer	USA	Germany	Germany-Japan	Sweden
02	Manufacturing site	USA-Mexico	Germany	Germany-Japan	Sweden
03	Compliance with defined Quality Standards	Complaine	Complaine	Complaine	Complaine
04	Specifications Compliance Feature Wise	(Integrating heating issue)	Complaine	Complaine	Complaine
05	Technical Eligibility of Product	Not Accepted	Accepted	Accepted	Demo Not Accepted
06	Technical Eligibility of Firm	Accepted	Accepted	Accepted	Accepted
Technical Status		Non Responsive	Responsive	Responsive	Non Responsive

Enduser Concerned / Co-opted Member

Dr. Farah Haroon
Associate Prof. of Neonatology
Department
Technical Advisory Committee

Dr. Saeedah Asaf
Associate Prof. Peadiatric Anesthesiology

Dr. Khawaja Ahmed Irfan Waheed
Prof. of Neonatology

Engr. Irfan Haider
Bio Medical Engineer / Purchase Officer / Coordinator

Engr. Shujaat Ali
Bio Medical Engineer / Technical
Officer

Riaz Shahid
Computer Programmer / Incharge
Payroll

Dr. Zeeshan Sarfraz
AMS (Purchase)

Shoukat Mustafa
Chief Pharmacist

Dr. Syed Salman Ahmad Shah
Assistant Prof. of Cardiovascular Surgery

Dr. Nazish Saqlan
Associate Prof. of Pathology

Dr. Sarfraz Ahmad
Associate Prof. of Peadiatric
PM&R

Dr. Farah Naz
Associate Prof. of Peads Medicine

Dr. Iqbal Hussain Dogar
Associate Prof. Peds Radiology

Dr. Syed Najam Hayder
Prof. of Pead Cardiology

Dr. Aizza Zafar
Prof. of Microbiology / Head of
Pathology Department

Dr. Nabila Talat
Prof. of Paediatric Surgery / Head of
Surgery Unit-II

Dr. Javed Rashid
Prof. of Paediatric Medicine / Head of
Medical Unit-I



The Children's Hospital & The Institute of Child Health Lahore

Technical Evaluation

Package No. / Bid No. :36375 /08/CH & ICH

Name of the Equipment:

Pulse Oximeter (Neonatal)

Required Quantity:

37 Nos.

Part - I

KNOCKED DOWN CRITERIA - (BID EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	M/s Hospicare Systems	M/s Al Kareem Medical Technologies	M/s Saarf Medical Solutions	M/s Radiant Medical (Pvt) Ltd
01	Complete Package / Tender (Mandatory)	Yes	Yes	Yes	Yes
02	Original Receipt of Tender (Mandatory)	Yes	Yes	Yes	Yes
03	Affidavit from Bidder	Yes	Yes	Yes	Yes
04	Bid Security CDR / Bank Guarantee (Mandatory)	Yes	Yes	Yes	Yes
05	Bid Validity (Mandatory)	Yes	Yes	Yes	Yes
06	Delivery Period (Mandatory)	Yes	Yes	Yes	Yes
Remarks:		Responsive	Responsive	Responsive	Responsive

PART - II

KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	M/s Hospicare Systems	M/s Al Kareem Medical Technologies	M/s Saarf Medical Solutions	M/s Radiant Medical (Pvt) Ltd
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	Yes	Yes	Yes	Yes
02	Technical & Engineering capability (As defined for the specific Bid in specifications)	yes	Yes	Yes	Yes
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal	yes		Yes	Yes
04	Bidder Past Performance for Quoted Product in Teaching Hospitals	yes		Yes	NO
05	Availability of relevant Tools and Testing / Calibration Equipment	yes		Yes	Yes
06	Warranty	yes		Yes	Yes
Remarks:		Responsive		Responsive	Non Responsive

PART - III

KNOCKED DOWN CRITERIA - (PRODCUT EVALUATION)

Evaluated by (TAC) Committee

1	Required Item Name	M/s Hospicare Systems	M/s Al Kareem Medical Technologies	M/s Saarf Medical Solutions	M/s Radiant Medical (Pvt) Ltd
	Pulse Oximeter (Neonatal)				
	Brand	Meditronic-Covidien	NOVA TECHINOCARE	aXcent Medical	ZUGMED medical system
Sr. #	Quoted Model	PM-100N	NPT-80	PAVO	SPLF-OX
01	Country of Manufacturer	USA	UK	Germany	France
02	Manufacturing site	Korea		Germany	France
03	Compliance with defined Quality Standards	Yes	Not Compliance	Yes	Not Compliance own under taking
04	Specifications Compliance Feature Wise	Compliance		Yes	Compliance
05	Technical Eligibility of Product	Accepted		Demo Not Accepted	Not Accepted
06	Technical Eligibility of Firm	Accepted		Accepted	Not Accepted
	Technical Status	Responsive	NonResponsive	Non Responsive	Non Responsive

Enduser Concerned / Co-opted Member

Dr. Farah Haroon
Associate Prof. of Neonatology
Department

Dr. Khawaja Ahmed Irfan Waheed
Prof. of Neonatology

Eng. Irfan Haider
Bio Medical Engineer / Purchase
Officer / Coordinator

Technical Advisory Committee

Engr. Shujaat Ali
Bio Medical Engineer / Technical Officer

Riaz Shahid
Computer Programmer / Incharge Payroll

Dr. Zeeshan Sarfraz
AMS (Purchase)

Sboukat Mustafa
Chief Pharmacist

Dr. Syed Salman Ahmad Shah
Assistant Prof. of Cardiovascular
Surgery

Dr. Nazish Saqlan
Associate Prof. of Pathology

Dr. Sarfraz Ahmad
Associate Prof. of Paediatric
PM&R

Dr. Farah Naz
Associate Prof. of Peds Medicine

Dr. Iqbal Hussain Dogar
Associate Prof. Peds Radiology

Dr. Syed Nafan Hayder
Prof. of Paed Cardiology

Dr. Aizza Zafar
Prof. of Microbiology / Head of Pathology
Department

Dr. Nabila Palat
Prof. of Paediatric Surgery / Head of
Surgery Unit-II

Dr. Junaid Rashid
Prof. of Paediatric Medicine /
Head of Medical Unit-I



The Children's Hospital & The Institute of Child Health Lahore

Technical Evaluation

Package No. / Bid No. :36375 /09/CH & ICH

Name of the Equipment:

Neonatal Transport Systems

Required Quantity:

02 Nos.

Part - I

KNOCKED DOWN CRITERIA - (BID EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	M/s T.K. Medical Instruments Co.	M/s Eastern Medical Technology Services	M/s Vertex Medical (Pvt) Ltd	M/s Saarf Medical Solutions	M/s Aair Medicals
01	Complete Package / Tender (Mandatory)	Yes	Yes	Yes	Yes	Yes
02	Original Receipt of Tender (Mandatory)	No	Yes	Yes	Yes	No
03	Affidavit from Bidder	Yes	Yes	Yes	Yes	Yes
04	Bid Security CDR / Bank Guarantee (Mandatory)	No	Yes	Yes	Yes	Less then 2%
05	Bid Validity (Mandatory)	Yes	Yes	Yes	Yes	No
06	Delivery Period (Mandatory)	Yes	Yes	Yes	Yes	No
Remarks:		Non Responsive	Responsive	Responsive	Responsive	Non Responsive

PART - II

KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	M/s T.K. Medical Instruments Co.	M/s Eastern Medical Technology Services	M/s Vertex Medical (Pvt) Ltd	M/s Saarf Medical Solutions	M/s Aair Medicals
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer		Yes	Yes	Yes	
02	Technical & Engineering capability (As defined for the specific Bid in specifications)		Yes	Yes	Yes	
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal		Yes	Yes	Yes	
04	Bidder Past Performance for Quoted Product in Teaching Hospitals		No	Yes	Yes	
05	Availability of relevant Tools and Testing / Calibration Equipment		Yes	Yes	Yes	
06	Warranty		Yes	Yes	Yes	
Remarks:		Non Responsive	Non Responsive	Responsive	Responsive	

PART - III

KNOCKED DOWN CRITERIA - (PRODUCT EVALUATION)

Evaluated by (TAC) Committee						
1	Required Item Name	M/s T.K. Medical Instruments Co.	M/s Eastern Medical Technology Services	M/s Vertex Medical (Pvt) Ltd	M/s Saarf Medical Solutions	M/s Aair Medicals
	Neonatal Transport Systems					
	Brand	Olidex Medical	TSE Medical	Dragerwerk	COBAMS	Astra Med
Sr. #	Quoted Model	RWT PLUS	Baby Active Ti 401	Isollete TI500	Stm Baby Cristina	AMT 700
01	Country of Manufacturer	USA	Czech Republic	Germany	Italy	
02	Manufacturing site	USA	Czech Republic	USA	Italy	
03	Compliance with defined Quality Standards			Compliance	Compliance	
04	Specifications Compliance Feature Wise			Compliance	Compliance	
05	Technical Eligibility of Product			Accepted	Accepted	
06	Technical Eligibility of Firm	Not Accepted	Not Accepted	Accepted	Accepted	
Technical Status		Non Responsive	Non Responsive	Responsive	Responsive	Non Responsive

Enduser Concerned / Co-opted Member

Dr. Farah Haroon
Associate Prof. of Neonatology
Department

Dr. Khawaja Ahmed Irfan Waheed
Prof. of Neonatology

Engr. Irfan Haider
Bio Medical Engineer / Purchase Officer /
Coordinator

Technical Advisory Committee

Engr. Shujaat Ali
Bio Medical Engineer / Technical Officer

Riaz Shahid
Computer Programmer / Incharge Payroll

Dr. Zeeshan Sarfraz
AMS (Purchase)

Shoukat Mustafa
Chief Pharmacist

Dr. Syed Salman Ahmad Shah
Assistant Prof. of Cardiovascular
Surgery

Dr. Nazish Saqlan
Associate Prof. of Pathology

Dr. Sarfraz Ahmad
Associate Prof. of Paediatric PM&R

Dr. Farah Naz
Associate Prof. of Peds Medicine

Dr. Iqbal Hussain Dogar
Associate Prof. Peds Radiology

Dr. Syed Najam Haider
Prof. of Peds Cardiology

Dr. Aizza Zafar
Prof. of Microbiology / Head of Pathology
Department

Dr. Nabila Talat
Prof. of Paediatric Surgery / Head of Surgery
Unit-II

Dr. Usaid Rashid
Prof. of Paediatric Medicine / Head of
Medical Unit-I



The Children's Hospital & The Institute of Child Health Lahore

Technical Evaluation

Package No. / Bid No. :36375 /10/CH & ICH

Name of the Equipment:

A) Bilirubinometer
B) Centrifuge Machine

Required Quantity:

02+02=4

Part - I

KNOCKED DOWN CRITERIA - (BID EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	M/s Al Kareem Medical Technologies	M/s Vital Care	M/s Human Health Care
01	Complete Package / Tender (Mandatory)	Yes	Yes	Yes(No technical offer)
02	Original Receipt of Tender (Mandatory)	Yes	Yes	Yes
03	Affidavit from Bidder	Yes	Yes	Yes
04	Bid Security CDR / Bank Guarantee (Mandatory)	Yes	Yes	No
05	Bid Validity (Mandatory)	Yes	Yes	Yes
06	Delivery Period (Mandatory)	Yes	Yes	Yes
Remarks:		Responsive	Responsive	Non Responsive

PART - II

KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	M/s Al Kareem Medical Technologies	M/s Vital Care	M/s Human Health Care
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	Yes	Yes	
02	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes	
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal	Yes	Yes	
04	Bidder Past Performance for Quoted Product in Teaching Hospitals	No	No	
05	Availability of relevant Tools and Testing / Calibration Equipment	Yes	Yes	
06	Warranty	Yes	Yes	
Remarks:		Non Responsive	Non Responsive	

PART - III

KNOCKED DOWN CRITERIA - (PRODUCT EVALUATION)

Evaluated by (TAC) Committee

1	Required Item Name	M/s Al Kareem Medical Technologies	M/s Vital Care	M/s Human Health Care
	A) Bilirubinometer B) Centrifuge Machine			
	Brand	DAS Sri/Labtron	Paramedical	
Sr. #	Quoted Model	NEO-BIL PLUS/LHS-A 10	PKL PPC 512(centrifuge only)	
01	Country of Manufacturer			
02	Manufacturing site			
03	Compliance with defined Quality Standards			
04	Specifications Compliance Feature Wise			
05	Technical Eligibility of Product			
06	Technical Eligibility of Firm			
Technical Status		Non Responsive	Non Responsive	Non Responsive

Enduser Concerned / Co-opted Member

Farah

Dr. Farah Haroon
Associate Prof. of Neonatology
Department

[Signature]

Dr. Khawaja Ahmed Irfan Waheed
Prof. of Neonatology

[Signature]

Engr. Irfan Haider
Bio Medical Engineer / Purchase
Officer / Coordinator

Technical Advisory Committee

[Signature]

Engr. Shujaat Ali
Bio Medical Engineer / Technical Officer

[Signature]

Riaz Shahid
Computer Programmer / Incharge
Payroll

[Signature]

Dr. Zeeshan Sarfraz
AMS (Purchase)

[Signature]

Shoukat Mustafa
Chief Pharmacist

[Signature]

Dr. Syed Salman Ahmad Shah
Assistant Prof. of Cardiovascular
Surgery

[Signature]

Dr. Nazish Saqlan
Associate Prof. of Pathology

[Signature]

Dr. Sarfraz Ahmad
Associate Prof. of Paediatric PM&R

[Signature]

Dr. Farah Naz
Associate Prof. of Peds Medicine

[Signature]

Dr. Iqbal Hussain Dogar
Associate Prof. Peds Radiology

[Signature]

Dr. Syed Najam Hayder
Prof. of Paed Cardiology

[Signature]

Dr. Aizza Zafar
Prof. of Microbiology / Head of
Pathology Department

[Signature]

Dr. Nabila Talat
Prof. of Paediatric Surgery / Head of
Surgery Unit-II

[Signature]

Dr. Junaid Rashid
Prof. of Paediatric Medicine /
Head of Medical Unit-I



The Children's Hospital & The Institute of Child Health Lahore

Technical Evaluation

Package No. / Bid No. :36375 /11 / CH & ICH

Name of the Equipment:

Transcutaneous Bilirubino Meter

Required Quantity:

02 Nos.

Part - I

KNOCKED DOWN CRITERIA - (BID EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	M/s Vertex Medical (Pvt) Ltd
01	Complete Package / Tender (Mandatory)	Yes
02	Original Receipt of Tender (Mandatory)	Yes
03	Affidavit from Bidder	Yes
04	Bid Security CDR / Bank Guarantee (Mandatory)	Yes
05	Bid Validity (Mandatory)	Yes
06	Delivery Period (Mandatory)	Yes
Remarks:		Responsive

PART - II

KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	M/s Vertex Medical (Pvt) Ltd
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	Yes
02	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal	Yes
04	Bidder Past Performance for Quoted Product in Teaching Hospitals	Yes
05	Availability of relevant Tools and Testing / Calibration Equipment	Yes
06	Warranty	Yes
Remarks:		Responsive

PART - III		
KNOCKED DOWN CRITERIA - (PRODCUT EVALUATION)		
Evaluated by (TAG) Committee		
1	Required Item Name	M/s Vertex Medical (Pvt) Ltd
	Transcutaneous Bilirubino Meter	
	Brand	Dragerwerk,
Sr. #	Quoted Model	JM-105
01	Country of Manufacturer	Germany
02	Manufacturing site	Germany
03	Compliance with defined Quality Standards	Compliance
04	Specifications Compliance Feature Wise	Compliance
05	Technical Eligibility of Product	Accepted
06	Technical Eligibility of Firm	Accepted
Technical Status		Responsive

Enduser Concerned / Co-opted Member

Dr. Farah Haroon
Associate Prof. of Neonatology Department

Dr. Khawaja Ahmed Irfan Wahced
Prof. of Neonatology

Engr. Irfan Haider
Bio Medical Engineer / Purchase Officer
/ Coordinator

Technical Advisory Committee

Engr. Shujaat Ali
Bio Medical Engineer / Technical Officer

Riaz Shahid
Computer Programmer / Incharge Payroll

Dr. Zeeshan Sarfraz
AMS (Purchase)

Shoukat Mustafa
Chief Pharmacist

Dr. Syed Salman Ahmad Shah
Assistant Prof. of Cardiovascular
Surgery

Dr. Nadish Saqlan
Associate Prof. of Pathology

Dr. Sarfraz Ahmad
Associate Prof. of Paediatric PM&R

Dr. Farah Naz
Associate Prof. of Peds Medicine

Dr. Iqbal Hussain Dogar
Associate Prof. Peds Radiology

Dr. Syed Najam Haider
Prof. of Paed. Cardiology

Dr. Aizza Zafar
Prof. of Microbiology / Head of Pathology
Department

Dr. Nabila Palat
Prof. of Paediatric Surgery / Head of
Surgery Unit-II

Dr. Junaid Rashid
Prof. of Paediatric Medicine /
Head of Medical Unit-I



The Children's Hospital & The Institute of Child Health Lahore

Technical Evaluation

Package No. / Bid No. :36375 /12 / CH & ICH

Name of the Equipment:

Low Flow Meters

Required Quantity:

50 Nos.

Part - I

KNOCKED DOWN CRITERIA - (BID EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	M/s Total Technologies (Pvt) Ltd	M/s Al Kareem Medical Technologies	M/s Advance System
01	Complete Package / Tender (Mandatory)	Yes	Yes	Yes
02	Original Receipt of Tender (Mandatory)	Yes	Yes	Yes
03	Affidavit from Bidder	Yes	Yes	Yes
04	Bid Security CDR / Bank Guarantee (Mandatory)	Yes	Yes	Yes
05	Bid Validity (Mandatory)	Yes	Yes	Yes
06	Delivery Period (Mandatory)	Yes	Yes	Yes
Remarks:		Responsive	Responsive	Responsive

PART - II

KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	M/s Total Technologies (Pvt) Ltd	M/s Al Kareem Medical Technologies	M/s Advance System
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	Yes	Yes	Yes
02	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes	Yes
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal	Yes	Yes	Yes
04	Bidder Past Performance for Quoted Product in Teaching Hospitals	Yes	Yes	Yes
05	Availability of relevant Tools and Testing / Calibration Equipment	Yes	Yes	Yes
06	Warranty	Yes	Yes	Yes
Remarks:		Responsive	Responsive	Responsive

PART - III

KNOCKED DOWN CRITERIA - (PRODCUT EVALUATION)

Evaluated by (TAC) Committee

1	Required Item Name	M/s Total Technologies (Pvt) Ltd	M/s Al Kareem Medical Technologies	M/s Advance System
	Low Flow Meters			
	Brand			
Sr. #	Quoted Model	Flowmeter	Maxtec	Muller France
01	Country of Manufacturer	Easymed+CH 200	RP34P03-005(bottel model missing)	Nanotube 10911-Microbulle
02	Manufacturing site	Italy	USA	France
03	Compliance with defined Quality Standards	Italy	USA	France
04	Specifications Compliance Feature Wise	Compliance	Compliance	compliance
05	Technical Eligibility of Product	Compliance		
06	Technical Eligibility of Firm	Demo not accepted	Demo not accepted	Demo not accepted
		Accepted	Accepted	Accepted
	Technical Status	Non Responsive	Non Responsive	Non Responsive

Enduser Concerned / Co-opted Member

Dr. Farah Haroon
Associate Prof. of Neonatology
Department

Dr. Khawaja Ahmed Irfan Waheed
Prof. of Neonatology

Engr. Irfan Haider
Bio Medical Engineer / Purchase Officer / Coordinator

Technical Advisory Committee

Engr. Shujaat Ali
Bio Medical Engineer / Technical Officer

Riaz Shahid
Computer Programmer /
Incharge Payroll

Dr. Zeeshan Sarfraz
AMS (Purchase)

Shoukat Mustafa
Chief Pharmacist

Dr. Syed Salman Ahmad Shah
Assistant Prof. of Cardiovascular Surgery

Dr. Nazish Saqlan
Associate Prof. of Pathology

Dr. Sarfraz Ahmad
Associate Prof. of
Padiatric PM&R

Dr. Farah Naz
Associate Prof. of Peds
Medicine

Dr. Iqbal Hussain Dogar
Associate Prof. Peds Radiology

Dr. Syed Najam Hayder
Prof. of Pead Cardiology

Dr. Aizza Zafar
Prof. of Microbiology / Head of Pathology
Department

Dr. Nabila Tahir
Prof. of Paediatric Surgery / Head
of Surgery Unit-II

Dr. Junaid Rashid
Prof. of Paediatric Medicine / Head of Medical
Unit-I



The Children's Hospital & The Institute of Child Health Lahore

Technical Evaluation

Package No. / Bid No. :36375 /13 / CH & ICH

Name of the Equipment:

Neonatal Overhead Warmer / Heater (Local)

Required Quantity:

02 Nos.

Part - I

KNOCKED DOWN CRITERIA - (BID EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	M/s Sci tech Enterprises	M/s Techno Orbits	M/s Al Kareem Medical Technologies
01	Complete Package / Tender (Mandatory)	Yes	Yes	Yes
02	Original Receipt of Tender (Mandatory)	Yes	Yes	Yes
03	Affidavit from Bidder	No	No	Yes
04	Bid Security CDR / Bank Guarantee (Mandatory)	No (Loose Bid)	No	Yes
05	Bid Validity (Mandatory)	Yes	Yes	Yes
06	Delivery Period (Mandatory)	Yes	Yes	Yes
Remarks:		Non Responsive	Non Responsive	Responsive

PART - II

KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	M/s Sci tech Enterprises	M/s Techno Orbits	M/s Al Kareem Medical Technologies
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer			N/A
02	Technical & Engineering capability (As defined for the specific Bid in specifications)			Yes
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal			Yes
04	Bidder Past Performance for Quoted Product in Teaching Hospitals			Yes
05	Availability of relevant Tools and Testing / Calibration Equipment			Yes
06	Warranty			Yes
Remarks:				Responsive

PART - III

KNOCKED DOWN CRITERIA - (PRODUCT EVALUATION)

Evaluated by (TAC) Committee				
1	Required Item Name	M/s Sci tech Enterprises	M/s Techno Orbits	M/s Al Kareem Medical Technologies
	Neonatal Overhead Warmer / Heater (Local)			
	Brand			
Sr. #	Quoted Model			Al Kareem
01	Country of Manufacturer			AK 174
02	Manufacturing site			local
03	Compliance with defined Quality Standards			local
04	Specifications Compliance Feature Wise			Yes
05	Technical Eligibility of Product			Yes
06	Technical Eligibility of Firm			Accepted
Technical Status		Non Responsive	Non Responsive	Responsive

Enduser Concerned / Co-opted Member

Dr. Farah Haroon
Associate Prof. of Neonatology
Department

Dr. Khawaja Ahmed Irfan Waheed
Prof. of Neonatology

Engr. Irfan Haider
Bio Medical Engineer / Purchase Officer

Technical Advisory Committee

Engr. Shujaat Ali
Bio Medical Engineer / Technical Officer

Riaz Shahid
Computer Programmer / Incharge Payroll

Dr. Zeeshan Sarfraz
AMS (Purchase)

Shoukat Mustafa
Chief Pharmacist

Dr. Syed Salman Ahmad Shah
Assistant Prof. of Cardiovascular
Surgery

Dr. Nazish Saqlan
Associate Prof. of Pathology

Dr. Sarfraz Ahmad
Associate Prof. of Paediatric PM&R

Dr. Farah Naz
Associate Prof. of Peds Medicine

Dr. Iqbal Hussain Dogar
Associate Prof. Peds Radiology

Dr. Syed Najam Hayder
Prof. of Paed Cardiology

Dr. Aizza Zafar
Prof. of Microbiology / Head of Pathology
Department

Dr. Nabila Talat
Prof. of Paediatric Surgery /
Head of Surgery Unit-II

Dr. Junaid Rashid
Prof. of Paediatric Medicine /
Head of Medical Unit-I



The Children's Hospital & The Institute of Child Health Lahore

Technical Evaluation

Package No. / Bid No. :36375 /15 / CH & ICH

Name of the Equipment:

ECG Machine (3 Channel)

Required Quantity:

01 No.

Part - I

KNOCKED DOWN CRITERIA - (BID EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	M/s Shirazi Trading Company (Pvt) Ltd	M/s Radiant Medical (Pvt) Ltd	M/s Kasbn International
01	Complete Package / Tender (Mandatory)	Yes	Yes	Yes
02	Original Receipt of Tender (Mandatory)	Yes	Yes	Yes
03	Affidavit from Bidder	Yes	Yes	Yes
04	Bid Security CDR / Bank Guarantee (Mandatory)	Yes	Yes	Yes
05	Bid Validity (Mandatory)	Yes	Yes	Yes
06	Delivery Period (Mandatory)	Yes	Yes	Yes
Remarks:		Responsive	Responsive	Responsive

PART - II

KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	M/s Shirazi Trading Company (Pvt) Ltd	M/s Radiant Medical (Pvt) Ltd	M/s Kasbn International
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	Yes	Yes	Yes
02	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	yes	Yes
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal	Yes	yes	Yes
04	Bidder Past Performance for Quoted Product in Teaching Hospitals	Yes	Yes	Yes
05	Availability of relevant Tools and Testing / Calibration Equipment	Yes	Yes	Yes
06	Warranty	Yes	Yes	Yes
Remarks:		Responsive	Responsive	Responsive


PART - III


KNOCKED DOWN CRITERIA - (PRODUCT EVALUATION)


Evaluated by (TAC) Committee

1	Required Item Name	M/s Shirazi Trading Company (Pvt) Ltd	M/s Radiant Medical (Pvt) Ltd	M/s Kasbn International
	ECG Machine (3 Channel)			
	Brand	GE Health care	Hill Rom (Welch Allyn)	LUMED
Sr. #	Quoted Model	MAC 600	CP 50	EURO ECG 3 veiw
01	Country of Manufacturer	USA	USA	Italy
02	Manufacturing site	INDIA	USA	Italy
03	Compliance with defined Quality Standards	Yes	Yes	Yes
04	Specifications Compliance Feature Wise	Compliance	Compliance	
05	Technical Eligibility of Product	Accepted	Accepted	yes
06	Technical Eligibility of Firm	Accepted	Accepted	Accepted
Technical Status		Responsive	Responsive	Responsive


Enduser Concerned / Co-opted Member



Dr. Farah Haroon
Associate Prof. of Neonatology
Department



Dr. Khawaja Ahmed Irfan Waheed
Prof. of Neonatology



Engr. Irfan Haider
Bio Medical Engineer / Purchase Officer /
Coordinator


Technical Advisory Committee



Engr. Shujaat Ali
Bio Medical Engineer / Technical Officer



Riaz Shahid
Computer Programmer / Incharge Payroll



Dr. Zeeshan Sarfraz
AMS (Purchase)



Shoukat Mustafa
Chief Pharmacist



Dr. Syed Salman Ahmad Shah
Assistant Prof. of Cardiovascular Surgery



Dr. Nazish Saqlian
Associate Prof. of Pathology



Dr. Sarfraz Ahmad
Associate Prof. of Peadiatric PM&R



Dr. Farah Naz
Associate Prof. of Peads Medicine


Dr. Iqbal Hussain Dogar
Associate Prof. Peds Radiology


Dr. Syed Najam Hayder
Prof. of Pead Cardiology


Dr. Aizza Zafar
Prof. of Microbiology / Head of Pathology
Department


Dr. Nabila Talat
Prof. of Paediatric Surgery / Head of Surgery Unit-II


Dr. Junaid Rashid
Prof. of Paediatric Medicine / Head of
Medical Unit-I



The Children's Hospital & The Institute of Child Health Lahore

Technical Evaluation

Package No. / Bid No. :36375 /16 / CH & ICH

Name of the Equipment:

Miscellaneous Items (As a package)

Required Quantity:

As per list

Part - I

KNOCKED DOWN CRITERIA - (BID EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	M/s 3C International	M/s Lifecare Technologies	M/s Al Kareem Medical Technologies
01	Complete Package / Tender (Mandatory)	Yes	Yes	Yes
02	Original Receipt of Tender (Mandatory)	Yes	Yes	Yes
03	Affidavit from Bidder	Yes	Yes	Yes
04	Bid Security CDR / Bank Guarantee (Mandatory)	No	No	Yes
05	Bid Validity (Mandatory)	Yes	Yes	Yes
06	Delivery Period (Mandatory)	Yes	Yes	Yes
Remarks:		Non Responsive	Non Responsive	Responsive

PART - II

KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	M/s 3C International	M/s Lifecare Technologies	M/s Al Kareem Medical Technologies
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	N/A	N/A	NA
02	Technical & Engineering capability (As defined for the specific Bid in specifications)	N/A	N/A	NA
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal	N/A	N/A	NA
04	Bidder Past Performance for Quoted Product in Teaching Hospitals	N/A	N/A	NA
05	Availability of relevant Tools and Testing / Calibration Equipment	N/A	N/A	NA
06	Warranty	N/A	N/A	NA
Remarks:		Non Responsive	Non-Responsive	Responsive


PART - III


KNOCKED DOWN CRITERIA - (PRODCUT EVALUATION)


Evaluated by (TAC) Committee

1	Required Item Name	M/s 3C International	M/s Lifecare Technologies	M/s Al Kareem Medical Technologies
	Miscellaneous Items (As a package)			
	Brand			
Sr. #	Quoted Model			
01	Country of Manufacturer			Local
02	Manufacturing site			Local
03	Compliance with defined Quality Standards			Compliance
04	Specifications Compliance Feature Wise			Compliance
05	Technical Eligibility of Product			Accepted
06	Technical Eligibility of Firm			Accepted
Technical Status		Non Responsive	Non Responsive	Responsive


Enduser Concerned / Co-opted Member



Dr. Farah Haroon
Associate Prof. of Neonatology
Department



Dr. Khawaja Ahmed Irfan Waheed
Prof. of Neonatology



Engr. Irfan Haider
Bio Medical Engineer / Purchase Officer /
Coordinator


Technical Advisory Committee



Engr. Shujaat Ali
Bio Medical Engineer / Technical Officer



Riaz Shahid
Computer Programmer / Incharge Payroll



Dr. Zeeshan Sarfraz
AMS (Purchase)



Shoukat Mustafa
Chief Pharmacist



Dr. Syed Salman Ahmad Shah
Assistant Prof. of Cardiovascular Surgery



Dr. Nazish Saqlan
Associate Prof. of Pathology



Dr. Sarfraz Ahmad
Associate Prof. of Paediatric PM&R



Dr. Farah Naz
Associate Prof. of Paeds Medicine


Dr. Iqbal Hussain Dogar
Associate Prof. of Paeds Radiology


Dr. Syed Najam Hayder
Prof. of Paed. Cardiology


Dr. Aizza Zafar
Prof. of Microbiology / Head of Pathology
Department


Dr. Nabila Talat
Prof. of Paediatric Surgery / Head of
Surgery Unit-II


Dr. Junaid Rashid
Prof. of Paediatric Medicine / Head of
Medical Unit-I



The Children's Hospital & The Institute of Child Health
Ferozpur Road, Lahore
Phone No. +92 42 99230901 – 23

Technical Evaluation

Package No. / Bid No. : PC/36375/17 / CH & ICH

Name of the Equipment: Computer Systems, Printers and UPS

Required Quantity: 10

Part - I

KNOCKED DOWN CRITERIA - (BID EVALUATION)

Evaluated by Purchase Committee

Sr. #	Evaluation Parameters	M/s Computer Point
01	Complete Package / Tender (Mandatory)	Yes
02	Original Receipt of Tender (Mandatory)	Yes
03	Affidavit from Bidder	Yes
04	Bid Security / CDR (Mandatory)	Yes
05	Bid Validity (Mandatory)	Yes
06	Delivery Period (Mandatory)	Yes
Remarks:		Responsive

PART - II

KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)


Evaluated by Technical Committee


Sr. #	Evaluation Parameters	M/s Computer Point
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	N/A
02	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal	N/A
04	Bidder Past Performance for Quoted item	Yes
05	Availability of relevant Tools and Testing / Calibration Equipment	Yes
06	Warranty	Yes
Remarks:		Responsive


PART - III


KNOCKED DOWN CRITERIA (PRODUCT EVALUATION)		
Evaluated by Technical Committee		
Required Item Name		Computer Systems, Printers and UPS
Brand		Dell /HP/ APC
Sr. 65	Quoted Model	Computer: Dell Vostro 3888, Printer: HP 107W, UPS: APC 650 VA
01	Country of Manufacturer	China
02	Country of Origin of Product	USA
03	Compliance with defined Quality Standards	Yes
04	Specifications Compliance Feature Wise	Yes
05	Technical Eligibility of Product	Yes
06	Technical Eligibility of Firm	Yes
Technical Status		Responsive

Enduser Concerned / Co-opted Member



Mazhar Iqbal Noor
I.T. Incharge / Enduser



Dr. Farah Haroon
Associate Prof. of Neonatology
Department



Dr. Khawaja Ahmed Irfan Waheed
Prof. of Neonatology



Engr. Irfan Haider
Bio Medical Engineer / Purchase Officer /
Coordinator


Technical Advisory Committee



Engr. Shujaat Ali
Bio Medical Engineering Technical Officer



Riaz Shahid
Computer Programmer / Incharge Payroll



Dr. Zeeshan Sarfraz
AMS (Purchase)



Shoukat Mustafa
Chief Pharmacist



Dr. Syed Salman Ahmad Shah
Assistant Prof. of Cardiovascular Surgery



Dr. Nazish Saqlan
Associate Prof. of Pathology



Dr. Sarfraz Ahmad
Associate Prof. of Paediatric PM&R



Dr. Farah Naz
Associate Prof. of Peds Medicine


Dr. Iqbal Hussain Dogar
Associate Prof. Peds Radiology


Dr. Syed Najam Hyyder
Prof. of Paed Cardiology


Dr. Aizza Zafar
Prof. of Microbiology / Head of Pathology
Department


Dr. Nabila Palat
Prof. of Paediatric Surgery / Head of
Surgery Unit-II


Dr. Junaid Rashid
Prof. of Paediatric Medicine / Head of
Medical Unit-I



The Children's Hospital & The Institute of Child Health
Ferozpur Road, Lahore
Phone No. +92 42 99230901 – 23

Technical Evaluation

Package No. / Bid No. : PC/36375/19 / CH & ICH

Name of the Equipment: LED Tv 42" = 03 , LED Tv 46 " = 02

Required Quantity: 03+02

Part - I

KNOCKED DOWN CRITERIA - (BID EVALUATION)			
Evaluated by Purchase Committee			
Sr. #	Evaluation Parameters	M/s Computer Point	M/s Mian & co.
01	Complete Package / Tender (Mandatory)	Yes	Yes
02	Original Receipt of Tender (Mandatory)	Yes	Yes
03	Affidavit from Bidder	Yes	Yes
04	Bid Security / CDR (Mandatory)	Yes	Yes
05	Bid Validity (Mandatory)	Yes	Yes
06	Delivery Period (Mandatory)	Yes	Yes
Remarks:		Responsive	Responsive

PART - II

KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)			
Evaluated by Technical Committee			
Sr. #	Evaluation Parameters	M/s Computer Point	M/s Mian & co.
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	N/A	N/A
02	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal	N/A	N/A
04	Bidder Past Performance for Quoted item	Yes	Yes
05	Availability of relevant Tools and Testing / Calibration Equipment	Yes	Yes
06	Warranty	Yes	Yes
Remarks:		Responsive	Responsive

Package No. / Bid No. : PC/36375/ 19/ CH & ICH

PART - III

KNOCKED DOWN CRITERIA (PRODUCT EVALUATION)			
Evaluated by Technical Committee		M/s Computer Point	M/s Mian & co.
Required Item Name		LED TV	LED TV
Brand		TCL	Izone
Sr. 65	Quoted Model	TCL 43" 43P615-4K, TCL 50" 50P615-4K	Model Not Menstioned
01	Country of Manufacturer	China	Pakistan
02	Country of Origin of Product	China	Pakistan
03	Compliance with defined Quality Standards	Yes	N/A
04	Specifications Compliance Feature Wise	Yes	N/A
05	Technical Eligibility of Product	Yes	N/A
06	Technical Eligibility of Firm	Yes	N/A
Technical Status		Responsive	Non-Responsive

Enduser Concerned / Committed Member

Mazhar Iqbal Noor
I.T. Incharge / Enduser

Dr. Farah Haroon
Associate Prof. of Neonatology
Department

Dr. Khawaja Ahmed Irfan Waheed
Prof. of Neonatology

Engr. Irfan Haider
Bio Medical Engineer / Purchase Officer /
Coordinator

Technical Advisory Committee

Engr. Shujaat Ali
Bio Medical Engineer / Technical Officer

Riaz Shahid
Computer Programmer / Incharge Payroll

Dr. Zeeshan Sarfraz
AMS (Purchase)

Shoukat Mustafa
Chief Pharmacist

Dr. Syed Salman Ahmad Shah
Assistant Prof. of Cardiovascular Surgery

Dr. Mazish Saqlan
Associate Prof. of Pathology

Dr. Sarfraz Ahmad
Associate Prof. of Peadiatric PM&R

Dr. Farah Naz
Associate Prof. of Peds Medicine

Dr. Iqbal Hussain Dogar
Associate Prof. Peds Radiology

Dr. Syed Najam Hayer
Prof. of Pead Cardiology

Dr. Aizza Zafar
f. of Microbiology / Head of Pathology
Department

Dr. Nabila Patai
Prof. of Paediatric Surgery / Head of
Surgery Unit-II

Dr. Junaid Rashid
Prof. of Paediatric Medicine / Head of
Medical Unit-I

Technical Evaluation

Package No. _____
 Part No. _____
 Name of the Equipment
 Required Quantity _____

THE CHILDREN'S HOSPITAL & ICH
 MILITARY PROCTOR

Part - I

TECHNICAL EVALUATION		M/s. Mian & Co.			
Evaluation Parameters		M/s. Mian & Co.			
No. 1	Weightage (0-100) (Marked)	Yes	Yes	Yes	Yes
No. 2	Weightage (0-100) (Marked)	Yes	Yes	Yes	Yes
No. 3	Weightage (0-100) (Marked)	Yes	Yes	Yes	Yes
No. 4	Weightage (0-100) (Marked)	Yes	Yes	Yes	Yes
No. 5	Weightage (0-100) (Marked)	Yes	Yes	Yes	Yes
No. 6	Weightage (0-100) (Marked)	Yes	Yes	Yes	Yes
No. 7	Weightage (0-100) (Marked)	Yes	Yes	Yes	Yes
No. 8	Weightage (0-100) (Marked)	Yes	Yes	Yes	Yes
Remarks		Responsive	Non-Responsive	Responsive	Responsive

Part - II

TECHNICAL EVALUATION		M/s. Mian & Co.			
Evaluation Parameters		M/s. Mian & Co.			
No. 1	Weightage (0-100) (Marked)	Yes	Yes	Yes	Yes
No. 2	Weightage (0-100) (Marked)	Yes	Yes	Yes	Yes
No. 3	Weightage (0-100) (Marked)	Yes	Yes	Yes	Yes
No. 4	Weightage (0-100) (Marked)	Yes	Yes	Yes	Yes
No. 5	Weightage (0-100) (Marked)	Yes	Yes	Yes	Yes
No. 6	Weightage (0-100) (Marked)	Yes	Yes	Yes	Yes
No. 7	Weightage (0-100) (Marked)	Yes	Yes	Yes	Yes
No. 8	Weightage (0-100) (Marked)	Yes	Yes	Yes	Yes
Remarks		Responsive	Non-Responsive	Responsive	Responsive

PART - III

KNOCKED DOWN CRITERIA (PROFIT EVALUATION)					
Evaluated by Technical Committee		M/s Mian & Co.	M/s Nedo Corporation	M/s ABM Data Systems (Pvt.) Ltd	M/s Computer point
Required Item Name					
Brand		Panasonic	Maxell	Epson	Panasonic
Sr. #	Quoted Model	Panasonic VW 540	Maxell EU5001	Epson EB-2265U	Panasonic PTVZ 580
01	Country of Manufacturer	China	N/A	China	China
02	Country of Origin of Product	Japan	N/A	Japan	Japan
03	Compliance with defined Quality Standards	Yes	N/A	Yes	Yes
04	Specifications Compliance Feature Wise	Yes	N/A	Yes	Yes
05	Technical Eligibility of Product	Yes	N/A	Yes	Yes
06	Technical Eligibility of Firm	Yes	N/A	Yes	Yes
Technical Status		Responsive	Non-Responsive	Responsive	Responsive

Enduser Concerned / Guided Member

Muzhar Iqbal Noor
I.T. Incharge / Enduser

Dr. Farah Haroon
Associate Prof. of Neonatology
Department

Dr. Khawaja Ahmed Irfan Waheed
Prof. of Neonatology

Eng. Irfan Haider
Bio Medical Engineer / Purchase Officer

Technical Advisory Committee

Dr. Shaukat Ali
Bio Medical Engineer / Technical Officer

Riaz Shahid
Computer Programmer / Incharge Payroll

Dr. Zeeshan Saifraz
AMS (Purchase)

Shaukat Mustafa
Chief Pharmacist

Dr. Syed Salman Ahmad Shah
Assistant Prof. of Cardiovascular Surgery

Dr. Nazish Saqlan
Associate Prof. of Pathology

Dr. Sarfraz Ahmad
Associate Prof. of Paediatric PM&R

Dr. Farah Naz
Associate Prof. of Paeds Medicine

Dr. Iqbal Hussain Dogar
Associate Prof. Paeds Radiology

Dr. Syed Najam Haider
Prof. of Paed Cardiology

Dr. Aizza Zahar
Prof. of Microbiology / Head of Pathology
Department

Dr. Nabila Salat
Prof. of Paediatric Surgery / Head of
Surgery Unit-II

Dr. Junaid Rashid
Prof. of Paediatric Medicine / Head of
Medical Unit-I