The Children's Hospital & The Institute of Child Health Ferozepur Road, Lahore Phone # (92) (42) 99230901-23 Fax # 99231560



No. P.C./ 26689 /CH&ICH Dated: 19-05-

2018

То

Mr. Abdul Wahab I/C Programmer (I.T.) Health Department, Civil Secretariat, Lahore.

Subject: -**REQUEST FOR UPLOADING THE TECHNICAL EVALUATION FOR** THE PROCUREMENT OF ELECTRO MEDICAL EQUIPMENT FOR SURGERY ESTABLISHMENT PAEDIATRIC AND OF ALLIED SECIALITIES IN THE CHILDREN'S HOSPITAL, LAHORE FOR THE WEBSITE/HEALTH DEPARTMENT YEAR 2017-2018 ON PPRA WEBSITE

Please find enclosed herewithTechnical Evsalution committee reports for the procurement of Electro Medical Equipment Establishment of Paediatric Surgery and Allied Specialities for the year 2017-2018 in The Children's Hospital, Lahore and needs to be uploaded in the Health Department / PPRA website.

Your positive response will be highly appreciated.

Prof. Dr. Muhammad Saleem Professor of Paediatric Surgery Medical Director 1 an

No. P.C./

/CH&ICH

Dated

2018

A copy is forwarded for information to Additional Secretary (Development), Government of the Punjab, Specialized Healthcare & Medical Education Department, Lahore.

Prof. Dr. Muhammad Saleem (FCPS_MME-HPE Professor of Paediatric Surgery Medical Director



The Children's Hospital & The Institute of Child Health Ferozpur Road, Lahore Phone No. +92 42 99230901 – 23

Technical Evaluation

Package No. / Bid No. : 10110 / 27 / CH & ICH

Name of the Equipment:

Instrument Set (Local) for Surgery

Required Quantity:

As per package

	Part - I				
			KNOCKED DOWN	CRITERIA - (BID EVALUATION)	
Evaluate	d by Purchase Committee				
Sr. #	Evaluation Parameters	Surgiquips	Medap International	Orient Medical	
01	Complete Package / Tender (Mandatory)	Yes	No	Yes	
02	Original Receipt of Tender (Mandatory)	Yes	Yes	Yes	
03	Affidavit from Bidder	Yes	Yes	Yes	
04	Bid Security / CDR (Mandatory)	Yes	Yes	Yes	
05	Bid Validity (Mandatory)	Yes	Yes	Yes	
06	Delivery Period (Mandatory)	Yes	Yes	Yes	
	Remarks:	Responsive	Non - Responsive	Responsive	

Package No. / Bid No. : 10110 / 27 / CH & ICH

	PART - II					
	KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)					
valuate	d by Technical Committee					
Sr. #	Evaluation Parameters	Surgiquips	Medap International	Orient Medical		
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	Yes	Yes	Yes		
02	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes	Yes		
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal	Yes	Yes	Yes		
04	Bidder Past Performance	Satisfactory	Satisfactory	Satisfactory		
05	Availability of relevant Tools and Testing / Calibration Equipment	Yes	Yes	Yes		
06	Warranty	Yes	Yes	Yes		
	Remarks:	Responsive	Responsive	Responsive		

Package No. / Bid No. : 10110 / 27 / CH & ICH

PART - III

KNOCKED DOWN CRITERIA - (PRODCUT EVALUATION)

eq. Item #	Required Item Name	Surgiquips	Medap International	Orient Medical
27	Instrument Set (Local) for Surgery	011	·	
27	Brand	Hilbro	STI Surgical	Falcon
Sr. #	Quoted Model	N/A	N/A	N/A
01	Country of Manufacturer	Pakistan	Pakistan	Pakistan
02	Country of Origin of Product	Pakistan	Pakistan	Pakistan
03	Compliance with defined Quality Standards	Compliance	Compliance	Compliance
04	Specifications Compliance Feature Wise	Accepted	Incomplete Offer (Sr. No. 1 Part B,C,E,F not quoted. Neurosurgery set not quoted.	Accepted
05	Technical Eligibility of Product	Accepted	Not Accepted	Accepted
06	Technical Eligibility of Firm	Accepted	Accepted	Accepted
	Technical Status	Responsive	Non - Responsive	Responsive

Ahmed Maqsood Electromedical Engineer Muhammad Umair Kabeer Bio Medical Engineer Muhammad Hafeez Purchase Officer / Coordinator Dr. Khalid Hameed Zaighum DMS (Purchase / Admin)

Dr. M. Aslam Khan Assistant Prof. ofPlastic & Reconstructive Surgery Dr. Saeed Ahmad Assistant Prof. of Maxillofacial Surgery

Dr. Imran Hashim Assistant Prof. of Plastic Surgery Dr. Mumtaz Hussain Assistant Prof. of Paediatric Orthopedic Surgery

Dr. M. Nadeem Malik Associate Prof. of Paeds Neurosurgery Dr. Seema Qayyum Prof. of Paeds Ophthalmology



The Children's Hospital & The Institute of Child Health Ferozpur Road, Lahore Phone No. +92 42 99230901 – 23

Technical Evaluation

Package No. / Bid No. : 10110 / 28 / CH & ICH

Name of the Equipment:

Pneumatic Tourniquet Digital

Required Quantity:

As per ad	vertisement
-----------	-------------

	Part - I					
	KNOCKED DOWN CRITERIA - (BID EVALUATION)					
Evaluated	by Purchase Committee					
Sr. #	Evaluation Parameters	Gmed	FDS			
01	Complete Package / Tender (Mandatory)	Yes	Yes			
02	Original Receipt of Tender (Mandatory)	Yes	Yes			
03	Affidavit from Bidder	Yes	Yes			
04	Bid Security / CDR (Mandatory)	Yes	Yes			
05	Bid Validity (Mandatory)	Yes	Yes			
06	Delivery Period (Mandatory)	Yes	Yes			
	Remarks: Responsive Responsive					

	PART - II					
	KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)					
Evaluated	by Technical Committee	_				
Sr. #	Evaluation Parameters	Gmed	FDS			
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	Yes	Yes			
02	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes			
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal	Yes	Yes			
04	Bidder Past Performance	Satisfactory	Satisfactory			
05	Availability of relevant Tools and Testing / Calibration Equipment	Yes	Yes			
06	Warranty	Yes	Yes			
	Remarks:	Responsive	Responsive			

Package No. / Bid No. : 10110 / 28 / CH & ICH

Package No. / Bid No. : 10110 / 28 / CH & ICH

	PART - III				
		KNOCKED DOW	VN CRITERIA - (PRODCUT EVALUATION)		
	by Technical Committee				
Req. Item #	Required Item Name	M/s Gmed Pvt	M/s FDS		
28	Pneumatic Tourniquet Digital	,	,		
20	Brand	D&D (Dessillons) & Dutrillaux	Anetic Aid		
Sr. #	Quoted Model	G10 706	AT4 Touniquet System		
01	Country of Manufacturer	France	UK		
02	Country of Origin of Product	France	UK		
03	Compliance with defined Quality Standards	Compliance	Compliance		
04	Specifications Compliance Feature Wise	Accepted	Accepted		
05	Technical Eligibility of Product	Accepted	Accepted		
06	Technical Eligibility of Firm	Accepted	Accepted		
	Technical Status	Responsive	Responsive		

Ahmed Maqsood Electromedical Engineer Muhammad Umair Kabeer Bio Medical Engineer Muhammad Hafeez Purchase Officer / Coordinator Dr. Khalid Hameed Zaighum DMS (Purchase / Admin)

Dr. M. Aslam Khan Assistant Prof. ofPlastic & Reconstructive Surgery

Dr. Imran Hashim Assistant Prof. of Plastic Surgery Dr. M. Nadeem Malik Associate Prof. of Paeds Neurosurgery Dr. Farrukh Mahmood Prof. of Plastic & Reconstructive Surgery



The Children's Hospital & The Institute of Child Health Ferozpur Road, Lahore Phone No. +92 42 99230901 – 23

Technical Evaluation

Package No. / Bid No. : 10110 / 31 / CH & ICH

Name of the Equipment:

Pneumatic Dermatome

Required Quantity:

As per advertisement

	Part - I				
	KNOCKED DOWN CRITERIA - (BID EVALUATION				
Evaluatea	l by Purchase Committee				
Sr. #	Evaluation Parameters	Medi Urge	Gmed	Ali Gohar & Co.	
01	Complete Package / Tender (Mandatory)	Yes	Yes	Yes	
02	Original Receipt of Tender (Mandatory)	Yes	Yes	Yes	
03	Affidavit from Bidder	Yes	Yes	Yes	
04	Bid Security / CDR (Mandatory)	Yes	Yes	Yes	
05	Bid Validity (Mandatory)	Yes	Yes	Yes	
06	Delivery Period (Mandatory)	Yes	Yes	Yes	
	Remarks:	Responsive	Responsive	Responsive	

Package No. / Bid No. : 10110 / 31 / CH & ICH

	PART - II					
	KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)					
Evaluatea	l by Technical Committee		_			
Sr. #	Evaluation Parameters	Medi Urge	Gmed	Ali Gohar & Co.		
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	Yes	Yes	Yes		
02	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes	Yes		
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal	Yes	Yes	Yes		
04	Bidder Past Performance	Satisfactory	Satisfactory	Satisfactory		
05	Availability of relevant Tools and Testing / Calibration Equipment	Yes	Yes	Yes		
06	Warranty	Yes	Yes	Yes		
	Remarks:	Responsive	Responsive	Responsive		

Package No. / Bid No. : 10110 / 31 / CH & ICH

PART - III

KNOCKED DOWN CRITERIA - (PRODCUT EVALUATION)

	Evaluated by Technical Committee					
Req. Item #	Required Item Name	Medi Urge	Gmed	Ali Gohar & Co.		
31	Dermatome					
51	Brand	Integra Padgett	deSoutter Medical Ltd	Zimmer		
Sr. #	Quoted Model	Air Dermatome Model C	Derma Drive GD - 113 EPV - 220	Air Dermatom 00-8801-000-00		
01	Country of Manufacturer	USA	UK	USA		
02	Country of Origin of Product	USA	UK	USA		
03	Compliance with defined Quality Standards	Compliance	Compliance	Compliance		
04	Specifications Compliance Feature Wise	1) Bid No. 126163 / 22 / CH & ICH Demo Remarks 2) 497-JB & RSC, Lahore (Unsatisfactory Performance Letter)	 Cutting width issue not as per requirement - 04 Plate System not available 3 - Not controlled by hand Also rejected in bid no 126163 	Accepted		
05	Technical Eligibility of Product	Not Accepted	Not Accepted	Accepted		
06	Technical Eligibility of Firm	Not Accepted	Accepted	Accepted		
	Technical Status	Non - Responsive	Non - Responsive	Responsive		

Ahmed Maqsood Electromedical Engineer Muhammad Umair Kabeer Bio Medical Engineer Muhammad Hafeez Purchase Officer / Coordinator Dr. Khalid Hameed Zaighum DMS (Purchase / Admin)

Dr. M. Aslam Khan Assistant Prof. ofPlastic & Reconstructive Surgery

Dr. Imran Hashim Assistant Prof. of Plastic Surgery

Dr. M. Nadeem Malik Associate Prof. of Paeds Neurosurgery Dr. Farrukh Mahmood Prof. of Plastic & Reconstructive Surgery



The Children's Hospital & The Institute of Child Health Ferozpur Road, Lahore Phone No. +92 42 99230901 – 23

Technical Evaluation

Package No. / Bid No. : 10110 / 33 / CH & ICH

Name of the Equipment:

Micro Drill

As per advertisement

Required Quantity:

	Part - I					
	KNOCKED DOWN CRITERIA - (BID EVALUATIO					
Evaluated	by Purchase Committee					
Sr. #	Evaluation Parameters	FDS	KASBN			
01	Complete Package / Tender (Mandatory)	Yes	Yes			
02	Original Receipt of Tender (Mandatory)	Yes	Yes			
03	Affidavit from Bidder	Yes	Yes			
04	Bid Security / CDR (Mandatory)	Yes	Yes			
05	Bid Validity (Mandatory)	Yes	Yes			
06	Delivery Period (Mandatory)	Yes	Yes			
	Remarks:	Responsive	Responsive			

Package No. / Bid No. : 10110 / 33 / CH & ICH

	PART - II					
		KNOCKED DO	OWN CRITERIA - (VENDOR EVALUATION)			
Evaluated	Evaluated by Technical Committee					
Sr. #	Evaluation Parameters	FDS	KASBN			
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	Yes	Yes			
02	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes			
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal	Yes	Yes			
04	Bidder Past Performance	Satisfactory	Satisfactory			
05	Availability of relevant Tools and Testing / Calibration Equipment	Yes	Yes			
06	Warranty	Yes	Yes			
	Remarks:	Responsive	Responsive			

Package No. / Bid No. : 10110 / 33 / CH & ICH

PART - III

KNOCKED DOWN CRITERIA - (PRODCUT EVALUATION)

Req. Item #	Required Item Name	FDS	KASBN
33	Micro Drill		
33	Brand	Conmed	NOUVAG AG
Sr. #	Quoted Model	D 4000	High Surge 30 & Vacuson 18
01	Country of Manufacturer	USA	Switzerland
02	Country of Origin of Product	USA	Switzerland
03	Compliance with defined Quality Standards	Compliance	Compliance
04	Specifications Compliance Feature Wise	Accepted	Accepted
05	Technical Eligibility of Product	Accepted	Accepted
06	Technical Eligibility of Firm	Accepted	Accepted
	Technical Status	Responsive	Responsive

Ahmed Maqsood Electromedical Engineer Muhammad Umair Kabeer Bio Medical Engineer Muhammad Hafeez Purchase Officer / Coordinator Dr. Khalid Hameed Zaighum DMS (Purchase / Admin)

Dr. M. Aslam Khan Assistant Prof. ofPlastic & Reconstructive Surgery

Dr. Imran Hashim Assistant Prof. of Plastic Surgery Dr. M. Nadeem Malik Associate Prof. of Paeds Neurosurgery Dr. Farrukh Mahmood Prof. of Plastic & Reconstructive Surgery



The Children's Hospital & The Institute of Child Health Ferozpur Road, Lahore

Phone No. +92 42 99230901 – 23

Technical Evaluation

Package No. / Bid No. : 10110 / 34 / CH & ICH

Name of the Equipment:

Oscillating Saw (Micro)

Required Quantity:

As per advertisement

		Part - I				
		KNOCKE	DOWN CRITERIA - (BID EVALUATION)			
Evaluated	Evaluated by Purchase Committee					
Sr. #	Evaluation Parameters	FDS	KASBN			
01	Complete Package / Tender (Mandatory)	Yes	Yes			
02	Original Receipt of Tender (Mandatory)	Yes	Yes			
03	Affidavit from Bidder	Yes	Yes			
04	Bid Security / CDR (Mandatory)	Yes	Yes			
05	Bid Validity (Mandatory)	Yes	Yes			
06	Delivery Period (Mandatory)	Yes	Yes			
	Remarks:	Responsive	Responsive			

Package No. / Bid No. : 10110 / 34 / CH & ICH

	PART - II					
	KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)					
valuate	aluated by Technical Committee					
Sr. #	Evaluation Parameters	FDS	KASBN			
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	Yes	Yes			
02	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes			
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal	Yes	Yes			
04	Bidder Past Performance	Satisfactory	Satisfactory			
05	Availability of relevant Tools and Testing / Calibration Equipment	Yes	Yes			
06	Warranty	Yes	Yes			
	Remarks:	Responsive	Responsive			

Package No. / Bid No. : 10110 / 34 / CH & ICH

PART - III

KNOCKED DOWN CRITERIA - (PRODCUT EVALUATION)

q. Item #	Required Item Name	FDS	KASBN
34	Oscillating Saw (Micro)		
	Brand	Conmed	NOUVAG AG
Sr. #	Quoted Model	D 4000	High Surge 30 & Vacuson 18
01	Country of Manufacturer	USA	Switzerland
02	Country of Origin of Product	USA	Switzerland
03	Compliance with defined Quality Standards	Compliance	Compliance
04	Specifications Compliance Feature Wise	Accepted	Accepted
05	Technical Eligibility of Product	Accepted	Accepted
06	Technical Eligibility of Firm	Accepted	Accepted
	Technical Status	Responsive	Responsive

Ahmed Maqsood Electromedical Engineer Muhammad Umair Kabeer Bio Medical Engineer Muhammad Hafeez Purchase Officer / Coordinator Dr. Khalid Hameed Zaighum DMS (Purchase / Admin)

Dr. M. Aslam Khan Assistant Prof. ofPlastic & Reconstructive Surgery

Dr. Imran Hashim Assistant Prof. of Plastic Surgery

Dr. M. Nadeem Malik Associate Prof. of Paeds Neurosurgery Dr. Farrukh Mahmood Prof. of Plastic & Reconstructive Surgery



The Children's Hospital & The Institute of Child Health Ferozpur Road, Lahore Phone No. +92 42 99230901 – 23

Technical Evaluation

Package No. / Bid No. : 10110 / 41/ CH & ICH

Name of the Equipment:

Wound Clouser System VAC

Required Quantity:

	Part - I				
	KNOCKED DOWN CRITERIA - (BID EVALUATION)				
Evaluated	by Purchase Committee				
Sr. #	Evaluation Parameters	Clinical Life Inc	Dymedix Systems	Medi Urge	
01	Complete Package / Tender (Mandatory)	Yes	No	Yes	
02	Original Receipt of Tender (Mandatory)	Yes	Yes	Yes	
03	Affidavit from Bidder	Yes	Yes	Yes	
04	Bid Security / CDR (Mandatory)	Yes	Yes	Yes	
05	Bid Validity (Mandatory)	Yes	Yes	Yes	
06	Delivery Period (Mandatory)	Yes	Yes	Yes	
	Remarks: Responsive Non - Responsive Responsive				

As per advertisement

Package No. / Bid No. : 10110 / 41 / CH & ICH

	PART - II				
	KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)				
Evaluated	by Technical Committee				
Sr. #	Evaluation Parameters	Clinical Life Inc	Dymedix Systems	Medi Urge	
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	Yes	No	Yes	
02	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes	Yes	
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal	Yes	No	Yes	
04	Bidder Past Performance	Satisfactory	No	Satisfactory	
05	Availability of relevant Tools and Testing / Calibration Equipment	Yes	Yes	Yes	
06	Warranty	Yes	Νο	Yes	
	Remarks:	Responsive	Non - Responsive	Responsive	

Package No. / Bid No. : 10110 / 41 / CH & ICH

PART - III

Req. Item #	Required Item Name	Clinical Life Inc	Dymedix Systems	Medi Urge
41	Wound Clouser System VAC		Dynical systems	mear orge
	Brand	Talley Group	Not Mentioned	Simex Medizin Tec
Sr. #	Quoted Model	Venturi Avanti	extriCARE 3600	Simex 301
01	Country of Manufacturer	UK	Not Mentioned	Germany
02	Country of Origin of Product	UK	Not Mentioned	Germany
03	Compliance with defined Quality Standards	Compliance	No	Compliance
04	Specifications Compliance Feature Wise	Suction unit needs change for every patient. Not feasible	 Provided product warranty is 1 year instead of 3 years 2) Product brand not mentioned No Authorization, CE / FDA Certificate of the product is attached No evidence of previous reference attached & Poor past performance in CH & ICH, Lahore 	Accepted
05	Technical Eligibility of Product	Not Accepted	Not Accepted	Accepted
06	Technical Eligibility of Firm	Accepted	Not Accepted	Accepted
	Technical Status	Non - Responsive	Non - Responsive	Responsive

Ahmed Maqsood Electromedical Engineer Muhammad Umair Kabeer Bio Medical Engineer Muhammad Hafeez Purchase Officer / Coordinator Dr. Khalid Hameed Zaighum DMS (Purchase / Admin)

Dr. M. Aslam Khan Assistant Prof. ofPlastic & Reconstructive Surgery Dr. Imran Hashim Assistant Prof. of Plastic Surgery

Dr. M. Nadeem Malik Associate Prof. of Paeds Neurosurgery Dr. Farrukh Mahmood Prof. of Plastic & Reconstructive Surgery



The Children's Hospital & The Institute of Child Health Ferozpur Road, Lahore

Phone No. +92 42 99230901 – 23

Technical Evaluation

Package No. / Bid No. : 10110 / 42 / CH & ICH

Name of the Equipment:

Cardiac Monitor (Invasive)

Required Quantity:

	Part - I				
	KNOCKED DOWN CRITERIA - (BID EVALUATION				
Evaluated	by Purchase Committee				
Sr. #	Evaluation Parameters	Al Kareem	Vital Care	Medequips	
01	Complete Package / Tender (Mandatory)	Yes	Yes	Yes	
02	Original Receipt of Tender (Mandatory)	Yes	Yes	Yes	
03	Affidavit from Bidder	Yes	Yes	Yes	
04	Bid Security / CDR (Mandatory)	Yes	Yes	Yes	
05	Bid Validity (Mandatory)	Yes	Yes	Yes	
06	Delivery Period (Mandatory)	Yes	Yes	Yes	
	Remarks: Responsive Responsive Responsive				

As per advertisement

Package No. / Bid No. : 10110 / 42 / CH & ICH

	PART - II				
	KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)				
Evaluated	valuated by Technical Committee				
Sr. #	Evaluation Parameters	Al Kareem	Vital Care	Medequips	
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	Yes	Yes	Yes	
02	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes	Yes	
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal	Yes	Yes	Yes	
04	Bidder Past Performance	Satisfactory	Satisfactory	Satisfactory	
05	Availability of relevant Tools and Testing / Calibration Equipment	Yes	Yes	Yes	
06	Warranty	Yes	Yes	Yes	
	Remarks:	Responsive	Responsive	Responsive	

Package No. / Bid No. : 10110 / 42 / CH & ICH

	PART - III KNOCKED DOWN CRITERIA - (PRODCUT EVALUATION)				
Evaluated	by Technical Committee				
Req. Item #	Required Item Name	Al Kareem	Vital Care	Medequips	
42	Cardiac Monitor (Invasive)				
	Brand	Veiele	Heyer Medical	Nihon Kohden	
Sr. #	Quoted Model	SE 1500	Scalis - 15	BSM - 6701K	
01	Country of Manufacturer	Sweden	Germany	Japan	
02	Country of Origin of Product	Sweden	Germany	Japan	
03	Compliance with defined Quality Standards	Compliance	Compliance	Compliance	
04	Specifications Compliance Feature Wise	Accepted	Accepted	Accepted	
05	Technical Eligibility of Product	Accepted	Accepted	Accepted	
06	Technical Eligibility of Firm	Accepted	Accepted	Accepted	
	Technical Status	Responsive	Responsive	Responsive	

Ahmed Maqsood Electromedical Engineer Muhammad Umair Kabeer Bio Medical Engineer Muhammad Hafeez Purchase Officer / Coordinator Dr. Khalid Hameed Zaighum DMS (Purchase / Admin)

Dr. M. Aslam Khan Assistant Prof. ofPlastic & Reconstructive Surgery

Dr. Imran Hashim Assistant Prof. of Plastic Surgery Dr. M. Nadeem Malik Associate Prof. of Paeds Neurosurgery Dr. Farrukh Mahmood Prof. of Plastic & Reconstructive Surgery



The Children's Hospital & The Institute of Child Health Ferozpur Road, Lahore Phone No. +92 42 99230901 – 23

Technical Evaluation

Package No. / Bid No. : 10110 / 46 / CH & ICH

Name of the Equipment:

Patient Bed with Drawer & Bedside Bench

As per advertisement

Required Quantity:

	Part - I						
	KNOCKED DOWN CRITERIA - (BID EVALUATION)						
Evaluated	Evaluated by Purchase Committee						
Sr. #	Evaluation Parameters	Orient Medical	Noorani Surgical				
01	Complete Package / Tender (Mandatory)	Yes	Yes				
02	Original Receipt of Tender (Mandatory)	Yes	Yes				
03	Affidavit from Bidder	Yes	Yes				
04	Bid Security / CDR (Mandatory)	Yes	Yes				
05	Bid Validity (Mandatory)	Yes	Yes				
06	Delivery Period (Mandatory)	Yes	Yes				
	Remarks:	Responsive	Responsive				

PART - II **KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)** *Evaluated by Technical Committee* Sr. # **Evaluation Parameters Orient Medical** Noorani Surgical 01 Exclusive Authorization / Sole Agent Certificate by the Manufacturer Yes Yes Technical & Engineering capability 02 Yes Yes (As defined for the specific Bid in specifications) Up-to-Date After Sales Service Certificate from Manufacturer / Principal 03 Yes Yes Bidder Past Performance Satisfactory Satisfactory 04 05 Availability of relevant Tools and Testing / Calibration Equipment Yes Yes Warranty 06 Yes Yes **Remarks:** Responsive Responsive

Package No. / Bid No. : 10110 / 46 / CH & ICH

Package No. / Bid No. : 10110 / 46 / CH & ICH

	PART - III					
	KNOCKED DOWN CRITERIA - (PRODCUT EVALUATI					
	by Technical Committee					
Req. Item #	Required Item Name	Orient Medical	Noorani Surgical			
46	Patient Bed with Drawer & Bedside Bench		5			
40	Brand	Local	Local			
Sr. #	Quoted Model	N/A	N/A			
01	Country of Manufacturer	Pakistan	Pakistan			
02	Country of Origin of Product	Pakistan	Pakistan			
03	Compliance with defined Quality Standards	N/A	N/A			
04	Specifications Compliance Feature Wise	Accepted	Accepted			
05	Technical Eligibility of Product	Accepted	Accepted			
06	Technical Eligibility of Firm	Accepted	Accepted			
	Technical Status	Responsive	Responsive			

Ahmed Maqsood Electromedical Engineer Muhammad Umair Kabeer Bio Medical Engineer Muhammad Hafeez Purchase Officer / Coordinator

Dr. Khalid Hameed Zaighum DMS (Purchase / Admin)

Dr. Zahid Dogar SMO, Ophthalmology Dr. Asma Mushtaq Senior Registrar Paediatric Ophthalmology

Dr. Seema Qayyum Prof. of Paeds Ophthalmology

The Children's Hospital & The Institute of Child Health Ferozpur Road, Lahore

Phone No. +92 42 99230901 - 23

Technical Evaluation

Package No. / Bid No. : 10110 / 51/ CH & ICH

Name of the Equipment:

Required Quantity:

	Part - I KNOCKED DOWN CRITERIA - (BID EVALUATION)					
Evaluated	d by Purchase Committee					
Sr. #	Evaluation Parameters	Radiant Medical	Latif Brothers	Jasani Scientific		
01	Complete Package / Tender (Mandatory)	Yes	Yes	Yes		
02	Original Receipt of Tender (Mandatory)	Yes	Yes	Yes		
03	Affidavit from Bidder	Yes	Yes	Yes		
04	Bid Security / CDR (Mandatory)	Yes	Yes	Yes		
05	Bid Validity (Mandatory)	Yes	Yes	Yes		
06	Delivery Period (Mandatory)	Yes	Yes	Yes		
	Remarks:ResponsiveResponsiveResponsiveResponsive					

Slit Lamp Microscope

1



	PART - II KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)						
Evaluated	Evaluated by Technical Committee						
Sr. #	Evaluation Parameters	Radiant Medical	Latif Brothers	Jasani Scientific			
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	Yes	Yes	Yes			
02	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes	Yes			
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal	Yes	Yes	Yes			
04	Bidder Past Performance	Satisfactory	Satisfactory	Satisfactory			
05	Availability of relevant Tools and Testing / Calibration Equipment	Yes	Yes	Yes			
06	Warranty	Yes	Yes	Yes			
	Remarks: Responsive Responsive Responsive						

Package No. / Bid No. : 10110 / 51/ CH & ICH

Package I	No. / Bid No. : 10110 / 51/ CH & ICH						
		PART - III					
KNOCKED DOWN CRITERIA - (PRODCUT EVALUATION)							
valuated	l by Technical Committee						
Req. Item #	Required Item Name						
	Slit Lamp Microscope	Radiant Medical	Latif Brothers	Jasani Scientific			
51							
	Brand	Inami & Co	Takagi	CSO			
Sr. #	Quoted Model	L-0185 LED	30-GI	SL 9900			
01	Country of Manufacturer	Japan	Japan	Italy			
02	Country of Origin of Product	Japan	Japan	Italy			
03	Compliance with defined Quality Standards	Compliance	Compliance	Compliance			
04	Specifications Compliance Feature Wise	Accepted	Accepted	Accepted			
05	Technical Eligibility of Product	Accepted	Accepted	Accepted			
06	Technical Eligibility of Firm	Accepted	Accepted	Accepted			
	Technical Status	Responsive	Responsive	Responsive			

Ahmed Maqsood Electromedical Engineer

Dr. Khalid Hameed Zaighum DMS (Purchase / Admin) Muhammad Umair Kabeer Bio Medical Engineer Muhammad Hafeez Purchase Officer / Coordinator

Dr. Zahid Dogar SMO, Ophthalmology Dr. Asma Mushtaq Senior Registrar Paediatric Ophthalmology

Dr. Seema Qayyum Prof. of Paeds Ophthalmology



The Children's Hospital & The Institute of Child Health *Ferozpur Road, Lahore*

Phone No. +92 42 99230901 – 23

Technical Evaluation

Package No. / Bid No. : 10110 / 52/ CH & ICH

Name of the Equipment:

Slit Lamp Bio Microscope With digital Imaging

Required Quantity:

As per advertisement

	Part - I						
	KNOCKED DOWN CRITERIA - (BID EVALUATION) Evaluated by Purchase Committee						
Evaluated by							
Sr. #	Evaluation Parameters	Radiant Medical	Latif Brothers	Haji S. Ameer ud Din	Jasani Scientific		
01	Complete Package / Tender (Mandatory)	Yes	Yes	Yes	Yes		
02	Original Receipt of Tender (Mandatory)	Yes	Yes	Yes	Yes		
03	Affidavit from Bidder	Yes	Yes	Yes	Yes		
04	Bid Security / CDR (Mandatory)	Yes	Yes	Yes	Yes		
05	Bid Validity (Mandatory)	Yes	Yes	Yes	Yes		
06	Delivery Period (Mandatory)	Yes	Yes	Yes	Yes		
	Remarks: Responsive Responsive Responsive Responsive						

Package No. / Bid No. : 10110 / 52 / CH & ICH

	PART - II						
	KNOCKED DOWN CRITERIA - (VENDOR EVALUATION) Evaluated by Technical Committee						
Evaluated by							
Sr. #	Evaluation Parameters	Radiant Medical	Latif Brothers	Haji S. Ameer ud Din	Jasani Scientific		
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	Yes	Yes	Yes	Yes		
02	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes	Yes	Yes		
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal	Yes	Yes	Yes	Yes		
04	Bidder Past Performance	Satisfactory	Satisfactory	Satisfactory	Satisfactory		
05	Availability of relevant Tools and Testing / Calibration Equipment	Yes	Yes	Yes	Yes		
06	Warranty	Yes	Yes	Yes	Yes		
	Remarks:	Responsive	Responsive	Responsive	Responsive		

Package No. / Bid No. : 10110 / 52 / CH & ICH

PART - III

KNOCKED DOWN CRITERIA - (PRODCUT EVALUATION)						
Evaluated by T Req. Item #	Technical Committee Required Item Name	Radiant Medical	Latif Brothers	Haji S. Ameer ud	Jasani	
	Slit Lamp Bio Microscope With digital Imaging	Kaulant Meulcai	Latii Brotners	Din	Scientific	
52	Brand	Right Mfg. Co.	Takagi	Topcon	CSO	
Sr. #	Quoted Model	MW 50D	Slit Lamp Microscope 700 GL	SL-D701	SL 9900	
01	Country of Manufacturer	Japan	Japan	Japan	Italy	
02	Country of Origin of Product	Japan	Japan	Japan	Italy	
03	Compliance with defined Quality Standards	Compliance	Compliance	Compliance	Compliance	
04	Specifications Compliance Feature Wise	End user not satisfied, Demo Remarks: it required additional USB dongle to run the software which is not user friendly	End user not satisfied, Demo Remarks: it required additional USB dongle to run the software which is not user friendly	Accepted	Accepted	
05	Technical Eligibility of Product	Not Accepted	Not Accepted	Accepted	Accepted	
06	Technical Eligibility of Firm	Accepted	Accepted	Accepted	Accepted	
	Technical Status	Non - Responsive	Non - Responsive	Responsive	Responsive	

Ahmed Maqsood Electromedical Engineer Muhammad Umair Kabeer Bio Medical Engineer Muhammad Hafeez Purchase Officer / Coordinator Dr. Khalid Hameed Zaighum DMS (Purchase / Admin)

Dr. Zahid Dogar SMO, Ophthalmology

Dr. Asma Mushtaq Senior Registrar Paediatric Ophthalmology Dr. Seema Qayyum Prof. of Paeds Ophthalmology