



**The Children's Hospital,
University of Child Health Sciences**
Ferozpur Road, Lahore Phone # (92) (42) 99230901-23 Fax # 99231560



No. P.C./ 13343 /CH&UCHS

Dated 11-03- 2024


To

Mr. Abdul Wahab
I/C Programmer (I.T.) Health Department,
Civil Secretariat, Lahore.

Subject: - **REQUEST FOR UPLOADING THE MINUTES OF TECHNICAL ADVISORY COMMITTEE (TAC) MEETING REGARDING THE PROCUREMENT OF ELECTRO MEDICAL EQUIPMENT, PLANT & MACHINERY AND IT EQUIPMENT FOR ESTABLISHMENT OF PAEDIATRIC BURN CENTRE & UP-GRADATION OF PEADIATRIC ENT WITH SPECIAL REFERENCE OF COCHLEAR IMPLANT & PEADIATRIC UROLOGY WITH SPECIAL REFERENCE TO NON-INVASIVE TREATMENT**

Please find enclosed herewith Minutes of Technical Advisory Committee (TAC) meeting regarding the **Procurement of Electro_Medical Equipment, Plant & Machinery and IT Equipment For Establishment of Paediatric Burn Centre & Up-Gradation of Peadiatric ENT With Special Reference of Cochlear Implant & Peadiatric Urology With Special Reference To Non-Invasive Treatment** for The Children's Hospital & University of Child Health Sciences, Lahore for the year 2023-2024 needs to be uploaded on the Health Department website.

Your early response will be highly appreciated.



PROF. DR. TIPU SULTAN
M.D.S., F.C.P.S., M.S., (U.K.), (F.R.C.P.) (LONDON)
Prof. of Paediatric Neurology
Medical Director

MINUTES OF TECHNICAL ADVISORY COMMITTEE (TAC) MEETING HELD ON 09-03-2024 AT 10:00 A.M. IN THE CONFERENCE ROOM OF THE ADMIN BLOCK THE CHILDREN'S HOSPITAL LAHORE REGARDING THE PROCUREMENT OF ELECTRO MEDICAL EQUIPMENT, PLANT & MACHINERY AND IT EQUIPMENT FOR UP-GRADATION OF PEADIATRIC ENT WITH SPECIAL REFERENCE OF COCHLEAR IMPLANT & PEADIATRIC UROLOGY WITH SPECIAL REFERENCE TO NON-INVASIVE TREATMENT OF CHILDREN'S HOSPITAL & UNIVERSITY OF CHILD HEALTH SCIENCES, LAHORE

In compliance to letter No. P.C./11393-406/CH&UCHS dated 29-02-2024, Technical Advisory Committee (TAC) meeting held on 09-03-2024 at 10:00 a.m. in conference room of the admin block of The Children's Hospital, University of Child Health Sciences, Lahore.

Meeting started with the name of Allah by recitation of the Holy Quran. Bio medical engineers/technical officers briefed the committee that the Procurement of Electro Medical Equipment, Plant & Machinery and IT Equipment for Up-gradation of Peadiatric ENT with Special Reference of Cochlear Implant & Peadiatric Urology with Special Reference to Non-Invasive Treatment to be purchased for the Children's Hospital, Lahore was advertised through press as well as on PPRA website. All the committee members, End-users and co-opted members scrutinized / evaluated and discussed all the quoted bids in detail and give the following recommendations for finalization of the technical advisory committee (TAC) reports. Following decision was made unanimously by all the members of TAC committee, co-opted members & end users.


Bid Sr. No.	Item Description	Remarks by TAC Committee
2.	Mastoidectomy Set	<ul style="list-style-type: none"> Three firms participated in the bid i.e. M/s Vertex Medical (Pvt) Ltd, M/s Mediland Pakistan (Pvt Ltd)& M/s FDS (Pvt) Ltd. Two Firms i.e. M/s Vertex Medical (Pvt) Ltd& M/s FDS (Pvt) Ltd declared technically responsive. M/s Mediland Pakistan (Pvt Ltd) declared technically non-responsive due to No bidders Past Performance for quoted product in Teaching Hospitals.
3.	Middle Ear Micro Ear Surgery Set	<ul style="list-style-type: none"> Three firms participated in the bid i.e. M/s Vertex Medical (Pvt) Ltd, M/s Mediland Pakistan (Pvt Ltd)& M/s FDS (Pvt) Ltd. Two Firms i.e. M/s Vertex Medical (Pvt) Ltd& M/s FDS (Pvt) Ltd declared technically responsive. M/s Mediland Pakistan (Pvt Ltd) declared technically non-responsive due to No bidders Past Performance for quoted product in Teaching Hospitals.
4.	Middle Ear Surgery Tympanoplasty And Stapedotomy Set	<ul style="list-style-type: none"> Three firms participated in the bid i.e. M/s Vertex Medical (Pvt) Ltd, M/s Mediland Pakistan (Pvt Ltd)& M/s FDS (Pvt) Ltd. Two Firms i.e. M/s Vertex Medical (Pvt) Ltd& M/s FDS (Pvt) Ltd declared technically responsive. M/s Mediland Pakistan (Pvt Ltd) declared technically non-responsive due to No bidders Past Performance for quoted product in Teaching Hospitals.
5.	Myringotomy Set	<ul style="list-style-type: none"> Only one Firm participated in the bid i.e. M/s Vertex Medical (Pvt) Ltd and declared technically responsive.



 M. Khan
 12-3-2024

MINUTES OF TECHNICAL ADVISORY COMMITTEE (TAC) MEETING HELD ON 09-03-2024 AT 10:00 A.M. IN THE CONFERENCE ROOM OF THE ADMIN BLOCK THE CHILDREN'S HOSPITAL LAHORE REGARDING THE PROCUREMENT OF ELECTRO MEDICAL EQUIPMENT, PLANT & MACHINERY AND IT EQUIPMENT FOR UP-GRADATION OF PEADIATRIC ENT WITH SPECIAL REFERENCE OF COCHLEAR IMPLANT & PEADIATRIC UROLOGY WITH SPECIAL REFERENCE TO NON-INVASIVE TREATMENT OF CHILDREN'S HOSPITAL & UNIVERSITY OF CHILD HEALTH SCIENCES, LAHORE

Bid Sr. No.	Item Description	Remarks by TAC Committee
6.	ENT unit combined	<ul style="list-style-type: none"> Only one Firm participated in the bid i.e. M/s Mediland Pakistan (Pvt Ltd) and declared technically responsive.
7.	Mastoid Drill System	<ul style="list-style-type: none"> Five firms participated in the bid i.e. M/s Allmed Solution, M/s Popular International, M/s KASBN International, M/s G Med (Pvt) Ltd & M/s FDS (Pvt) Ltd. Four Firms i.e. M/s Allmed Solution, M/s Popular International, M/s KASBN International & M/s FDS (Pvt) Ltd. declared technically responsive. M/s G Med (Pvt Ltd) declared technically non-responsive due to noncompliance on quality certificate i.e. the certificate attached in bid is expire .
8.	Cochlear Implantation Set	<ul style="list-style-type: none"> Only one Firm participated in the bid i.e. M/s Allmed Solution and declared technically responsive.
10.	Operating Microscope for ENT	<ul style="list-style-type: none"> Four firms participated in the bid i.e. M/s Jasani Scientifics, M/s Eastern Optical Co, M/s Latif Brothers, & M/s Medical Equipment & System. Two Firms i.e. M/s Jasani Scientifics & M/s Latif Brothers declared technically responsive. M/s Eastern Optical Co & M/s Medical Equipment & System declared technically non-responsive on following grounds. No bidders Past Performance for quoted product in Teaching Hospitals. Bid Authorization as both quoted the same product and model
11.	Operating head lights	<ul style="list-style-type: none"> Two firms participated in the bid i.e. M/s Allmed Solution & M/s Mediland Pakistan (Pvt) Ltd both Firms declared technically responsive.
12.	Endoscope cameras system (High Definition)	<ul style="list-style-type: none"> Only one Firm participated in the bid i.e. M/s Allmed Solution and declared technically responsive.
13.	Coblation equipment	<ul style="list-style-type: none"> Only one Firm participated in the bid i.e. M/s Popular International (Pvt) Ltd and declared technically responsive.
14.	Oto Endoscope	<ul style="list-style-type: none"> Only one Firm participated in the bid i.e. M/s Allmed Solution and declared technically responsive.
15.	Pulse Oximeter	<ul style="list-style-type: none"> Four firms participated in the bid i.e. M/s Eastern Medical Technologies, M/s Popular International (Pvt) Ltd, M/s Al Kareem Medical Technologies & M/s Hospicare System. Two Firms i.e. M/s Hospicare System & M/s Popular International (Pvt) Ltd declared technically responsive. M/s Eastern Medical Technologies & M/s Al Kareem Medical Technologies declared technically non-responsive due to No bidders Past Performance for quoted product in Teaching Hospitals.



 M. N. Nadeem
 12-3-2024

MINUTES OF TECHNICAL ADVISORY COMMITTEE (TAC) MEETING HELD ON 09-03-2024 AT 10:00 A.M. IN THE CONFERENCE ROOM OF THE ADMIN BLOCK THE CHILDREN'S HOSPITAL LAHORE REGARDING THE PROCUREMENT OF ELECTRO MEDICAL EQUIPMENT, PLANT & MACHINERY AND IT EQUIPMENT FOR UP-GRADATION OF PEADIATRIC ENT WITH SPECIAL REFERENCE OF COCHLEAR IMPLANT & PEADIATRIC UROLOGY WITH SPECIAL REFERENCE TO NON-INVASIVE TREATMENT OF CHILDREN'S HOSPITAL & UNIVERSITY OF CHILD HEALTH SCIENCES, LAHORE


Bid Sr. No.	Item Description	Remarks by TAC Committee
16.	Rigid bronchoscope set with Optical Forceps	<ul style="list-style-type: none"> Only one Firm participated in the bid i.e. M/s Allmed Solution and declared technically responsive.
17.	Laryngoscope set	<ul style="list-style-type: none"> Only one Firm participated in the bid i.e. M/s Allmed Solution and declared technically responsive.
18.	Regid Esophagoscope set with Optical Forceps	<ul style="list-style-type: none"> Only one Firm participated in the bid i.e. M/s Allmed Solution and declared technically responsive.
19.	Bronchial endoscope	<ul style="list-style-type: none"> Only one Firm participated in the bid i.e. M/s Allmed Solution and declared technically responsive.
20.	Micro Laryngeal Surgery Set	<ul style="list-style-type: none"> Only one Firm participated in the bid i.e. M/s Allmed Solution and declared technically responsive.
21.	Tracheostomy Set	<ul style="list-style-type: none"> Two firms participated in the bid i.e. M/s Vertex Medical (Pvt) Ltd & M/s FDS (Pvt) Ltd and declared technically responsive.
23.	Fess equipment Functional endoscopic sinus surgery	<ul style="list-style-type: none"> Only one Firm participated in the bid i.e. M/s Allmed Solution and declared technically responsive.
24.	Flexible endoscope equipment	<ul style="list-style-type: none"> Two firms participated in the bid i.e. M/s Allmed Solution & M/s Mediland Pakistan (Pvt) Ltd and declared technically responsive.
25.	ENT treatment units	<ul style="list-style-type: none"> Only one Firm participated in the bid i.e. M/s Mediland Pakistan (Pvt) Ltd and declared technically responsive.
26.	Adeno- Tonsillectomy Set	<ul style="list-style-type: none"> Only one Firm participated in the bid i.e. M/s Vertex Medical (Pvt) Ltd and declared technically responsive.
27.	Nasal Surgery Set	<ul style="list-style-type: none"> Two firms participated in the bid i.e. M/s Vertex Medical (Pvt) Ltd & M/s FDS (Pvt) Ltd and declared technically responsive.
28.	Shaver Debrider with components	<ul style="list-style-type: none"> Four firms participated in the bid i.e. M/s M/s. Allmed Solution, M/s Popular International (Pvt) Ltd, M/s KASBN International & M/s FDS (Pvt) Ltd. All Firms declared technically responsive.




 M. Zahid
 M. Naveed
 12-3-2024

MINUTES OF TECHNICAL ADVISORY COMMITTEE (TAC) MEETING HELD ON 09-03-2024 AT 10:00 A.M. IN THE CONFERENCE ROOM OF THE ADMIN BLOCK THE CHILDREN'S HOSPITAL LAHORE REGARDING THE PROCUREMENT OF ELECTRO MEDICAL EQUIPMENT, PLANT & MACHINERY AND IT EQUIPMENT FOR UP-GRADATION OF PEADIATRIC ENT WITH SPECIAL REFERENCE OF COCHLEAR IMPLANT & PEADIATRIC UROLOGY WITH SPECIAL REFERENCE TO NON-INVASIVE TREATMENT OF CHILDREN'S HOSPITAL & UNIVERSITY OF CHILD HEALTH SCIENCES, LAHORE

Bid Sr. No.	Item Description	Remarks by TAC Committee
29.	Stroboscope System	<ul style="list-style-type: none"> Two firms participated in the bid i.e. M/s Allmed Solution & M/s KASBN International. Only one Firm i.e. M/s Allmed Solution declared technically responsive. M/s KASBN International declared technically non-responsive due to No bidders Past Performance for quoted product in Teaching Hospitals.
30.	Syringe Pump with Docking station	<ul style="list-style-type: none"> Four firms participated in the bid i.e. M/s Medical Equipment & System, M/s Al Kareem Medical Technologies, M/s Hospicare System & M/s Total Technologies (Pvt) Ltd. Three Firms i.e. M/s Medical Equipment & System, M/s Hospicare System & M/s Total Technologies (Pvt) Ltd declared technically responsive. M/s Al Kareem Medical Technologies declared technically non-responsive due to No bidders Past Performance for quoted product in Teaching Hospitals.
33-37	Cystoscope Set for Neonates	<ul style="list-style-type: none"> Two firms participated in the bid i.e. M/s Allmed Solution & M/s Verizon. Both Firms declared technically responsive.
	Cystoscope for Infants and Pediatrics	
	Resectoscope for Neonates	
	Resectoscope for Infants and Pediatrics	
	Optical Urethrotome for Pediatrics	
38.	Cystoscope & Resectoscope for Adolescent	<ul style="list-style-type: none"> Only one Firm participated in the bid i.e. M/s Verizon and declared technically responsive.
39.	Irrigation pump with pressure monitor system for endoscopy	<ul style="list-style-type: none"> Two firms participated in the bid i.e. M/s Allmed Solution & M/s Verizon. Both Firms declared technically responsive.
40.	Pediatric Mini Nephroscope set	<ul style="list-style-type: none"> Two firms participated in the bid i.e. M/s Allmed Solution & M/s Verizon. Both Firms declared technically responsive.
41.	Pediatric Micro PCNL Set	<ul style="list-style-type: none"> Only one Firm participated in the bid i.e. M/s Allmed Solution and declared technically responsive.
42.	Pediatric Uretero-Renoscope Semi rigid URS 4.5-7.3 Fr, 50 - 60, Length 25-57cm, working channel 3-4 Fr or better	<ul style="list-style-type: none"> Two firms participated in the bid i.e. M/s Allmed Solution & M/s Verizon. Both Firms declared technically responsive.




 Mr. Aliwan
 12-3-2024

MINUTES OF TECHNICAL ADVISORY COMMITTEE (TAC) MEETING HELD ON 09-03-2024 AT 10:00 A.M. IN THE CONFERENCE ROOM OF THE ADMIN BLOCK THE CHILDREN'S HOSPITAL LAHORE REGARDING THE PROCUREMENT OF ELECTRO MEDICAL EQUIPMENT, PLANT & MACHINERY AND IT EQUIPMENT FOR UP-GRADATION OF PEADIATRIC ENT WITH SPECIAL REFERENCE OF COCHLEAR IMPLANT & PEADIATRIC UROLOGY WITH SPECIAL REFERENCE TO NON-INVASIVE TREATMENT OF CHILDREN'S HOSPITAL & UNIVERSITY OF CHILD HEALTH SCIENCES, LAHORE


Bid Sr. No.	Item Description	Remarks by TAC Committee
43.	Pediatric Uretero-Renoscope Semirigid URS 6.5-9 Fr, 6 - 120, Length 25-57cm, working channel 3-4 Fr or better	<ul style="list-style-type: none"> Two firms participated in the bid i.e. M/s Allmed Solution & M/s Verizon. Both Firms declared technically responsive.
44.	Flexible digital ureterosopr 6-8 fr (URS)	Only one Firm participated in the bid i.e. M/s Ontech Corporation declared technically non responsive. Due to non-compliance on country of manufacturer.
45.	Pediatric Laparoscope Set / System (4-k) having 1-tower 2- Instruments	<ul style="list-style-type: none"> Three firms participated in the bid i.e. M/s Allmed Solution, M/s Verizon & M/s Vertex Medical (Pvt) Ltd. Two Firms i.e. M/s Allmed Solution, M/s Verizon declared technically responsive. M/s Vertex Medical (Pvt) Ltd declared technically non-responsive due to No bidders Past Performance for quoted product in Teaching Hospitals and non-compliance on specification.
46.	Advanced Bipolar diathermy	<ul style="list-style-type: none"> Three firms participated in the bid i.e. M/s Radiant Medical, M/s Vertex Medical (Pvt) Ltd & M/s Mediland Pakistan (Pvt) Ltd. Only one Firm i.e. M/s Vertex Medical (Pvt) Ltd declared technically responsive. M/s Radiant Medical & M/s Mediland Pakistan (Pvt) Ltd declared technically non-responsive due to No bidders Past Performance for quoted product in Teaching Hospitals.
47.	Mobile Ultrasound Machine with Doppler	<ul style="list-style-type: none"> Two firms participated in the bid i.e. M/s Vertex Medical (Pvt) Ltd & M/s Medical Equipment System. Both Firms declared technically responsive.
49.	Extra-Corporeal shock waves Lithotripter (ESWL) integrated type for Pediatric patient with General anesthesia set up as pr PVMS	<ul style="list-style-type: none"> Two firms participated in the bid i.e. M/s Allmed Solution & M/s Hospital Supply Corporation. Only one Firm i.e. M/s Allmed Solution declared technically responsive. M/s Hospital Supply Corporation declared technically non-responsive due to Non Compliance on technical specifications.
50.	Computer System Core i7 (Latest Generator) 16 GB Ram 1 TB Hard disk 19" LCD with UPS (APC650VA)	<ul style="list-style-type: none"> Only one Firm participated in the bid i.e. M/s PTCL. Said Firm declared technically responsive.
51.	Laser printer (Black & White)	<ul style="list-style-type: none"> Only one Firm participated in the bid i.e. M/s Apex Traders. Said Firm declared technically responsive.
52.	Multimedia Projector	<ul style="list-style-type: none"> Only one Firm participated in the bid i.e. M/s Apex Traders. Said Firm declared technically responsive.
53.	Laser printer with scanner and photocopier (Black & White)	<ul style="list-style-type: none"> Only one Firm participated in the bid i.e. M/s Apex Traders. Said Firm declared technically responsive.


MINUTES OF TECHNICAL ADVISORY COMMITTEE (TAC) MEETING HELD ON 09-03-2024 AT 10:00 A.M. IN THE CONFERENCE ROOM OF THE ADMIN BLOCK THE CHILDREN'S HOSPITAL LAHORE REGARDING THE PROCUREMENT OF ELECTRO MEDICAL EQUIPMENT, PLANT & MACHINERY AND IT EQUIPMENT FOR UP-GRADATION OF PEADIATRIC ENT WITH SPECIAL REFERENCE OF COCHLEAR IMPLANT & PEADIATRIC UROLOGY WITH SPECIAL REFERENCE TO NON-INVASIVE TREATMENT OF CHILDREN'S HOSPITAL & UNIVERSITY OF CHILD HEALTH SCIENCES, LAHORE


Bid Sr. No.	Item Description	Remarks by TAC Committee
54.	Internet networking	<ul style="list-style-type: none"> Two firms participated in the bid i.e. M/s Apex Traders & M/s PTCL. Both Firms declared technically responsive
55.	Digital Camera high definition	<ul style="list-style-type: none"> Only one Firm participated in the bid i.e. M/s Apex Traders. Said Firm declared technically non responsive.

Enduser (Concerned Department/ Co-opted Member)



Eng. Arian Haider
Bio Medical Engineer
Co-opted Member


Mazhar Iqbal Noor
Computer Programmer
IT Incharge


Dr. Ghulam Mujtaba Zafar
Associate Prof. Paediatric Urology



Dr. Imran Saeed
Prof. of Paediatric ENT


Technical Advisory Committee TAC



Eng. Shujaat Ali
Bio Medical Engineer/
Technical Officer


Muhammad Riaz Shahid
Incharge Pay Roll Section



Dr. Syed Ali Abu Zar
AMS (Coordination)



Dr. Faeza Hasnain
Chief Pharmacist



Dr. Muhammad Zeeshan Khan
Assistant Professor Paediatric
Medicine



Dr. Mumtaz Hussain
Assistant Professor Paediatric
Orthopedic Surgery



Dr. Muhammad Bilal Mirza
Associate Prof. of Pead
Surgery


Dr. Zia ur Rehman
Associate Prof. of Peads
Neurology



Dr. Nazish Saqlan
Associate Prof. of Pathology
Department


Dr. Sarfraz Ahmad
Associate Prof. Peadiatric PM&R


Dr. Farah Haroon
Assocaite Prof. of Neonatology


Dr. Uzma Kazmi
Professor of Paediatric
Cardiology


Dr. Syed Muhammad Javed Iqbal
Prof. of Peadiatric Medicine Unit-II


Dr. Junaid Rashid
Prof. of Paediatric Medicine / Chairman
Department of Paediatric Medicine

The Children's Hospital & University of Child Health Sciences Lahore

Technical Evaluation

Package No. / Bid No. PC/2020/2/CH & UCHS

Mostoidectmy Set

Name of the Equipment:

1

Required Quantity:

Part - I

KNOCKED DOWN CRITERIA - (BID EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	Vertex Medical (Pvt) ltd	Medilan pakistan (Pvt) Ltd	FDS (Pvt) Ltd
01	Complete Package / Tender (Mandatory)	Yes	Yes	Yes
02	Original Receipt of Tender (Mandatory)	Yes	Yes	Yes
03	Affidavit from Bidder	Yes	Yes	Yes
04	Bid Security CDR / Bank Guarantee (Mandatory)	Yes	Yes	Yes
05	Bid Validity (Mandatory)	Yes	Yes	Yes
06	Delivery Period (Mandatory)	Yes	Yes	Yes
Remarks:		Responsive	Responsive	Responsive

PART - II

KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	Vertex Medical (Pvt) ltd	Medilan pakistan (Pvt) Ltd	FDS (Pvt) Ltd
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	Yes	Yes	Yes
02	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes	Yes
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal	Yes	Yes	Yes
04	Bidder Past Performance for Quoted Product in Teaching Hospitals	Yes	No	Yes
05	Availability of relevant Tools and Testing / Calibration Equipment	Yes	Yes	Yes
06	Warranty	Yes	Yes	Yes
Remarks:		Responsive	Non-Responsive	Responsive


PART - III				
KNOCKED DOWN CRITERIA - (PRODCUT EVALUATION)				
Evaluated by (TAC) Committee				
2	Required Item Name	Vertex Medical (Pvt) Ltd	Medilan pakistan (Pvt) Ltd	FDS (Pvt) Ltd
		Mostoidectmy Set - 1		
	Brand	B.Braun	KLS Martin	Medicon
Sr. #	Quoted Model	various	Various	Various
01	Country of Manufacturer	Germany	Germany	Germany
02	Manufacturing site	Germany	Germany	Germany
03	Compliance with defined Quality Standards	Compliance		Compliance
04	Specifications Compliance Feature Wise	Compliance		Compliance
05	Technical Eligibility of Product	Accepted		Accepted
06	Technical Eligibility of Firm	Accepted		Accepted
Technical Status		Responsive	Non-Responsive	Responsive

Enduser (Concerned Department/ Co-opted Member)



Engr. Imran Haider
Bio Medical Engineer/
Co-opted Member

Dr. Imran Saeed
Prof. of Peadiatric ENT

Technical Advisory Committee (TAC)


Engr. Shujaat Ali
Bio Medical Engineer/
Technical Officer


Muhammad Riaz Shahid
Incharge Pay Roll Section



Dr. Syed Ali Abu Zar
AMS (Coordination)


Dr. Faeza Hasnain
Chief Pharmacist


Dr. Muhammad Zeeshan Khan
Assistant Professor Paediatric
Medicine


Dr. Mumtaz Hussain
Assistant Professor Paediatric
Orthopedic Surgery



Dr. Muhammad Bilal Mirza
Associate Prof. of Pead Surgery



Dr. Zia ur Rehman
Associate Prof. of Peads
Neurology



Dr. Nazish Saqlan
Associate Prof. of Pathology
Department

Dr. Sarfraz Ahmad
Associate Prof. Peadiatric
PM&R


Dr. Farah Haroon
Assocaite Prof. of Neonatology


Dr. Uzma Kazmi
Professor of Paediatric Cardiology


Dr. Syed Muhammad Javed Iqbal
Prof. of Peadiatric Medicine Unit-II


Dr. Junaid Rashid
Prof. of Paediatric Medicine / Chairman
Department of Paediatric Medicine

The Children's Hospital & University of Child Health Sciences Lahore

Technical Evaluation

Package No. / Bid No. PC/2020/3/CH & UCHS

Name of the Equipment:

Middle Ear Micro Ear Surgery Set

Required Quantity:

1

Part - I

KNOCKED DOWN CRITERIA - (BID EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	Vertex Medical (Pvt) Ltd	FDS (Pvt) Ltd	Mediland Pakistan (Pvt) Ltd
01	Complete Package / Tender (Mandatory)	Yes	Yes	Yes
02	Original Receipt of Tender (Mandatory)	Yes	Yes	Yes
03	Affidavit from Bidder	Yes	Yes	Yes
04	Bid Security CDR / Bank Guarantee (Mandatory)	Yes	Yes	Yes
05	Bid Validity (Mandatory)	Yes	Yes	Yes
06	Delivery Period (Mandatory)	Yes	Yes	Yes
Remarks:		Responsive	Responsive	Responsive

PART - II

KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	Vertex Medical (Pvt) Ltd	FDS (Pvt) Ltd	Mediland Pakistan (Pvt) Ltd
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	Yes	Yes	Yes
02	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes	Yes
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal	Yes	Yes	Yes
04	Bidder Past Performance for Quoted Product in Teaching Hospitals	Yes	Yes	No
05	Availability of relevant Tools and Testing / Calibration Equipment	Yes	Yes	Yes
06	Warranty	Yes	Yes	Yes
Remarks:		Responsive	Responsive	Non-Responsive

PART - III				
KNOCKED DOWN CRITERIA - (PRODUCT EVALUATION)				
Evaluated by (TAC) Committee				
3	Required Item Name	Vertex Medical (Pvt) Ltd	FDS (Pvt) Ltd	Mediland Pakistan (Pvt) Ltd
	Middle Ear Micro Ear Surgery Set -1			
Sr. #	Brand	Bova	Medicon	KLS Martin
	Quoted Model	Various	Various	Various
01	Country of Manufacturer	Germany	Germany	Germany
02	Manufacturing site	Germany	Germany	Germany
03	Compliance with defined Quality Standards	Compliance	Compliance	
04	Specifications Compliance Feature Wise	Compliance	Compliance	
05	Technical Eligibility of Product	Accepted	Accepted	
06	Technical Eligibility of Firm	Accepted	Accepted	
Technical Status		Responsive	Responsive	Non-Responsive

Enduser (Concerned Department/ Co-opted Member)

Engr. Imran Haider
Bio Medical Engineer/
Co-opted Member

Dr. Imran Saeed
Prof. of Paediatric ENT

Technical Advisory Committee (TAC)

Engr. Syed Ali
Bio Medical Engineer/
Technical Officer

Muhammad Riaz Shahid
Incharge Pay Roll Section

Dr. Syed Ali Abu Zar
AMS (Coordination)

Dr. Faeza Hasnain
Chief Pharmacist

Dr. Muhammad Zeeshan Khan
Assistant Professor Paediatric
Medicine

Dr. Mumtaz Hussain
Assistant Professor Paediatric
Orthopedic Surgery

Dr. Muhammad Bilal Mirza
Associate Prof. of Paed
Surgery

Dr. Zia ur Rehman
Associate Prof. of Peads Neurology

Dr. Nazish Saqlan
Associate Prof. of Pathology
Department

Dr. Sarfraz Ahmad
Associate Prof. Paediatric
PM&R

Dr. Farah Haroon
Associate Prof. of Neonatology

Dr. Uzma Kazmi
Professor of Paediatric
Cardiology

Dr. Syed Muhammad Javed Iqbal
Prof. of Paediatric Medicine Unit-II

Dr. Junaid Rashid
Prof. of Paediatric Medicine / Chairman
Department of Paediatric Medicine

The Children's Hospital & University of Child Health Sciences Lahore

Technical Evaluation

Package No. / Bid No. PC/2020/4/CH & UCHS	Middle Ear Surgery Tympanoplasty & Stapedotomy Set
Name of the Equipment:	1
Required Quantity:	

Part - I

KNOCKED DOWN CRITERIA - (BID EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	FDS (Pvt) Ltd	Mediland Pakistan (Pvt) Ltd	Vertex Medical (Pvt) Ltd
01	Complete Package / Tender (Mandatory)	Yes	Yes	Yes
02	Original Receipt of Tender (Mandatory)	Yes	Yes	Yes
03	Affidavit from Bidder	Yes	Yes	Yes
04	Bid Security CDR / Bank Guarantee (Mandatory)	Yes	Yes	Yes
05	Bid Validity (Mandatory)	Yes	Yes	Yes
06	Delivery Period (Mandatory)	Responsive	Responsive	Responsive
Remarks:				

PART - II

KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	FDS (Pvt) Ltd	Mediland Pakistan (Pvt) Ltd	Vertex Medical (Pvt) Ltd
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	Yes	Yes	Yes
02	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes	Yes
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal	Yes	No	Yes
04	Bidder Past Performance for Quoted Product in Teaching Hospitals	Yes	Yes	Yes
05	Availability of relevant Tools and Testing / Calibration Equipment	Yes	Yes	Yes
06	Warranty	Responsive	Non-Responsive	Responsive
Remarks:				

PART - III

KNOCKED DOWN CRITERIA - (PRODCUT EVALUATION)

Evaluated by (TAC) Committee

4	Required Item Name	FDS (Pvt) Ltd	Mediland Pakistan (Pvt) Ltd	Vertex Medical (Pvt) Ltd
	Middle Ear Surgery Tympanoplasty & Stapedotomy Set - 1			
	Brand	Medicon	KLS martin	B. Braun
	Quoted Model	Various	Various	Various
Sr. #	Country of Manufacturer	Germany	Germany	Germany
01	Manufacturing site	Germany	Germany	Germany
02	Compliance with defined Quality Standards	Compliance		Compliance
03	Specifications Compliance Feature Wise	Compliance		Compliance
04	Technical Eligibility of Product	Accepted		Accepted
05	Technical Eligibility of Firm	Accepted		Accepted
06	Technical Status	Responsive	Non-Responsive	Responsive

Enduser (Concerned Department/ Co-opted Member)

Eng. Imran Haider
Bio Medical Engineer/
Co-opted Member

Dr. Imran Saeed
Prof. of Paediatric ENT

Technical Advisory Committee (TAC)

Eng. Shujaat Ali
Bio Medical Engineer/
Technical Officer

Muhammad Riaz Shahid
Incharge Pay Roll Section

Dr. Syed Ali Abu Zar
AMS (Coordination)

Dr. Faeza Hasnain
Chief Pharmacist

Dr. Muhammad Zeeshan Khan
Assistant Professor Paediatric
Medicine

Dr. Mumtaz Hussain
Assistant Professor Paediatric
Orthopedic Surgery

Dr. Muhammad Bilal Mirza
Associate Prof. of Pead Surgery

Dr. Zia ur Rihman
Associate Prof. of Peads Neurology

Dr. Nazish Saqlan
Associate Prof. of Pathology
Department

Dr. Sarfraz Ahmad
Associate Prof. Peadiatric
PM&R

Dr. Farah Haroon
Assocaite Prof. of Neonatology

Dr. Uzma Kazmi
Professor of Paediatric Cardiology

Dr. Syed Muhammad Javed Iqbal
Prof. of Paediatric Medicine Unit-II

Dr. Junaid Rashid
Prof. of Paediatric Medicine / Chairman
Department of Paediatric Medicine

The Children's Hospital & University of Child Health Sciences Lahore

Technical Evaluation

Package No. / Bid No. PC/2020/5/CH & UCHS

Name of the Equipment:

Myringotomy Set

Required Quantity:

1

Part - I

KNOCKED DOWN CRITERIA - (BID EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	Vertex Medical (Pvt) Ltd
01	Complete Package / Tender (Mandatory)	Yes
02	Original Receipt of Tender (Mandatory)	Yes
03	Affidavit from Bidder	Yes
04	Bid Security CDR / Bank Guarantee (Mandatory)	Yes
05	Bid Validity (Mandatory)	Yes
06	Delivery Period (Mandatory)	Yes
Remarks:		Responsive

PART - II

KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	Vertex Medical (Pvt) Ltd
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	Yes
02	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal	Yes
04	Bidder Past Performance for Quoted Product in Teaching Hospitals	Yes
05	Availability of relevant Tools and Testing / Calibration Equipment	Yes
06	Warranty	Yes
Remarks:		Responsive




PART - III

KNOCKED DOWN CRITERIA - (PRODCUT EVALUATION)

Evaluated by (TAC) Committee

5	Required Item Name	Vertex Medical (Pvt) Ltd
	Myringotomy Set - 1	
	Brand	B. Braun
Sr. #	Quoted Model	Various
01	Country of Manufacturer	Germany
02	Manufacturing site	Germany
03	Compliance with defined Quality Standards	Compliance
04	Specifications Compliance Feature Wise	Compliance
05	Technical Eligibility of Product	Accepted
06	Technical Eligibility of Firm	Accepted
Technical Status		Responsive

Enduser (Concerned Department/ Co-opted Member)



Dr. Irfan Haider
Bio Medical Engineer/
Co-opted Member


Dr. Imran Saeed
Prof. of Paediatric ENT

Technical Advisory Committee (TAC)




Eng. Shujaat Ali
Bio Medical Engineer/
Technical Officer

Muhammad Riaz Shahid
Incharge Pay Roll Section




Dr. Syed Ali Abu Zar
AMS (Coordination)

Dr. Faeza Hasnain
Chief Pharmacist




Dr. Muhammad Zeeshan Khan
Assistant Professor Paediatric
Medicine


Dr. Mumtaz Hussain
Assistant Professor Paediatric
Orthopedic Surgery



Dr. Muhammad Bilal Mirza
Associate Prof. of Paed Surgery




Dr. Zia ur Rehman
Associate Prof. of Paeds
Neurology




Dr. Nazish Saqlan
Associate Prof. of Pathology
Department


Dr. Sarfraz Ahmad
Associate Prof. Paediatric
PM&R




Dr. Farah Haroon
Associate Prof. of Neonatology



Dr. Uzma Kazmi
Professor of Paediatric Cardiology



Dr. Syed Muhammad Javed Iqbal
Prof. of Paediatric Medicine Unit-II



Dr. Junaid Rashid
Prof. of Paediatric Medicine /
Chairman Department of Paediatric
Medicine

The Children's Hospital & University of Child Health Sciences Lahore

Technical Evaluation

Package No. / Bid No. PC/2020/6/CH & UCHS

Name of the Equipment:

ENT Unit Combined

Required Quantity:

3

Part - I

KNOCKED DOWN CRITERIA - (BID EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	Mediland Pakistan (Pvt) Ltd
01	Complete Package / Tender (Mandatory)	Yes
02	Original Receipt of Tender (Mandatory)	Yes
03	Affidavit from Bidder	Yes
04	Bid Security CDR / Bank Guarantee (Mandatory)	Yes
05	Bid Validity (Mandatory)	Yes
06	Delivery Period (Mandatory)	Yes
Remarks:		Responsive

PART - II

KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	Mediland Pakistan (Pvt) Ltd
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	Yes
02	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal	Yes
04	Bidder Past Performance for Quoted Product in Teaching Hospitals	Yes
05	Availability of relevant Tools and Testing / Calibration Equipment	Yes
06	Warranty	Yes
Remarks:		Responsive


PART - III

KNOCKED DOWN CRITERIA - (PRODCUT EVALUATION)

Evaluated by (TAC) Committee

6	Required Item Name	Mediland Pakistan (Pvt) Ltd
	ENT Unit Combined - 3	Euro Clinic
	Brand	Otocompact Professional
Sr. #	Quoted Model	Italy
01	Country of Manufacturer	Italy
02	Manufacturing site	Compliance
03	Compliance with defined Quality Standards	Compliance
04	Specifications Compliance Feature Wise	Accepted
05	Technical Eligibility of Product	Accepted
06	Technical Eligibility of Firm	Responsive
Technical Status		

Enduser (Concerned Department/ Co-opted Member)



Eng. Irfan Haider
Bio Medical Engineer/
Co-opted Member


Dr. Imran Saeed
Prof. of Paediatric ENT

Technical Advisory Committee (TAC)




Eng. Shujat Ali
Bio Medical Engineer/
Technical Officer

Muhammad Riaz Shahid
Incharge Pay Roll Section




Dr. Syed Ali Abu Zar
AMS (Coordination)

Dr. Faeza Hasnain
Chief Pharmacist




Dr. Muhammad Zeeshan Khan
Assistant Professor Paediatric Medicine


Dr. Mumtaz Hussain
Assistant Professor Paediatric
Orthopedic Surgery



Dr. Muhammad Bilal Mirza
Associate Prof. of Paed Surgery




Dr. Zia ur Rehman
Associate Prof. of Paeds
Neurology




Dr. Mazish Saqlan
Associate Prof. of Pathology
Department


Dr. Sarfraz Ahmad
Associate Prof. Paediatric
PM&R



Dr. Farah Haroon
Associate Prof. of Neonatology



Dr. Uzma Kazmi
Professor of Paediatric Cardiology



Dr. Syed Muhammad Javed Iqbal
Prof. of Paediatric Medicine Unit-II



Dr. Javed Rashid
Prof. of Paediatric Medicine /
Chairman Department of Paediatric
Medicine

The Children's Hospital & University of Child Health Sciences Lahore

Technical Evaluation

Package No. / Bid No. PC/2020/7/CH & UCHS	
Name of the Equipment:	Mastoid Drill System
Required Quantity:	2

Part - I

KNOCKED DOWN CRITERIA - (BID EVALUATION)

Evaluated by (TAC) Committee


Sr. #	Evaluation Parameters	FDS (Pvt) Ltd	G Med (Pvt) Ltd	Allmed Solutions	Kasbn International	Popular International
01	Complete Package / Tender (Mandatory)	Yes	Yes	Yes	Yes	Yes
02	Original Receipt of Tender (Mandatory)	Yes	Yes	Yes	Yes	Yes
03	Affidavit from Bidder	Yes	Yes	Yes	Yes	Yes
04	Bid Security CDR / Bank Guarantee (Mandatory)	Yes	Yes	Yes	Yes	Yes
05	Bid Validity (Mandatory)	Yes	Yes	Yes	Yes	Yes
06	Delivery Period (Mandatory)	Yes	Yes	Yes	Yes	Yes
Remarks:		Responsive	Responsive	Responsive	Responsive	Responsive

PART - II

KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)


Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	FDS (Pvt) Ltd	G Med (Pvt) Ltd	Allmed Solutions	Kasbn International	Popular International
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	Yes	Yes	Yes	Yes	Yes
02	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes	Yes	Yes	Yes
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal	Yes	Yes	Yes	Yes	Yes
04	Bidder Past Performance for Quoted Product in Teaching Hospitals	Yes	Yes	Yes	Yes	Yes
05	Availability of relevant Tools and Testing / Calibration Equipment	Yes	Yes	Yes	Yes	Yes
06	Warranty	Yes	Yes	Yes	Yes	Yes
Remarks:		Responsive	Responsive	Responsive	Responsive	Responsive




PART - III						
KNOCKED DOWN CRITERIA - (PRODCUT EVALUATION)						
Evaluated by (TAC) Committee						
7	Required Item Name	FDS (Pvt) Ltd	G Med (Pvt) Ltd	Allmed Solutions	Kasbn International	Popular International
	Mastoid Drill System - 2					
	Brand	Conmed	De Soutter	Karl Storz	Nouvag AG	Medtronic
Sr. #	Quoted Model	D4000A	HDS-	UM600	High Surg 30	Not Quoted
01	Country of Manufacturer	USA	UK	Switzerland	Switzerland	USA
02	Manufacturing site	USA	UK	Switzerland	Switzerland	USA
03	Compliance with defined Quality Standards	Compliance	Expire	Compliance	Compliance	Compliance
04	Specifications Compliance Feature Wise	Compliance		Compliance	Compliance	Compliance
05	Technical Eligibility of Product	Accepted		Accepted	Accepted	Accepted
06	Technical Eligibility of Firm	Accepted		Accepted	Accepted	Accepted
Technical Status		Responsive	Non Responsive	Responsive	Responsive	Responsive

Enduser (Concerned Department/ Co-opted Member



Eng. Irfan Haider
Bio Medical Engineer/
Co-opted Member

Dr. Imran Saeed
Prof. of Peadiatric ENT


Technical Advisory Committee (TAC)


Eng. Shujaat Ali
Bio Medical Engineer/
Technical Officer


Muhammad Riaz Shahid
Incharge Pay Roll Section



Dr. Syed Ali Abu Zar
AMS (Coordination)


Dr. Faeza Hasnain
Chief Pharmacist


Dr. Muhammad Zeeshan Khan
Assistant Professor Paediatric
Medicine


Dr. Mumtaz Hussain
Assistant Professor Paediatric
Orthopedic Surgery



Dr. Muhammad Bilal Mirza
Associate Prof. of Pead Surgery



Dr. Zia ur Rehman
Associate Prof. of Peads Neurology



Dr. Nazish Saqlan
Associate Prof. of Pathology
Department

Dr. Sarfraz Ahmad
Associate Prof. Peadiatric PM&R


Dr. Farah Haroon
Assocaite Prof. of Neonatology


Dr. Uzma Kazmi
Professor of Paediatric
Cardiology


Dr. Syed Muhammad Javed Iqbal
Prof. of Peadiatric Medicine Unit-II


Dr. Junaid Rashid
Prof. of Paediatric Medicine / Chairman
Department of Paediatric Medicine

The Children's Hospital & University of Child Health Sciences Lahore

Technical Evaluation

Package No. / Bid No. PC/2020/8/CH & UCHS

Name of the Equipment:

Cochlear Implantation Set

Required Quantity:

2

Part - I

KNOCKED DOWN CRITERIA - (BID EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	Allmed Solutions
01	Complete Package / Tender (Mandatory)	Yes
02	Original Receipt of Tender (Mandatory)	Yes
03	Affidavit from Bidder	Yes
04	Bid Security CDR / Bank Guarantee (Mandatory)	Yes
05	Bid Validity (Mandatory)	Yes
06	Delivery Period (Mandatory)	Yes

Remarks:

Responsive

PART - II

KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	Allmed Solutions
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	Yes
02	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal	Yes
04	Bidder Past Performance for Quoted Product in Teaching Hospitals	Yes
05	Availability of relevant Tools and Testing / Calibration Equipment	Yes
06	Warranty	Yes

Remarks:

Responsive

PART - III		
KNOCKED DOWN CRITERIA - (PRODCUT EVALUATION)		
Evaluated by (TAC) Committee		
Sr. #	Required Item Name	Allmed Solutions
8	Cochlear Implantation Set - 2	
	Brand	Karl Storz
	Quoted Model	208000
01	Country of Manufacturer	Germany
02	Manufacturing site	Germany
03	Compliance with defined Quality Standards	Compliance
04	Specifications Compliance Feature Wise	Compliance
05	Technical Eligibility of Product	Accepted
06	Technical Eligibility of Firm	Accepted
Technical Status		Responsive

Enduser (Concerned Department/ Co-opted Member)

Engr. Irfan Haider
Bio-Medical Engineer/
Co-opted Member

Dr. Imran Saeed
Prof. of Peadiatric ENT

Technical Advisory Committee (TAC)

Engr. Shujaat Ali
Bio Medical Engineer/
Technical Officer

Muhammad Riaz Shahid
Incharge Pay Roll Section

Dr. Syed Ali Abu Zar
AMS (Coordination)

Dr. Faeza Hasnain
Chief Pharmacist

Dr. Muhammad Zeeshan Khan
Assistant Professor Paediatric
Medicine

Dr. Mumtaz Hussain
Assistant Professor Paediatric
Orthopedic Surgery

Dr. Muhammad Bilal Mirza
Associate Prof. of Pead Surgery

Dr. Zia ur Rehman
Associate Prof. of Peads
Neurology

Dr. Nazish Saqlan
Associate Prof. of Pathology
Department

Dr. Sarfraz Ahmad
Associate Prof.
Peadiatric PM&R

Dr. Farah Haroon
Assocaite Prof. of Neonatology

Dr. Uzma Kazmi
Professor of Paediatric Cardiology

Dr. Syed Muhammad Javed Iqbal
Prof. of Peadiatric Medicine Unit-II

Dr. Junaid Rashid
Prof. of Paediatric Medicine / Chairman
Department of Paediatric Medicine

The Children's Hospital & University of Child Health Sciences Lahore

Technical Evaluation

Package No. / Bid No. PC/2020/10/CH & UCHS

Name of the Equipment:

Operating Microscope for ENT

Required Quantity:

1

Part - I

KNOCKED DOWN CRITERIA - (BID EVALUATION)

Evaluated by (TAC) Committee


Sr. #	Evaluation Parameters	Eastern Optical Co.	Latif Brothers	Medical equipment & System	Jasani Scientific
01	Complete Package / Tender (Mandatory)	Yes	Yes	Yes	Yes
02	Original Receipt of Tender (Mandatory)	Yes	Yes	Yes	Yes
03	Affidavit from Bidder	Yes	Yes	Yes	Yes
04	Bid Security CDR / Bank Guarantee (Mandatory)	No	Yes	Yes	Yes
05	Bid Validity (Mandatory)	Yes	Yes	Yes	Yes
06	Delivery Period (Mandatory)	Yes	Yes	Yes	Yes
Remarks:		Non-Responsive	Responsive	Responsive	Responsive

PART - II

KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	Eastern Optical Co.	Latif Brothers	Medical equipment & System	Jasani Scientific
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	No	Yes	No	Yes
02	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes	Yes	Yes
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal	Yes	Yes	Yes	Yes
04	Bidder Past Performance for Quoted Product in Teaching Hospitals	NO	Yes	NO	Yes
05	Availability of relevant Tools and Testing / Calibration Equipment	Yes	Yes	Yes	Yes
06	Warranty	Yes	Yes	Yes	Yes
Remarks:		Non-Responsive	Responsive	Non-Responsive	Responsive



PART - III					
KNOCKED DOWN CRITERIA - (PRODCUT EVALUATION)					
Evaluated by (TAC) Committee					
10	Required Item Name	Eastern Optical Co.	Latif Brothers	Medical equipment & System	Jasani Scientific
	Operating Microscope for ENT - 1				
	Brand	Karl Kaps	Carl Zeiss	Karl Kaps	Leica
Sr. #	Quoted Model	Kaps Pro ENT Motorized	OPMI-Sensera	Kaps Pro ENT Motorized	M525F20
01	Country of Manufacturer	Germany	Germany	Germany	Germany
02	Manufacturing site	Germany	Germany	Germany	Germany
03	Compliance with defined Quality Standards		Compliance		Compliance
04	Specifications Compliance Feature Wise		Compliance		Compliance
05	Technical Eligibility of Product		Accepted		Accepted
06	Technical Eligibility of Firm		Accepted		Accepted
Technical Status		Non-Responsive	Responsive	Non-Responsive	Responsive

Enduser (Concerned Department/ Co-opted Member)

Eng. Irfan Haider
Bio-Medical Engineer/
Co-opted Member

Dr. Imran Saeed
Prof. of Paediatric ENT

Technical Advisory Committee (TAC)

Engr. Shujaat Ali
Bio Medical Engineer/
Technical Officer

Muhammad Riaz Shahid
Incharge Pay Roll Section

Dr. Syed Ali Abu Zar
AMS (Coordination)

Dr. Faeza Hasnain
Chief Pharmacist

Dr. Muhammad Zeeshan Khan
Assistant Professor Paediatric
Medicine

Dr. Mumtaz Hussain
Assistant Professor Paediatric
Orthopedic Surgery

Dr. Muhammad Bilal Mirza
Associate Prof. of Paed
Surgery

Dr. Zia ur Rehman
Associate Prof. of Paeds
Neurology

Dr. Nazish Saqlan
Associate Prof. of Pathology
Department

Dr. Sarfraz Ahmad
Associate Prof. Paediatric PM&R

Dr. Farah Haroon
Associate Prof. of Neonatology

Dr. Uzma Kazmi
Professor of Paediatric
Cardiology

Dr. Syed Muhammad Javed Iqbal
Prof. of Paediatric Medicine Unit-II

Dr. Junaid Rashid
Prof. of Paediatric Medicine / Chairman
Department of Paediatric Medicine

The Children's Hospital & University of Child Health Sciences Lahore

Technical Evaluation

Package No. / Bid No. PC/2020/11/CH & UCHS	
Name of the Equipment:	Operating Head Lights
Required Quantity:	4

Part - I

KNOCKED DOWN CRITERIA - (BID EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	Medilan Pakistan (Pvt) Ltd	Allmed Solutions
01	Complete Package / Tender (Mandatory)	Yes	Yes
02	Original Receipt of Tender (Mandatory)	Yes	Yes
03	Affidavit from Bidder	Yes	Yes
04	Bid Security CDR / Bank Guarantee (Mandatory)	Yes	Yes
05	Bid Validity (Mandatory)	Yes	Yes
06	Delivery Period (Mandatory)	Yes	Yes
Remarks:		Responsive	Responsive

PART - II

KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	Medilan Pakistan (Pvt) Ltd	Allmed Solutions
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	Yes	Yes
02	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal	Yes	Yes
04	Bidder Past Performance for Quoted Product in Teaching Hospitals	Yes	Yes
05	Availability of relevant Tools and Testing / Calibration Equipment	Yes	Yes
06	Warranty	Yes	Yes
Remarks:		Responsive	Responsive



PART - III

KNOCKED DOWN CRITERIA - (PRODCUT EVALUATION)

Evaluated by (TAC) Committee


11	Required Item Name	Medilan Pakistan (Pvt) Ltd	Allmed Solutions
	Operating Head Lights - 4		
	Brand		
Sr. #	Quoted Model	KLS Martin	Karl Storz
		MEDLED Chrome MC6	310060/20161201
01	Country of Manufacturer	Germany	Germany
02	Manufacturing site	USA	Germany
03	Compliance with defined Quality Standards	Compliance	Compliance
04	Specifications Compliance Feature Wise	Compliance	Compliance
05	Technical Eligibility of Product	Accepted	Accepted
06	Technical Eligibility of Firm	Accepted	Accepted
Technical Status		Responsive	Responsive

Enduser (Concerned Department/ Co-opted Member)



Engr. Ifan Haider
Bio-Medical Engineer/
Co-opted Member

Dr. Imran Saeed
Prof. of Paediatric ENT


Technical Advisory Committee (TAC)


Engr. Shujat Ali
Bio Medical Engineer/
Technical Officer


Muhammad Riaz Shahid
Incharge Pay Roll Section



Dr. Syed Ali Abu Zar
AMS (Coordination)


Dr. Faeza Hasnain
Chief Pharmacist


Dr. Muhammad Zeeshan Khan
Assistant Professor Paediatric
Medicine


Dr. Mumtaz Hussain
Assistant Professor Paediatric
Orthopedic Surgery



Dr. Muhammad Bilal Mirza
Associate Prof. of Pead
Surgery



Dr. Zia ur Rehman
Associate Prof. of Peads Neurology



Dr. Nazish Saqlan
Associate Prof. of
Pathology Department

Dr. Sarfraz Ahmad
Associate Prof. Paediatric
PM&R


Dr. Farah Haroon
Assocate Prof. of Neonatology


Dr. Uzma Kazmi
Professor of Paediatric
Cardiology


Dr. Syed Muhammad Javed Iqbal
Prof. of Paediatric Medicine Unit-II


Dr. Junaid Rashid
Prof. of Paediatric Medicine / Chairman
Department of Paediatric Medicine

The Children's Hospital & University of Child Health Sciences Lahore

Technical Evaluation

Package No. / Bid No. PC/2020/12/CH & UCHS

Name of the Equipment: Endoscope Camera System (High Defination)

Required Quantity: 1

Part - I

KNOCKED DOWN CRITERIA - (BID EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	Allmed Solutions
01	Complete Package / Tender (Mandatory)	Yes
02	Original Receipt of Tender (Mandatory)	Yes
03	Affidavit from Bidder	Yes
04	Bid Security CDR / Bank Guarantee (Mandatory)	Yes
05	Bid Validity (Mandatory)	Yes
06	Delivery Period (Mandatory)	Yes
Remarks:		Responsive

PART - II

KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	Allmed Solutions
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	Yes
02	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal	Yes
04	Bidder Past Performance for Quoted Product in Teaching Hospitals	Yes
05	Availability of relevant Tools and Testing / Calibration Equipment	Yes
06	Warranty	Yes
Remarks:		Responsive



PART - III

KNOCKED DOWN CRITERIA - (PRODCUT EVALUATION)

Evaluated by (TAC) Committee

12	Required Item Name	Allmed Solutions
	Endoscope Camera System (High Definition) - 1	
	Brand	Karl storz
Sr. #	Quoted Model	TC100EN
01	Country of Manufacturer	Germany
02	Manufacturing site	Germany
03	Compliance with defined Quality Standards	Compliance
04	Specifications Compliance Feature Wise	Compliance
05	Technical Eligibility of Product	Accepted
06	Technical Eligibility of Firm	Accepted
Technical Status		Responsive

Enduser (Concerned Department/ Co-opted Member)

Technical Advisory Committee (TAC)

Eng. Shuraf Ali
Bio Medical Engineer/
Technical Officer

Dr. Mumtaz Hussain
Assistant Professor Paediatric
Orthopedic Surgery

Dr. Farah Haroon
Associate Prof. of Neonatology

Muhammad Riaz Shahid
Incharge Pay Roll Section

Dr. Muhammad Bilal Mirza
Associate Prof. of Pead Surgery

Dr. Uzma Kazmi
Professor of Paediatric Cardiology

Eng. Irfan Haider
Bio Medical Engineer/
Co-opted Member

Dr. Syed Ali Abu Zar
AMS (Coordination)

Dr. Zia ur Rehman
Associate Prof. of Peads
Neurology

Dr. Syed Muhammad Javed Iqbal
Prof. of Paediatric Medicine Unit-II

Dr. Imran Saeed
Prof. of Peadiatric ENT

Dr. Faeza Hasnain
Chief Pharmacist

Dr. Nazish Saqlan
Associate Prof. of Pathology
Department

Dr. Junaid Rashid
Prof. of Paediatric Medicine / Chairman
Department of Paediatric Medicine

Dr. Muhammad Zeeshan Khan
Assistant Professor Paediatric
Medicine

Dr. Sarfraz Ahmad
Associate Prof. Peadiatric
PM&R

The Children's Hospital & University of Child Health Sciences Lahore

Technical Evaluation

Package No. / Bid No. PC/2020/13/CH & UCHS

Name of the Equipment:

Coblation Equipment

Required Quantity:

2

Part - I

KNOCKED DOWN CRITERIA - (BID EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	Popular International
01	Complete Package / Tender (Mandatory)	Yes
02	Original Receipt of Tender (Mandatory)	Yes
03	Affidavit from Bidder	Yes
04	Bid Security CDR / Bank Guarantee (Mandatory)	Yes
05	Bid Validity (Mandatory)	Yes
06	Delivery Period (Mandatory)	Yes
Remarks:		Responsive

PART - II

KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	Popular International
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	Yes
02	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal	Yes
04	Bidder Past Performance for Quoted Product in Teaching Hospitals	Yes
05	Availability of relevant Tools and Testing / Calibration Equipment	Yes
06	Warranty	Yes
Remarks:		Responsive


PART - III		
KNOCKED DOWN CRITERIA - (PRODCUT EVALUATION)		
Evaluated by (TAC) Committee		
13	Required Item Name	Popular International
	Coblation Equipment - 2	
	Brand	
Sr. #	Quoted Model	AEX TM Generator
01	Country of Manufacturer	USA
02	Manufacturing site	USA
03	Compliance with defined Quality Standards	Compliance
04	Specifications Compliance Feature Wise	Compliance
05	Technical Eligibility of Product	Accepted
06	Technical Eligibility of Firm	Accepted
Technical Status		Responsive

Enduser (Concerned Department/ Co-opted Member)

Technical Advisory Committee (TAC)



Engr. Shujaat Ali
Bio Medical Engineer/
Technical Officer

Muhammad Riaz Shahid
Incharge Pay Roll Section



Dr. Syed Ali Abu Zar
AMS (Coordination)


Dr. Faeza Hasnain
Chief Pharmacist


Dr. Imran Saeed
Prof. of Peadiatric ENT


Dr. Muhammad Zeeshan Khan
Assistant Professor Paediatric
Medicine


Dr. Mumtaz Hussain
Assistant Professor Paediatric
Orthopedic Surgery



Dr. Muhammad Bilal Mirza
Associate Prof. of Pead Surgery



Dr. Zia ur Rehman
Associate Prof. of Peads
Neurology



Dr. Nazish Saqlan
Associate Prof. of Pathology
Department

Dr. Sarfraz Ahmad
Associate Prof. Peadiatric
PM&R


Dr. Farah Haroon
Assocaite Prof. of Neonatology


Dr. Uzma Kazmi
Professor of Paediatric Cardiology


Dr. Syed Muhammad Javed Iqbal
Prof. of Peadiatric Medicine Unit-II


Dr. Junaid Rashid
Prof. of Paediatric Medicine / Chairman
Department of Paediatric Medicine

The Children's Hospital & University of Child Health Sciences Lahore

Technical Evaluation

Package No. / Bid No. PC/2020/14/CH & UCHS

Name of the Equipment:

OTO Endoscope

Required Quantity:

1

Part - I

KNOCKED DOWN CRITERIA - (BID EVALUATION)

Evaluated by (TAC) Committee


Sr. #	Evaluation Parameters	Allmed Solutions
01	Complete Package / Tender (Mandatory)	Yes
02	Original Receipt of Tender (Mandatory)	Yes
03	Affidavit from Bidder	Yes
04	Bid Security CDR / Bank Guarantee (Mandatory)	Yes
05	Bid Validity (Mandatory)	Yes
06	Delivery Period (Mandatory)	Yes
Remarks:		Responsive

PART - II

KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	Allmed Solutions
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	Yes
02	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal	Yes
04	Bidder Past Performance for Quoted Product in Teaching Hospitals	Yes
05	Availability of relevant Tools and Testing / Calibration Equipment	Yes
06	Warranty	Yes
Remarks:		Responsive



PART - III

KNOCKED DOWN CRITERIA - (PRODCUT EVALUATION)

Evaluated by (TAC) Committee


Required Item Name		Allmed Solutions
14	OTO Endoscope - 1	Karl Storz
	Brand	Various
	Quoted Model	Germany
Sr. #	Country of Manufacturer	Germany
01	Manufacturing site	Compliance
02	Compliance with defined Quality Standards	Compliance
03	Specifications Compliance Feature Wise	Accepted
04	Technical Eligibility of Product	Accepted
05	Technical Eligibility of Firm	Responsive
06	Technical Status	

Enduser (Concerned Department/ Co-opted Member)



Engr. Irfan Haider
Bio Medical Engineer/
Co-opted Member

Dr. Imran Saeed
Prof. of Paediatric ENT


Technical Advisory Committee (TAC)


Engr. Shujaat Ali
Bio Medical Engineer/
Technical Officer


Muhammad Riaz Shahid
Incharge Pay Roll Section



Dr. Syed Ali Abu Zar
AMS (Coordination)


Dr. Faeza Hasnain
Chief Pharmacist


Dr. Muhammad Zeeshan Khan
Assistant Professor Paediatric
Medicine


Dr. Mumtaz Hussain
Assistant Professor Paediatric
Orthopedic Surgery



Dr. Muhammad Bilal Mirza
Associate Prof. of Pead Surgery



Dr. Zia ur Rehman
Associate Prof. of Peads
Neurology



Dr. Nazish Saqlan
Associate Prof. of Pathology
Department

Dr. Sarfraz Ahmad
Associate Prof.
Paediatric PM&R


Dr. Farah Haroon
Assocaita Prof. of Neonatology


Dr. Uzma Kazmi
Professor of Paediatric Cardiology


Dr. Syed Muhammad Javed Iqbal
Prof. of Paediatric Medicine Unit-II


Dr. Junaid Rashid
Prof. of Paediatric Medicine / Chairman
Department of Paediatric Medicine

The Children's Hospital & University of Child Health Sciences Lahore

Technical Evaluation

Package No. / Bid No. PC/2020/15/CH & UCHS

Name of the Equipment:

Pulse Oximeter

Required Quantity:

4

Part - I

KNOCKED DOWN CRITERIA - (BID EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	Eastern Medical Tech	Al-Kareem Medical Tech	Hospicare System	Popular International
01	Complete Package / Tender (Mandatory)	Yes	Yes	Yes	Yes
02	Original Receipt of Tender (Mandatory)	Yes	Yes	Yes	Yes
03	Affidavit from Bidder	Yes	Yes	Yes	Yes
04	Bid Security CDR / Bank Guarantee (Mandatory)	Yes	Yes	Yes	Yes
05	Bid Validity (Mandatory)	Yes	Yes	Yes	Yes
06	Delivery Period (Mandatory)	Yes	Yes	Yes	Yes
Remarks:		Responsive	Responsive	Responsive	Responsive

PART - II

KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	Eastern Medical Tech	Al-Kareem Medical Tech	Hospicare System	Popular International
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	Yes	Yes	Yes	Yes
02	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes	Yes	Yes
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal	Yes	Yes	Yes	Yes
04	Bidder Past Performance for Quoted Product in Teaching Hospitals	NO	No	Yes	Yes
05	Availability of relevant Tools and Testing / Calibration Equipment	Yes	Yes	Yes	Yes
06	Warranty	Yes	Yes	Yes	Yes
Remarks:		Non-Responsive	Non-Responsive	Responsive	Responsive


PART - III

KNOCKED DOWN CRITERIA - (PRODCUT EVALUATION)

Evaluated by (TAC) Committee


15	Required Item Name	Eastern Medical Tech	Al-Kareem Medical Tech	Hospicare System	Popular International
	Pulse Oximeter - 4				
	Brand				
Sr. #	Quoted Model	PAVO	NTP-80	Rad-97	Nellcor
01	Country of Manufacturer	Germany	UK	USA	USA
02	Manufacturing site	Germany	China	USA	Korea
03	Compliance with defined Quality Standards			Compliance	Compliance
04	Specifications Compliance Feature Wise			Compliance	Compliance
05	Technical Eligibility of Product			Accepted	Accepted
06	Technical Eligibility of Firm			Accepted	Accepted
Technical Status		Non Responsive	Non-Responsive	Responsive	Responsive

Enduser (Concerned Department/ Co-opted Member)



Engr. Irfan Haider
Bio Medical Engineer/
Co-opted Member

Dr. Imran Saeed
Prof. of Paediatric ENT


Technical Advisory Committee (TAC)


Engr. Shujaat
Bio Medical Engineer/
Technical Officer


Muhammad Riaz Shahid
Incharge Pay Roll Section



Dr. Syed Ali Abu Zar
AMS (Coordination)


Dr. Faeza Hasnain
Chief Pharmacist


Dr. Muhammad Zeeshan Khan
Assistant Professor Paediatric Medicine


Dr. Mumtaz Hussain
Assistant Professor Paediatric
Orthopedic Surgery



Dr. Muhammad Bilal Mirza
Associate Prof. of Pead
Surgery



Dr. Zia ur Rehman
Associate Prof. of Peads Neurology



Dr. Nazish Saqlan
Associate Prof. of Pathology
Department

Dr. Sarfraz Ahmad
Associate Prof.
Paediatric PM&R


Dr. Farah Haroon
Associate Prof. of Neonatology


Dr. Uzma Kazmi
Professor of Paediatric Cardiology


Dr. Syed Muhammad Javed Iqbal
Prof. of Paediatric Medicine Unit-II


Dr. Junaid Rashid
Prof. of Paediatric Medicine / Chairman
Department of Paediatric Medicine

The Children's Hospital & University of Child Health Sciences Lahore

Technical Evaluation

Package No. / Bid No. PC/2020/16/CH & UCHS

Name of the Equipment:

Rigid Bronchoscope Set with Optical Forceps

Required Quantity:

1

Part - I

KNOCKED DOWN CRITERIA - (BID EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	Allmed Solutions
01	Complete Package / Tender (Mandatory)	Yes
02	Original Receipt of Tender (Mandatory)	Yes
03	Affidavit from Bidder	Yes
04	Bid Security CDR / Bank Guarantee (Mandatory)	Yes
05	Bid Validity (Mandatory)	Yes
06	Delivery Period (Mandatory)	Yes
Remarks:		Responsive

PART - II

KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	Allmed Solutions
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	Yes
02	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal	Yes
04	Bidder Past Performance for Quoted Product in Teaching Hospitals	Yes
05	Availability of relevant Tools and Testing / Calibration Equipment	Yes
06	Warranty	Yes
Remarks:		Responsive

PART - III		
KNOCKED DOWN CRITERIA - (PRODCUT EVALUATION)		
Evaluated by (TAC) Committee		
	Required Item Name	Allmed Solutions
16	Rigid Bronchoscope Set - 1	Karl Storz
	Brand	10339F
	Quoted Model	Germany
Sr. #	Country of Manufacturer	Germany
01	Manufacturing site	Compliance
02	Compliance with defined Quality Standards	Compliance
03	Specifications Compliance Feature Wise	Accepted
04	Technical Eligibility of Product	Accepted
05	Technical Eligibility of Firm	Responsive
06		
Technical Status		
Enduser (Concerned Department/ Co-opted Member)		

Technical Advisory Committee (TAC)

Engr. Shujaa Ali
Bio Medical Engineer/
Technical Officer

Dr. Mumtaz Hussain
Assistant Professor Paediatric
Orthopedic Surgery

Dr. Farah Haroon
Assocaite Prof. of Neonatology

Muhammad Riaz Shahid
Incharge Pay Roll Section

Dr. Muhammad Bilal Mirza
Associate Prof. of Pead Surgery

Dr. Uzma Kazmi
Professor of Paediatric Cardiology

Engr. Irfan Haider
Bio Medical Engineer/
Co-opted Member

Dr. Syed Ali Abu Zar
AMS (Coordination)

Dr. Zia ur Rehman
Associate Prof. of Peads
Neurology

Dr. Syed Muhammad Javed Iqbal
Prof. of Paediatric Medicine Unit-II

Dr. Faeza Hasnain
Chief Pharmacist

Dr. Nazish Saqlan
Associate Prof. of Pathology
Department

Dr. Junaid Rashid
Prof. of Paediatric Medicine / Chairman
Department of Paediatric Medicine

Dr. Imran Saeed
Prof. of Peadiatric ENT

Dr. Muhammad Zeeshan Khan
Assistant Professor Paediatric
Medicine

Dr. Sarfraz Ahmad
Associate Prof. Peadiatric
PM&R

The Children's Hospital & University of Child Health Sciences Lahore

Technical Evaluation

Package No. / Bid No. PC/2020/17/CH & UCHS

Name of the Equipment:

Laryngoscope Set

Required Quantity:

2

Part - I

KNOCKED DOWN CRITERIA - (BID EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	Allimed Solutions
01	Complete Package / Tender (Mandatory)	Yes
02	Original Receipt of Tender (Mandatory)	Yes
03	Affidavit from Bidder	Yes
04	Bid Security CDR / Bank Guarantee (Mandatory)	Yes
05	Bid Validity (Mandatory)	Yes
06	Delivery Period (Mandatory)	Yes
Remarks:		Responsive

PART - II

KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	Allimed Solutions
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	Yes
02	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal	Yes
04	Bidder Past Performance for Quoted Product in Teaching Hospitals	Yes
05	Availability of relevant Tools and Testing / Calibration Equipment	Yes
06	Warranty	Yes
Remarks:		Responsive


PART - III

KNOCKED DOWN CRITERIA - (PRODCUT EVALUATION)

Evaluated by (TAC) Committee


17	Required Item Name	Allmed Solutions
	Laryngoscope Set - 2	
	Brand	Karl Storz
Sr. #	Quoted Model	8590J
01	Country of Manufacturer	Germany
02	Manufacturing site	Germany
03	Compliance with defined Quality Standards	Compliance
04	Specifications Compliance Feature Wise	Compliance
05	Technical Eligibility of Product	Accepted
06	Technical Eligibility of Firm	Accepted
Technical Status		Responsive

Enduser (Concerned Department/ Co-opted Member)



Eng. Irfan Haider
Bio Medical Engineer/
Co-opted Member

Dr. Imran Saeed
Prof. of Peadiatric ENT

Technical Advisory Committee (TAC)



Eng. Shujaat Ali
Bio Medical Engineer/
Technical Officer

Muhammad Riaz Shahid
Incharge Pay Roll Section




Dr. Syed Ali Abu Zar
AMS (Coordination)

Dr. Faeza Hasnain
Chief Pharmacist




Dr. Muhammad Zeeshan Khan
Assistant Professor Paediatric Medicine


Dr. Mumtaz Hussain
Assistant Professor Paediatric
Orthopedic Surgery



Dr. Muhammad Bilal Mirza
Associate Prof. of Pead Surgery




Dr. Zia ur Rehman
Associate Prof. of Peads
Neurology




Dr. Nazish Saqlan
Associate Prof. of Pathology
Department

Dr. Sarfraz Ahmad
Associate Prof. Peadiatric
PM&R




Dr. Farah Haroon
Assocaite Prof. of Neonatology



Dr. Azma Kazmi
Professor of Paediatric
Cardiology



Dr. Syed Muhammad Javed Iqbal
Prof. of Peadiatric Medicine Unit-II



Dr. Junaid Rashid
Prof. of Paediatric Medicine / Chairman
Department of Paediatric Medicine

The Children's Hospital & University of Child Health Sciences Lahore

Technical Evaluation

Package No. / Bid No. PC/2020/18/CH & UCHS

Name of the Equipment:

Regid Esophagoscope Set with Optical Forceps

Required Quantity:

1

Part - I

KNOCKED DOWN CRITERIA - (BID EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	Allmed Solutions
01	Complete Package / Tender (Mandatory)	Yes
02	Original Receipt of Tender (Mandatory)	Yes
03	Affidavit from Bidder	Yes
04	Bid Security CDR / Bank Guarantee (Mandatory)	Yes
05	Bid Validity (Mandatory)	Yes
06	Delivery Period (Mandatory)	Responsive

Remarks:

PART - II

KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)


Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	Allmed Solutions
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	Yes
02	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal	Yes
04	Bidder Past Performance for Quoted Product in Teaching Hospitals	Yes
05	Availability of relevant Tools and Testing / Calibration Equipment	Yes
06	Warranty	Responsive

Remarks:


PART - III		
KNOCKED DOWN CRITERIA - (PRODCUT EVALUATION)		
Evaluated by (TAC) Committee		
	Required Item Name	Allmed Solutions
1	Regid esophagoscope Set with Optical Forceps - 1	Karl Storz 12030A Germany
	Brand	
	Quoted Model	
Sr. #	Country of Manufacturer	Germany
01	Manufacturing site	Compliance
02	Compliance with defined Quality Standards	Compliance
03	Specifications Compliance Feature Wise	Accepted
04	Technical Eligibility of Product	Accepted
05	Technical Eligibility of Firm	Responsive
Technical Status		

Enduser (Concerned Department/ Co-opted Member)



Eng. Imran Haider
Bio-Medical Engineer/
Co-opted Member

Dr. Imran Saeed
Prof. of Paediatric ENT


Technical Advisory Committee (TAC)


Eng. Shujaat Ali
Bio Medical Engineer/
Technical Officer


Muhammad Riaz Shahid
Incharge Pay Roll Section



Dr. Syed Ali Abu Zar
AMS (Coordination)


Dr. Faeza Hasnain
Chief Pharmacist


Dr. Muhammad Zeeshan Khan
Assistant Professor Paediatric
Medicine


Dr. Mumtaz Hussain
Assistant Professor Paediatric
Orthopedic Surgery



Dr. Muhammad Bilal Mirza
Associate Prof. of Paed Surgery



Dr. Zia ur Rehman
Associate Prof. of Paeds
Neurology



Dr. Nazish Saqlan
Associate Prof. of Pathology
Department

Dr. Sarfraz Ahmad
Associate Prof. Paediatric
PM&R


Dr. Farah Haroon
Associate Prof. of Neonatology


Dr. Uzma Kazmi
Professor of Paediatric Cardiology


Dr. Syed Muhammad Javed Iqbal
Prof. of Paediatric Medicine Unit-II


Dr. Junaid Rashid
Prof. of Paediatric Medicine / Chairman
Department of Paediatric Medicine

The Children's Hospital & University of Child Health Sciences Lahore

Technical Evaluation

Package No. / Bid No. PC/2020/19/CH & UCHS

Bronchial Endoscope

Name of the Equipment:

6

Required Quantity:

Part - I

KNOCKED DOWN CRITERIA - (BID EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	Allmed Solutions
01	Complete Package / Tender (Mandatory)	Yes
02	Original Receipt of Tender (Mandatory)	Yes
03	Affidavit from Bidder	Yes
04	Bid Security CDR / Bank Guarantee (Mandatory)	Yes
05	Bid Validity (Mandatory)	Yes
06	Delivery Period (Mandatory)	Responsive

Remarks:

PART - II

KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	Allmed Solutions
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	Yes
02	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal	Yes
04	Bidder Past Performance for Quoted Product in Teaching Hospitals	Yes
05	Availability of relevant Tools and Testing / Calibration Equipment	Yes
06	Warranty	Responsive

Remarks:

PART - III		
KNOCKED DOWN CRITERIA - (PRODCUT EVALUATION)		
Evaluated by (TAC) Committee		
	Required Item Name	Allmed Solutions
19	Bronchial Endoscope - 6	
	Brand	Karl Storz
	Quoted Model	10339B
01	Country of Manufacturer	Germany
02	Manufacturing site	Germany
03	Compliance with defined Quality Standards	Compliance
04	Specifications Compliance Feature Wise	Compliance
05	Technical Eligibility of Product	Accepted
06	Technical Eligibility of Firm	Accepted
Technical Status		Responsive

Enduser (Concerned Department/ Co-opted Member)



Engr. Irfan Haider
Bio Medical Engineer/
Co-opted Member

Dr. Imran Saeed
Prof. of Paediatric ENT


Technical Advisory Committee (TAC)


Engr. Shujaat Ali
Bio Medical Engineer/
Technical Officer


Muhammad Riaz Shahid
Incharge Pay Roll Section



Dr. Syed Ali Abu Zar
AMS (Coordination)

Dr. Faeza Hasnain
Chief Pharmacist


Dr. Muhammad Zeeshan Khan
Assistant Professor Paediatric
Medicine


Dr. Mumtaz Hussain
Assistant Professor Paediatric
Orthopedic Surgery



Dr. Muhammad Bilal Mirza
Associate Prof. of Pead Surgery



Dr. Zia ur Rehman
Associate Prof. of Peads
Neurology



Dr. Nazish Saqlan
Associate Prof. of Pathology
Department

Dr. Sarfraz Ahmad
Associate Prof.
Paediatric PM&R


Dr. Farah Haroon
Assocaite Prof. of Neonatology


Dr. Uzma Kazmi
Professor of Paediatric Cardiology


Dr. Syed Muhammad Javed Iqbal
Prof. of Paediatric Medicine Unit-II


Dr. Junaid Rashid
Prof. of Paediatric Medicine / Chairman
Department of Paediatric Medicine

The Children's Hospital & University of Child Health Sciences Lahore

Technical Evaluation

Package No. / Bid No. PC/2020/20/CH & UCHS Micro Layrngcal System Set

Name of the Equipment:

2

Required Quantity:

Part - I

KNOCKED DOWN CRITERIA - (BID EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	Allmed Solutions
01	Complete Package / Tender (Mandatory)	Yes
02	Original Receipt of Tender (Mandatory)	Yes
03	Affidavit from Bidder	Yes
04	Bid Security CDR / Bank Guarantee (Mandatory)	Yes
05	Bid Validity (Mandatory)	Yes
06	Delivery Period (Mandatory)	Responsive

Remarks:

PART - II

KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	Allmed Solutions
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	Yes
02	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal	Yes
04	Bidder Past Performance for Quoted Product in Teaching Hospitals	Yes
05	Availability of relevant Tools and Testing / Calibration Equipment	Yes
06	Warranty	Responsive

Remarks:


PART - III

KNOCKED DOWN CRITERIA - (PRODCUT EVALUATION)

Evaluated by (TAC) Committee


Required Item Name		Allmed Solutions
20	Micro Layrngcal System Set - 2	
	Brand	Karl Storz
	Quoted Model	8590C
Sr. #	Country of Manufacturer	Germany
01	Manufacturing site	Germany
02	Compliance with defined Quality Standards	Compliance
03	Specifications Compliance Feature Wise	Compliance
04	Technical Eligibility of Product	Accepted
05	Technical Eligibility of Firm	Accepted
06	Technical Status	Responsive

Enduser (Concerned Department/ Co-opted Member)



Engr. Irfan Haider
Bio Medical Engineer/
Co-opted Member

Dr. Imran Saeed
Prof. of Peadiatric ENT


Technical Advisory Committee (TAC)


Engr. Sanaullah
Bio Medical Engineer/
Technical Officer


Muhammad Riaz Shahid
Incharge Pay Roll Section



Dr. Syed Ali Abu Zar
AMS (Coordination)


Dr. Faeza Hasnain
Chief Pharmacist


Dr. Muhammad Zeeshan Khan
Assistant Professor Paediatric Medicine


Dr. Mumtaz Hussain
Assistant Professor Paediatric
Orthopedic Surgery



Dr. Muhammad Bilal Mirza
Associate Prof. of Pead Surgery



Dr. Zia ur Rehman
Associate Prof. of Peads
Neurology



Dr. Nazish Saqlan
Associate Prof. of Pathology
Department

Dr. Sarfraz Ahmad
Associate Prof. Peadiatric
PM&R


Dr. Farah Haroon
Assocaite Prof. of Neonatology


Dr. Uzma Kazmi
Professor of Paediatric
Cardiology


Dr. Syed Muhammad Javed Iqbal
Prof. of Peadiatric Medicine Unit-II


Dr. Junaid Rashid
Prof. of Paediatric Medicine / Chairman
Department of Paediatric Medicine

The Children's Hospital & University of Child Health Sciences Lahore

Technical Evaluation

Package No. / Bid No. PC/2020/21/CH & UCHS

Name of the Equipment:

Tracheostomy Set

Required Quantity:

2

Part - I

KNOCKED DOWN CRITERIA - (BID EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	FDS (Pvt) Ltd	Vertex Medical (Pvt) Ltd
01	Complete Package / Tender (Mandatory)	Yes	Yes
02	Original Receipt of Tender (Mandatory)	Yes	Yes
03	Affidavit from Bidder	Yes	Yes
04	Bid Security CDR / Bank Guarantee (Mandatory)	Yes	Yes
05	Bid Validity (Mandatory)	Yes	Yes
06	Delivery Period (Mandatory)	Yes	Yes
Remarks:		Responsive	Responsive

PART - II


KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	FDS (Pvt) Ltd	Vertex Medical (Pvt) Ltd
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	Yes	Yes
02	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal	Yes	Yes
04	Bidder Past Performance for Quoted Product in Teaching Hospitals	Yes	Yes
05	Availability of relevant Tools and Testing / Calibration Equipment	Yes	Yes
06	Warranty	Yes	Yes
Remarks:		Responsive	Responsive

PART - III			
KNOCKED DOWN CRITERIA - (PRODCUT EVALUATION)			
Evaluated by (TAC) Committee			
1	Required Item Name	FDS (Pvt) Ltd	Vertex Medical (Pvt) Ltd
	Tracheostomy Set - 2		
	Brand	Medicon	B. braun
Sr. #	Quoted Model	Various	Various
01	Country of Manufacturer	Germany	Germany
02	Manufacturing site	Germany	Germany
03	Compliance with defined Quality Standards	Compliance	Compliance
04	Specifications Compliance Feature Wise	Compliance	Compliance
05	Technical Eligibility of Product	Accepted	Accepted
06	Technical Eligibility of Firm	Accepted	Accepted
	Technical Status	Responsive	Responsive

Enduser (Concerned Department/ Co-opted Member)



Eng. Irfan Haider
Bio Medical Engineer/
Co-opted Member

Dr. Imran Saeed
Prof. of Paediatric ENT


Technical Advisory Committee (TAC)


Eng. Shujaat Ali
Bio Medical Engineer/ Technical
Officer


Muhammad Riaz Shahid
Incharge Pay Roll Section



Dr. Syed Ali Abu Zar
AMS (Coordination)


Dr. Faeza Hasnain
Chief Pharmacist


Dr. Muhammad Zeeshan Khan
Assistant Professor Paediatric
Medicine


Dr. Mumtaz Hussain
Assistant Professor Paediatric
Orthopedic Surgery



Dr. Muhammad Bilal Mirza
Associate Prof. of Pead
Surgery


Dr. Zia ur Rehman
Associate Prof. of Peads Neurology



Dr. Nazish Saqlan
Associate Prof. of Pathology
Department

Dr. Sarfraz Ahmad
Associate Prof. Paediatric
PM&R


Dr. Farah Haroon
Associate Prof. of Neonatology


Dr. Uzma Kazmi
Professor of Paediatric
Cardiology


Dr. Syed Muhammad Javed Iqbal
Prof. of Paediatric Medicine Unit-II


Dr. Junaid Rashid
Prof. of Paediatric Medicine / Chairman
Department of Paediatric Medicine

The Children's Hospital & University of Child Health Sciences Lahore

Technical Evaluation

Package No. / Bid No. PC/2020/23/CH & UCHS

Name of the Equipment:

Fess Equipment Function Endoscopic Sinus Surgery

Required Quantity:

1

Part - I

KNOCKED DOWN CRITERIA - (BID EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	Allmed Solutions
01	Complete Package / Tender (Mandatory)	Yes
02	Original Receipt of Tender (Mandatory)	Yes
03	Affidavit from Bidder	Yes
04	Bid Security CDR / Bank Guarantee (Mandatory)	Yes
05	Bid Validity (Mandatory)	Yes
06	Delivery Period (Mandatory)	Yes
Remarks:		Responsive

PART - II

KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	Allmed Solutions
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	Yes
02	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes
03	Up-to-Date After Sales Service Certificate from Manufacturer / Pncipal	Yes
04	Bidder Past Performance for Quoted Product in Teaching Hospitals	Yes
05	Availability of relevant Tools and Testing / Calibration Equipment	Yes
06	Warranty	Yes
Remarks:		Responsive

PART - III		
KNOCKED DOWN CRITERIA - (PRODCUT EVALUATION)		
Evaluated by (TAC) Committee		
Sr. #	Required Item Name	Allmed Solutions
23	Fess Equipment Function Endoscopic Sinus Surgery - 1	
	Brand	Karl Storz
	Quoted Model	7230AWA
01	Country of Manufacturer	Germany
02	Manufacturing site	Germany
03	Compliance with defined Quality Standards	Compliance
04	Specifications Compliance Feature Wise	Compliance
05	Technical Eligibility of Product	Accepted
06	Technical Eligibility of Firm	Accepted
Technical Status		Responsive

Enduser (Concerned Department/ Co-opted Member)



 Eng. Arfan Haider
 Bio Medical Engineer/
 Co-opted Member

Dr. Imran Saeed
 Prof. of Paediatric ENT

Technical Advisory Committee (TAC)


 Eng. Sanaat Ali
 Bio Medical Engineer/
 Technical Officer


Muhammad Riaz Shahid
 Incharge Pay Roll Section



 Dr. Syed Ali Abu Zar
 AMS (Coordination)


Dr. Faëza Hasnain
 Chief Pharmacist


 Dr. Muhammad Zeeshan Khan
 Assistant Professor Paediatric
 Medicine


Dr. Mumtaz Hussain
 Assistant Professor Paediatric
 Orthopedic Surgery



 Dr. Muhammad Bilal Mirza
 Associate Prof. of Pead Surgery



 Dr. Zia ur Rahman
 Associate Prof. of Peads
 Neurology



 Dr. Nazish Saqlan
 Associate Prof. of Pathology
 Department

Dr. Sarfraz Ahmad
 Associate Prof.
 Paediatric PM&R


 Dr. Farah Haroon
 Assocalte Prof. of Neonatology


 Dr. Uzma Kazmi
 Professor of Paediatric Cardiology


 Dr. Syed Muhammad Javed Iqbal
 Prof. of Paediatric Medicine Unit-II


 Dr. Junaid Rashid
 Prof. of Paediatric Medicine / Chairman
 Department of Paediatric Medicine

The Children's Hospital & University of Child Health Sciences Lahore

Technical Evaluation

Package No. / Bid No. PC/2020/24/CH & UCHS

Name of the Equipment:

Flexible Endoscope Equipment

Required Quantity:

1

Part - I

KNOCKED DOWN CRITERIA - (BID EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	Allmed Solutions	Mediland Pakistan (Pvt) Ltd
01	Complete Package / Tender (Mandatory)	Yes	Yes
02	Original Receipt of Tender (Mandatory)	Yes	Yes
03	Affidavit from Bidder	Yes	Yes
04	Bid Security CDR / Bank Guarantee (Mandatory)	Yes	Yes
05	Bid Validity (Mandatory)	Yes	Yes
06	Delivery Period (Mandatory)	Yes	Yes
Remarks:		Responsive	Responsive

PART - II

KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	Allmed Solutions	Mediland Pakistan (Pvt) Ltd
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	Yes	Yes
02	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal	Yes	Yes
04	Bidder Past Performance for Quoted Product in Teaching Hospitals	Yes	Yes
05	Availability of relevant Tools and Testing / Calibration Equipment	Yes	Yes
06	Warranty	Yes	Yes
Remarks:		Responsive	Responsive



PART - III			
KNOCKED DOWN CRITERIA - (PRODCUT EVALUATION)			
Evaluated by (TAC) Committee			
24	Required Item Name	Allmed Solutions	Mediland Pakistan (Pvt) Ltd
	Flexible Endoscope Equipment - 1		
	Brand	Karl Storz	Pentax
Sr. #	Quoted Model	11102CMM	VNL9-CP
01	Country of Manufacturer	Germany	Japan
02	Manufacturing site	Germany	Spain
03	Compliance with defined Quality Standards	Compliance	Compliance
04	Specifications Compliance Feature Wise	Compliance	Compliance
05	Technical Eligibility of Product	Accepted	Accepted
06	Technical Eligibility of Firm	Accepted	Accepted
	Technical Status	Responsive	Responsive

Enduser (Concerned Department/ Co-opted Member)

Engr. Imran Haider
Bio Medical Engineer/
Co-opted Member

Dr. Imran Saeed
Prof. of Paediatric ENT

Technical Advisory Committee (TAC)

Engr. Shujaat Ali
Bio Medical Engineer/
Technical Officer

Muhammad Riaz Shahid
Incharge Pay Roll Section

Dr. Syed Ali Abu Zar
AMS (Coordination)

Dr. Faeza Hasnain
Chief Pharmacist

Dr. Muhammad Zeeshan Khan
Assistant Professor Paediatric
Medicine

Dr. Mumtaz Hussain
Assistant Professor Paediatric
Orthopedic Surgery

Dr. Muhammad Bilal Mirza
Associate Prof. of Paed
Surgery

Dr. Zia ur Rehman
Associate Prof. of Paeds Neurology

Dr. Nazish Saqlan
Associate Prof. of Pathology
Department

Dr. Sarfraz Ahmad
Associate Prof. Paediatric
PM&R

Dr. Farah Haroon
Associate Prof. of Neonatology

Dr. Uzma Kazmi
Professor of Paediatric
Cardiology

Dr. Syed Muhammad Javed Iqbal
Prof. of Paediatric Medicine Unit-II

Dr. Junaid Rashid
Prof. of Paediatric Medicine / Chairman
Department of Paediatric Medicine

The Children's Hospital & University of Child Health Sciences Lahore

Technical Evaluation

Package No. / Bid No. PC/2020/25/CH & UCHS

Name of the Equipment:

ENT Treatment Unit

Required Quantity:

1

Part - I

KNOCKED DOWN CRITERIA - (BID EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	Modilan Pakistan (Pvt) Ltd
01	Complete Package / Tender (Mandatory)	Yes
02	Original Receipt of Tender (Mandatory)	Yes
03	Affidavit from Bidder	Yes
04	Bid Security CDR / Bank Guarantee (Mandatory)	Yes
05	Bid Validity (Mandatory)	Yes
06	Delivery Period (Mandatory)	Yes
Remarks:		Responsive

PART - II

KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	Medilan Pakistan (Pvt) Ltd
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	Yes
02	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal	Yes
04	Bidder Past Performance for Quoted Product in Teaching Hospitals	Yes
05	Availability of relevant Tools and Testing / Calibration Equipment	Yes
06	Warranty	Yes
Remarks:		Responsive




PART - III		
KNOCKED DOWN CRITERIA - (PRODCUT EVALUATION)		
Evaluated by (TAC) Committee		
25	Required Item Name	Medilan Pakistan (Pvt) Ltd
	ENT Treatment Unit - 1	
	Brand	Euro Clinic
Sr. #	Quoted Model	Otocompact Professional
01	Country of Manufacturer	Italy
02	Manufacturing site	Italy
03	Compliance with defined Quality Standards	Compliance
04	Specifications Compliance Feature Wise	Compliance
05	Technical Eligibility of Product	Accepted
06	Technical Eligibility of Firm	Accepted
Technical Status		Responsive

Enduser (Concerned Department/ Co-opted Member)


Engr. Irfan Haider
Bio Medical Engineer/
Co-opted Member

Dr. Imran Saeed
Prof. of Paediatric ENT


Technical Advisory Committee (TAC)


Engr. Shujaat Ali
Bio Medical Engineer/
Technical Officer


Muhammad Riaz Shahid
Incharge Pay Roll Section



Dr. Syed Ali Abu Zar
AMS (Coordination)


Dr. Faeza Hasnain
Chief Pharmacist


Dr. Muhammad Zeeshan Khan
Assistant Professor Paediatric
Medicine


Dr. Mumtaz Hussain
Assistant Professor Paediatric
Orthopedic Surgery



Dr. Muhammad Bilal Mirza
Associate Prof. of Pead Surgery



Dr. Zia ur Rehman
Associate Prof. of Peads
Neurology



Dr. Nazish Saqlan
Associate Prof. of Pathology
Department

Dr. Sarfraz Ahmad
Associate Prof. Paediatric
PM&R


Dr. Farah Haroon
Assocaite Prof. of Neonatology


Dr. Uzma Kazmi
Professor of Paediatric Cardiology


Dr. Syed Muhammad Javed Iqbal
Prof. of Paediatric Medicine Unit-II


Dr. Junaid Rashid
Prof. of Paediatric Medicine / Chairman
Department of Paediatric Medicine

The Children's Hospital & University of Child Health Sciences Lahore

Technical Evaluation

Package No. / Bid No. PC/2020/26/CH & UCHS

Name of the Equipment: Adeno Tonsillectomy Set

Required Quantity: 4

Part - I

KNOCKED DOWN CRITERIA - (BID EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	Vertex Medical (Pvt) Ltd
01	Complete Package / Tender (Mandatory)	Yes
02	Original Receipt of Tender (Mandatory)	Yes
03	Affidavit from Bidder	Yes
04	Bid Security CDR / Bank Guarantee (Mandatory)	Yes
05	Bid Validity (Mandatory)	Yes
06	Delivery Period (Mandatory)	Yes
Remarks:		Responsive

PART - II

KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	Vertex Medical (Pvt) Ltd
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	Yes
02	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes
03	Up-to-Date After Sales Service Certificate from Manufacturer /	Yes
04	Bidder Past Performance for Quoted Product in Teaching	Yes
05	Availability of relevant Tools and Testing / Calibration	Yes
06	Warranty	Yes
Remarks:		Responsive

PART - III		
KNOCKED DOWN CRITERIA - (PRODCUT EVALUATION)		
Evaluated by (TAC) Committee		
26	Required Item Name	Vertex Medical (Pvt) Ltd
	Adeno Tonsillectomy Set - 4	
	Brand	
Sr. #	Quoted Model	Various
01	Country of Manufacturer	Germany
02	Manufacturing site	Germany
03	Compliance with defined Quality Standards	Compliance
04	Specifications Compliance Feature Wise	Compliance
05	Technical Eligibility of Product	Accepted
06	Technical Eligibility of Firm	Accepted
Technical Status		Responsive

Enduser (Concerned Department/ Co-opted Member)



Engr. Irfan Haider
Bio Medical Engineer/
Co-opted Member

Dr. Imran Saeed
Prof. of Peadiatric ENT


Technical Advisory Committee (TAC)


Engr. Syed Ali
Bio Medical Engineer/
Technical Officer


Muhammad Riaz Shahid
Incharge Pay Roll Section



Dr. Syed Ali Abu Zar
AMS (Coordination)


Dr. Faeza Hasnain
Chief Pharmacist


Dr. Muhammad Zeeshan Khan
Assistant Professor Paediatric
Medicine


Dr. Mumtaz Hussain
Assistant Professor Paediatric
Orthopedic Surgery



Dr. Muhammad Bilal Mirza
Associate Prof. of Pead Surgery



Dr. Zia ur Rehman
Associate Prof. of Peads
Neurology



Dr. Nazish Saqlan
Associate Prof. of Pathology
Department

Dr. Sarfraz Ahmad
Associate Prof. Peadiatric
PM&R


Dr. Farah Haroon
Assocate Prof. of Neonatology


Dr. Uzma Kazmi
Professor of Paediatric Cardiology


Dr. Syed Muhammad Javed Iqbal
Prof. of Peadiatric Medicine Unit-II


Dr. Junaid Rashid
Prof. of Paediatric Medicine / Chairman
Department of Paediatric Medicine

The Children's Hospital & University of Child Health Sciences Lahore

Technical Evaluation

Package No. / Bid No. PC/2020/27/CH & UCHS	
Name of the Equipment:	Nasal Surgery Set
Required Quantity:	1

Part - I

KNOCKED DOWN CRITERIA - (BID EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	Vertex Medical (Pvt) Ltd	FDS (Pvt) Ltd
01	Complete Package / Tender (Mandatory)	Yes	Yes
02	Original Receipt of Tender (Mandatory)	Yes	Yes
03	Affidavit from Bidder	Yes	Yes
04	Bid Security CDR / Bank Guarantee (Mandatory)	Yes	Yes
05	Bid Validity (Mandatory)	Yes	Yes
06	Delivery Period (Mandatory)	Yes	Yes
Remarks:		Responsive	Responsive

PART - II

KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	Vertex Medical (Pvt) Ltd	FDS (Pvt) Ltd
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	Yes	Yes
02	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal	Yes	Yes
04	Bidder Past Performance for Quoted Product in Teaching Hospitals	Yes	Yes
05	Availability of relevant Tools and Testing / Calibration Equipment	Yes	Yes
06	Warranty	Yes	Yes
Remarks:		Responsive	Responsive


PART - III			
KNOCKED DOWN CRITERIA - (PRODCUT EVALUATION)			
Evaluated by (TAC) Committee			
1	Required Item Name	Vertex Medical (Pvt) Ltd	FDS (Pvt) Ltd
	Nasal Surgery Set - 1		
	Brand	B.Braun	Medicon
Sr. #	Quoted Model	Various	Various
01	Country of Manufacturer	Germany	Germany
02	Manufacturing site	Germany	Germany
03	Compliance with defined Quality Standards	Compliance	Compliance
04	Specifications Compliance Feature Wise	Compliance	Compliance
05	Technical Eligibility of Product	Accepted	Accepted
06	Technical Eligibility of Firm	Accepted	Accepted
Technical Status		Responsive	Responsive

Enduser (Concerned Department/ Co-opted Member



Engh Irfan Haider
Bio Medical Engineer/
Co-opted Member

Dr. Imran Saeed
Prof. of Peadiatric ENT


Technical Advisory Committee (TAC)


Engh Shujab Ali
Bio Medical Engineer/
Technical Officer


Muhammad Riaz Shahid
Incharge Pay Roll Section



Dr. Syed Ali Abu Zar
AMS (Coordination)


Dr. Faeza Hasnain
Chief Pharmacist


Dr. Muhammad Zeeshan Khan
Assistant Professor Paediatric Medicine


Dr. Mumtaz Hussain
Assistant Professor Paediatric
Orthopedic Surgery



Dr. Muhammad Bilal Mirza
Associate Prof. of Pead Surgery


Dr. Zia ur Rehman
Associate Prof. of Peads Neurology



Dr. Nazish Saqlian
Associate Prof. of
Pathology Department

Dr. Sarfraz Ahmad
Associate Prof. Peadiatric PM&R


Dr. Farah Haroon
Assocaite Prof. of Neonatology


Dr. Uzma Kazmi
Professor of Paediatric Cardiology


Dr. Syed Muhammad Javed Iqbal
Prof. of Peadiatric Medicine Unit-II


Dr. Junaid Rashid
Prof. of Paediatric Medicine / Chairman
Department of Paediatric Medicine

The Children's Hospital & University of Child Health Sciences Lahore

Technical Evaluation

Package No. / Bid No. PC/2020/28/CH & UCHS	
Name of the Equipment:	Shaver Debrider with Comonents
Required Quantity:	1

Part - I

KNOCKED DOWN CRITERIA - (BID EVALUATION)

Evaluated by (TAC) Committee


Sr. #	Evaluation Parameters	FDS (Pvt) Ltd	Kasbn International	Allmed Solutions	Popular International
01	Complete Package / Tender (Mandatory)	Yes	Yes	Yes	Yes
02	Original Receipt of Tender (Mandatory)	Yes	Yes	Yes	Yes
03	Affidavit from Bidder	Yes	Yes	Yes	Yes
04	Bid Security CDR / Bank Guarantee (Mandatory)	Yes	Yes	Yes	Yes
05	Bid Validity (Mandatory)	Yes	Yes	Yes	Yes
06	Delivery Period (Mandatory)	Yes	Yes	Yes	Yes
Remarks:		Responsive	Responsive	Responsive	Responsive

PART - II

KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)


Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	FDS (Pvt) Ltd	Kasbn International	Allmed Solutions	Popular International
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	Yes	Yes	Yes	Yes
02	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes	Yes	Yes
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal	Yes	Yes	Yes	Yes
04	Bidder Past Performance for Quoted Product in Teaching Hospitals	Yes	Yes	Yes	Yes
05	Availability of relevant Tools and Testing / Calibration Equipment	Yes	Yes	Yes	Yes
06	Warranty	Yes	Yes	Yes	Yes
Remarks:		Responsive	Responsive	Responsive	Responsive




PART - III					
KNOCKED DOWN CRITERIA - (PRODCUT EVALUATION)					
Evaluated by (TAC) Committee					
28	Required Item Name	FDS (Pvt) Ltd	Kasbn International	Alimed Solutions	Popular International
	Shaver Debrider with Components - 1				
	Brand	Conmed	Nouvag	Karl Storz	Medtronic
Sr. #	Quoted Model	Various	High Surge 30	UM600	Not Quoted
01	Country of Manufacturer	USA	Switzerland	Switzerland/Germany	USA
02	Manufacturing site	USA	Switzerland	Germany	USA
03	Compliance with defined Quality Standards	Compliance	Compliance	Compliance	Compliance
04	Specifications Compliance Feature Wise	Compliance	Compliance	Compliance	Compliance
05	Technical Eligibility of Product	Accepted	Accepted	Accepted	Accepted
06	Technical Eligibility of Firm	Accepted	Accepted	Accepted	Accepted
Technical Status		Responsive	Responsive	Responsive	Responsive

Enduser (Concerned Department/ Co-opted Member)



Engr. Irfan Haider
Bio Medical Engineer/
Co-opted Member

Dr. Imran Saeed
Prof. of Paediatric ENT


Technical Advisory Committee (TAC)


Engr. Shaukat Ali
Bio Medical Engineer/
Technical Officer


Muhammad Riaz Shahid
Incharge Pay Roll Section



Dr. Syed Ali Abu Zar
AMS (Coordination)


Dr. Faeza Hasnain
Chief Pharmacist


Dr. Muhammad Zeeshan Khan
Assistant Professor Paediatric
Medicine


Dr. Mumtaz Hussain
Assistant Professor Paediatric
Orthopedic Surgery



Dr. Muhammad Bilal Mirza
Associate Prof. of Paed
Surgery



Dr. Zia ur Rehman
Associate Prof. of Paeds
Neurology



Dr. Nazish Saqlan
Associate Prof. of Pathology
Department

Dr. Sarfraz Ahmad
Associate Prof. Paediatric
PM&R


Dr. Farah Haroon
Associate Prof. of Neonatology


Dr. Uzma Kazmi
Professor of Paediatric
Cardiology


Dr. Syed Muhammad Javed Iqbal
Prof. of Paediatric Medicine Unit-II


Dr. Junaid Rashid
Prof. of Paediatric Medicine / Chairman
Department of Paediatric Medicine

The Children's Hospital & University of Child Health Sciences Lahore

Technical Evaluation

Package No. / Bid No. PC/2020/29/CH & UCHS	
Name of the Equipment:	Stroboscope System
Required Quantity:	1

Part - I

KNOCKED DOWN CRITERIA - (BID EVALUATION)

Evaluated by (TAC) Committee


Sr. #	Evaluation Parameters	Allmed Solutions	Kasbn International
01	Complete Package / Tender (Mandatory)	Yes	Yes
02	Original Receipt of Tender (Mandatory)	Yes	Yes
03	Affidavit from Bidder	Yes	Yes
04	Bid Security CDR / Bank Guarantee (Mandatory)	Yes	Yes
05	Bid Validity (Mandatory)	Yes	Yes
06	Delivery Period (Mandatory)	Yes	Yes
Remarks:		Responsive	Responsive

PART - II

KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	Allmed Solutions	Kasbn International
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	Yes	Yes
02	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal	Yes	Yes
04	Bidder Past Performance for Quoted Product in Teaching Hospitals	Yes	No
05	Availability of relevant Tools and Testing / Calibration Equipment	Yes	Yes
06	Warranty	Yes	Yes
Remarks:		Responsive	Non-Responsive




PART - III			
KNOCKED DOWN CRITERIA - (PRODCUT EVALUATION)			
Evaluated by (TAC) Committee			
29	Required Item Name	Allmed Solutions	Kasbn International
	Stroboscope System - 1		
	Brand	Karl Storz	Optomic
Sr. #	Quoted Model	40160040	Strobolux III LED+HD
01	Country of Manufacturer	Germany	Spain
02	Manufacturing site	Germany	Spain
03	Compliance with defined Quality Standards	Compliance	
04	Specifications Compliance Feature Wise	Compliance	
05	Technical Eligibility of Product	Accepted	
06	Technical Eligibility of Firm	Accepted	
Technical Status		Responsive	Non-Responsive

Enduser (Concerned Department/ Co-opted Member)



Engr. Irfan Haider
Bio Medical Engineer/
Co-opted Member

Dr. Imran Saeed
Prof. of Paediatric ENT


Technical Advisory Committee (TAC)


Engr. Shumail Ali
Bio Medical Engineer/
Technical Officer


Muhammad Riaz Shahid
Incharge Pay Roll Section



Dr. Syed Ali Abu Zar
AMS (Coordination)


Dr. Faeza
Hasnain
Chief Pharmacist


Dr. Muhammad Zeeshan Khan
Assistant Professor Paediatric
Medicine


Dr. Mumtaz Hussain
Assistant Professor Paediatric
Orthopedic Surgery



Dr. Muhammad Bilal Mirza
Associate Prof. of Pead
Surgery



Dr. Zia ur Rehman
Associate Prof. of Peads Neurology



Dr. Nazish Saqlan
Associate Prof. of
Pathology Department

Dr. Sarfraz Ahmad
Associate Prof. Paediatric
PM&R


Dr. Farah Haroon
Associate Prof. of Neonatology


Dr. Uzma Kazmi
Professor of Paediatric
Cardiology


Dr. Syed Muhammad Javed Iqbal
Prof. of Paediatric Medicine Unit-II


Dr. Junaid Rashid
Prof. of Paediatric Medicine / Chairman
Department of Paediatric Medicine

The Children's Hospital & University of Child Health Sciences Lahore

Technical Evaluation

Package No. / Bid No. PC/2020/30/CH & UCHS

Name of the Equipment:

Syringe Pump with Docking Station

Required Quantity:

10 + 3

Part - I

KNOCKED DOWN CRITERIA - (BID EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	Hospicare Systems	Al-karem Medical Tech	Medical equipment & System	Total Technologies
01	Complete Package / Tender (Mandatory)	Yes	Yes	Yes	Yes
02	Original Receipt of Tender (Mandatory)	Yes	Yes	Yes	Yes
03	Affidavit from Bidder	Yes	Yes	Yes	Yes
04	Bid Security CDR / Bank Guarantee (Mandatory)	Yes	Yes	Yes	Yes
05	Bid Validity (Mandatory)	Yes	Yes	Yes	Yes
06	Delivery Period (Mandatory)	Yes	Yes	Yes	Yes
Remarks:		Responsive	Responsive	Responsive	Responsive

PART - II


KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	Hospicare Systems	Al-karem Medical Tech	Medical equipment & System	Total Technologies
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	Yes	Yes	Yes	Yes
02	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes	Yes	Yes
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal	Yes	Yes	Yes	Yes
04	Bidder Past Performance for Quoted Product in Teaching Hospitals	Yes	No	Yes	Yes
05	Availability of relevant Tools and Testing / Calibration Equipment	Yes	Yes	Yes	Yes
06	Warranty	Yes	Yes	Yes	Yes
Remarks:		Responsive	Non-Responsive	Responsive	Responsive


PART - III					
KNOCKED DOWN CRITERIA - (PRODCUT EVALUATION)					
Evaluated by (TAC) Committee					
30	Required Item Name	Hospicare Systems	Al-karem Medical Tech	Medical equipment & System	Total Technologies
		Syringe Pump with Docking Station - 10 + 3			
	Brand	B. Braun	Novatech	Fresenius	Terumo
Sr. #	Quoted Model	Perfusor Compact Plus	NTS-40	Agilia SP MC	TE-SS830
01	Country of Manufacturer	Germany	UK	Germany	Japan
02	Manufacturing site	Germany	China	France	Japan
03	Compliance with defined Quality Standards	Compliance		Compliance	Compliance
04	Specifications Compliance Feature Wise	Compliance		Compliance	Compliance
05	Technical Eligibility of Product	Accepted		Accepted	Accepted
06	Technical Eligibility of Firm	Accepted		Accepted	Accepted
Technical Status		Responsive	Non Responsive	Responsive	Responsive

Enduser (Concerned Department/ Co-opted Member)



Eng. Imran Haider
Bio Medical Engineer/
Co-opted Member

Dr. Imran Saeed
Prof. of Paediatric ENT


Technical Advisory Committee (TAC)


Engr. Shujat Ali
Bio Medical Engineer/
Technical Officer


Muhammad Riaz Shahid
Incharge Pay Roll Section



Dr. Syed Ali Abu Zar
AMS (Coordination)


Dr. Faeza
Hasnain
Chief Pharmacist


Dr. Muhammad Zeeshan Khan
Assistant Professor Paediatric Medicine


Dr. Mumtaz Hussain
Assistant Professor Paediatric
Orthopedic Surgery



Dr. Muhammad Bilal Mirza
Associate Prof. of Paed
Surgery


Dr. Zia ur Rehman
Associate Prof. of Paeds
Neurology



Dr. Nazish Saqlan
Associate Prof. of
Pathology Department

Dr. Sarfraz Ahmad
Associate Prof. Paediatric PM&R


Dr. Farah Haroon
Associate Prof. of Neonatology


Dr. Uzma Kazmi
Professor of Paediatric
Cardiology


Dr. Syed Muhammad Javed Iqbal
Prof. of Paediatric Medicine Unit-II


Dr. Junaid Rashid
Prof. of Paediatric Medicine / Chairman
Department of Paediatric Medicine

The Children's Hospital & University of Child Health Sciences Lahore

Technical Evaluation

Package No. / Bid No. PC/2020/33-37/CH & UCHS	33-37
Name of the Equipment:	33. Cystoscope Set for Neonates 34. Cystoscope Infants & pediatrics 35. Resectoscope for Neonate 36. Resectoscope for Infant & Pediatrics 37. Optical Urethrotome for Pediatrics (As a Package Items)
Required Quantity:	2+2+2+2+1

Part - I

KNOCKED DOWN CRITERIA - (BID EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	Allmed Solutions	Verizon
01	Complete Package / Tender (Mandatory)	Yes	Yes
02	Original Receipt of Tender (Mandatory)	Yes	Yes
03	Affidavit from Bidder	Yes	Yes
04	Bid Security CDR / Bank Guarantee (Mandatory)	Yes	Yes
05	Bid Validity (Mandatory)	Yes	Yes
06	Delivery Period (Mandatory)	Yes	Yes
Remarks:		Responsive	Responsive

PART - II

KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)

Evaluated by (TAC) Committee


Sr. #	Evaluation Parameters	Allmed Solutions	Verizon
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	Yes	Yes
02	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal	Yes	Yes
04	Bidder Past Performance for Quoted Product in Teaching Hospitals	Yes	Yes
05	Availability of relevant Tools and Testing / Calibration Equipment	Yes	Yes
06	Warranty	Yes	Yes
Remarks:		Responsive	Responsive

M. Naveed 12.3.2024.


PART - III			
KNOCKED DOWN CRITERIA - (PRODCUT EVALUATION)			
Evaluated by (TAC) Committee			
33-37	Required Item Name	Allmed Solutions	Verizon
	Cystoscope Set for Neonates / Resectoscope Set for Neonates / Recestoscope Set for Infants & pediatrics / Optical Urethrotome for Pediatrics / Cystoscope Set for Infants & pediatrics - 2+2+2+2+1		
	Brand	Karl Storz	R. Wolf
Sr. #	Quoted Model	Various	Various
01	Country of Manufacturer	Germany	Germany
02	Manufacturing site	Germany	Germany
03	Compliance with defined Quality Standards	Compliance	Compliance
04	Specifications Compliance Feature Wise	Compliance	Compliance
05	Technical Eligibility of Product	Accepted	Accepted
06	Technical Eligibility of Firm	Accepted	Accepted
Technical Status		Responsive	Responsive

Enduser (Concerned Department/ Co-opted Member)



Eng. Inan Haider
Bio-Medical Engineer/
Co-opted Member


12.3.2024
Dr. Ghulam Mujtaba Zafar
Associate Prof. Paediatric Urology


Technical Advisory Committee (TAC)


Eng. Shujda Ali
Bio Medical Engineer/
Technical Officer


Muhammad Riaz Shahid
Incharge Pay Roll Section



Dr. Syed Ali Abu Zar
AMS (Coordination)


Dr. Faeza Hasnain
Chief Pharmacist


Dr. Muhammad Zeeshan Khan
Assistant Professor Paediatric Medicine


Dr. Mumtaz Hussain
Assistant Professor
Paediatric Orthopedic
Surgery


Dr. Muhammad Bilal Mirza
Associate Prof. of Pead
Surgery



Dr. Zia ur Rehman
Associate Prof. of Peads Neurology



Dr. Nazish Saqlan
Associate Prof. of
Pathology Department

Dr. Sarfraz Ahmad
Associate Prof. Peadiatric PM&R


Dr. Farah Haroon
Associate Prof. of Neonatology


Dr. Uzma Kazmi
Professor of Paediatric Cardiology


Dr. Syed Muhammad Javed Iqbal
Prof. of Peadiatric Medicine Unit-II


Dr. Junaid Rashid
Prof. of Paediatric Medicine / Chairman
Department of Paediatric Medicine

The Children's Hospital & University of Child Health Sciences Lahore

Technical Evaluation

Package No. / Bid No. PC/2020/38/CH & UCHS

Name of the Equipment:

Cystoscope & Resectoscope for Adolescent

Required Quantity:

1

Part - I

KNOCKED DOWN CRITERIA - (BID EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	Verizon
01	Complete Package / Tender (Mandatory)	Yes
02	Original Receipt of Tender (Mandatory)	Yes
03	Affidavit from Bidder	Yes
04	Bid Security CDR / Bank Guarantee (Mandatory)	Yes
05	Bid Validity (Mandatory)	Yes
06	Delivery Period (Mandatory)	Yes
Remarks:		Responsive

PART - II

KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	Verizon
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	Yes
02	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal	Yes
04	Bidder Past Performance for Quoted Product in Teaching Hospitals	Yes
05	Availability of relevant Tools and Testing / Calibration Equipment	Yes
06	Warranty	Yes
Remarks:		Responsive





 M. T. Q. R.

M. A. Khan
 23-12-24


PART - III		
KNOCKED DOWN CRITERIA - (PRODCUT EVALUATION)		
Evaluated by (TAC) Committee		
38	Required Item Name	Verizon
	Cystoscope & Resectoscope for Adolescent - 1	
	Brand	R. Wolf
Sr. #	Quoted Model	
01	Country of Manufacturer	Germany
02	Manufacturing site	Germany
03	Compliance with defined Quality Standards	Compliance
04	Specifications Compliance Feature Wise	Compliance
05	Technical Eligibility of Product	Accepted
06	Technical Eligibility of Firm	Accepted
Technical Status		Responsive

Enduser (Concerned Department/ Co-opted Member)



Engr. Irfan Haider
Bio Medical Engineer/
Co-opted Member


12-3-2024
Dr. Ghulam Mujtaba Zafar
Associate Prof. Paediatric
Urology


Technical Advisory Committee (TAC)


Engr. Shujaat Ali
Bio Medical Engineer/
Technical Officer


Muhammad Riaz Shahid
Incharge Pay Roll Section



Dr. Syed Ali Abu Zar
AMS (Coordination)


Dr. Faeza Hasnain
Chief Pharmacist


Dr. Muhammad Zeeshan Khan
Assistant Professor Paediatric
Medicine


Dr. Mumtaz Hussain
Assistant Professor Paediatric
Orthopedic Surgery



Dr. Muhammad Bilal Mirza
Associate Prof. of Pead Surgery


Dr. Zia ur Rehman
Associate Prof. of Peads
Neurology



Dr. Nazish Saqlan
Associate Prof. of Pathology
Department

Dr. Sarfraz Ahmad
Associate Prof. Peadiatric
PM&R


Dr. Farah Haroon
Associate Prof. of Neonatology


Dr. Uzma Kazmi
Professor of Paediatric
Cardiology


Dr. Syed Muhammad Javed Iqbal
Prof. of Peadiatric Medicine Unit-II


Dr. Junaid Rashid
Prof. of Paediatric Medicine / Chairman
Department of Paediatric Medicine

The Children's Hospital & University of Child Health Sciences Lahore

Technical Evaluation

Package No. / Bid No. PC/2020/39/CH & UCHS

Name of the Equipment:

Irrigation pump with Pressure Monitor System for Endoscopy

Required Quantity:

1

Part - I

KNOCKED DOWN CRITERIA - (BID EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	Allmed Solutions	Verizon
01	Complete Package / Tender (Mandatory)	Yes	Yes
02	Original Receipt of Tender (Mandatory)	Yes	Yes
03	Affidavit from Bidder	Yes	Yes
04	Bid Security CDR / Bank Guarantee (Mandatory)	Yes	Yes
05	Bid Validity (Mandatory)	Yes	Yes
06	Delivery Period (Mandatory)	Yes	Yes
Remarks:		Responsive	Responsive

PART - II

KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	Allmed Solutions	Verizon
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	Yes	Yes
02	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal	Yes	Yes
04	Bidder Past Performance for Quoted Product in Teaching Hospitals	Yes	Yes
05	Availability of relevant Tools and Testing / Calibration Equipment	Yes	Yes
06	Warranty	Yes	Yes
Remarks:		Responsive	Responsive




 f, lb, Z, WE, M. W. E. F. 91, 2


PART - III			
KNOCKED DOWN CRITERIA - (PRODCUT EVALUATION)			
Evaluated by (TAC) Committee			
39	Required Item Name	Allmed Solutions	Verizon
	Irrigation pump with Pressure Monitor System for Endoscopy - 1		
	Brand	Karl Storz	R. Wolf
Sr. #	Quoted Model	UP220	22250011
01	Country of Manufacturer	Switzerland / Germany	Germany
02	Manufacturing site	Switzerland / Germany	Germany
03	Compliance with defined Quality Standards	Compliance	Compliance
04	Specifications Compliance Feature Wise	Compliance	Compliance
05	Technical Eligibility of Product	Compliance	Compliance
06	Technical Eligibility of Firm	Compliance	Compliance
Technical Status		Responsive	Responsive

Enduser (Concerned Department/ Co-opted Member)



 Eng. Irfan Haider
 Bio Medical Engineer/
 Co-opted Member


 Dr. Ghulam Mujtaba Zafar
 Associate Prof. Paediatric Urology


Technical Advisory Committee (TAC)


 Eng. Shaza Ali
 Bio Medical Engineer/
 Technical Officer


Muhammad Riaz Shahid
 Incharge Pay Roll Section



 Dr. Syed Ali Abu Zar
 AMS (Coordination)


Dr. Faeza
 Hasnain
 Chief Pharmacist


 Dr. Muhammad Zeeshan Khan
 Assistant Professor Paediatric
 Medicine


Dr. Mumtaz Hussain
 Assistant Professor Paediatric
 Orthopedic Surgery



 Dr. Muhammad Bilal Mirza
 Associate Prof. of Pead
 Surgery



 Dr. Zia ur Rehman
 Associate Prof. of Peads Neurology



 Dr. Nazish Saqlan
 Associate Prof. of
 Pathology Department

Dr. Sarfraz Ahmad
 Associate Prof. Peadiatric
 PM&R


 Dr. Farah Haroon
 Associate Prof. of Neonatology


 Dr. Uzma Kazmi
 Professor of Paediatric
 Cardiology


 Dr. Syed Muhammad Javed Iqbal
 Prof. of Paediatric Medicine Unit-II


 Dr. Junaid Rashid
 Prof. of Paediatric Medicine / Chairman
 Department of Paediatric Medicine

The Children's Hospital & University of Child Health Sciences Lahore

Technical Evaluation

Package No. / Bid No. PC/2020/40/CH & UCHS	
Name of the Equipment:	Pediatric Mini Nephroscope Set
Required Quantity:	2

Part - I

KNOCKED DOWN CRITERIA - (BID EVALUATION)

Evaluated by (TAC) Committee


Sr. #	Evaluation Parameters	Allmed Solutions	Verizon
01	Complete Package / Tender (Mandatory)	Yes	Yes
02	Original Receipt of Tender (Mandatory)	Yes	Yes
03	Affidavit from Bidder	Yes	Yes
04	Bid Security CDR / Bank Guarantee (Mandatory)	Yes	Yes
05	Bid Validity (Mandatory)	Yes	Yes
06	Delivery Period (Mandatory)	Yes	Yes
Remarks:		Responsive	Responsive

PART - II

KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)

Evaluated by (TAC) Committee


Sr. #	Evaluation Parameters	Allmed Solutions	Verizon
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	Yes	Yes
02	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal	Yes	Yes
04	Bidder Past Performance for Quoted Product in Teaching Hospitals	Yes	Yes
05	Availability of relevant Tools and Testing / Calibration Equipment	Yes	Yes
06	Warranty	Yes	Yes
Remarks:		Responsive	Responsive




 M. N. Khan
 12.3.2020


PART - III			
KNOCKED DOWN CRITERIA - (PRODCUT EVALUATION)			
Evaluated by (TAC) Committee			
40	Required Item Name	Allmed Solutions	Verizon
	Pediatric Mini Nephroscope Set -2		
	Brand		
Sr. #	Quoted Model	27830KAK	8968.421
01	Country of Manufacturer	Germany	Germany
02	Manufacturing site	Germany	Germany
03	Compliance with defined Quality Standards	Compliance	Compliance
04	Specifications Compliance Feature Wise	Compliance	Compliance
05	Technical Eligibility of Product	Accepted	Accepted
06	Technical Eligibility of Firm	Accepted	Accepted
Technical Status		Responsive	Responsive

Enduser (Concerned Department/ Co-opted Member)



Engr. Irfan Haider
Bio-Medical Engineer/
Co-opted Member


12-3-2024
Dr. Ghulam Mujtaba Zafar
Associate Prof. Paediatric Urology


Technical Advisory Committee (TAC)


Engr. Saikat Ali
Bio Medical Engineer/
Technical Officer


Muhammad Riaz Shahid
Incharge Pay Roll Section



Dr. Syed Ali Abu Zar
AMS (Coordination)


Dr. Faeza Hasnain
Chief Pharmacist


Dr. Muhammad Zeeshan Khan
Assistant Professor Paediatric
Medicine


Dr. Mumtaz Hussain
Assistant Professor Paediatric
Orthopedic Surgery



Dr. Muhammad Bilal Mirza
Associate Prof. of Pead Surgery



Dr. Zia ur Rehman
Associate Prof. of Peads Neurology



Dr. Nazish Saqlan
Associate Prof. of
Pathology Department

Dr. Sarfraz Ahmad
Associate Prof. Paediatric
PM&R


Dr. Farah Haroon
Assocaite Prof. of Neonatology


Dr. Uzma Kazmi
Professor of Paediatric Cardiology


Dr. Syed Muhammad Javed Iqbal
Prof. of Peadiatric Medicine Unit-II


Dr. Junaid Rashid
Prof. of Paediatric Medicine / Chairman
Department of Paediatric Medicine

The Children's Hospital & University of Child Health Sciences Lahore

Technical Evaluation

Package No. / Bid No. PC/2020/41/CH & UCHS

Name of the Equipment:

Pediatric Micro PCNL Set

Required Quantity:

5

Part - I

KNOCKED DOWN CRITERIA - (BID EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	Allmed Solutions
01	Complete Package / Tender (Mandatory)	Yes
02	Original Receipt of Tender (Mandatory)	Yes
03	Affidavit from Bidder	Yes
04	Bid Security CDR / Bank Guarantee (Mandatory)	Yes
05	Bid Validity (Mandatory)	Yes
06	Delivery Period (Mandatory)	Yes
Remarks:		Responsive

PART - II

KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)

Evaluated by (TAC) Committee


Sr. #	Evaluation Parameters	Allmed Solutions
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	Yes
02	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal	Yes
04	Bidder Past Performance for Quoted Product in Teaching Hospitals	Yes
05	Availability of relevant Tools and Testing / Calibration Equipment	Yes
06	Warranty	Yes
Remarks:		Responsive




 M. I. Q. 16/12/2021


PART - III			
KNOCKED DOWN CRITERIA - (PRODCUT EVALUATION)			
Evaluated by (TAC) Committee			
	Required Item Name	Allmed Solutions	
41	Pediatric Micro PCNL Set - 5		
	Brand	Karl Storz	
Sr. #	Quoted Model	27820KAK	
01	Country of Manufacturer	Germany	
02	Manufacturing site	Germany	
03	Compliance with defined Quality Standards	Compliance	
04	Specifications Compliance Feature Wise	Compliance	
05	Technical Eligibility of Product	Accepted	
06	Technical Eligibility of Firm	Accepted	
Technical Status		Responsive	


Enduser (Concerned Department/ Co-opted Member)


Eng. Aman Haider
Bio Medical Engineer/
Co-opted Member



12-3-2024
Dr. Ghulam Mujtaba Zafar
Associate Prof. Paediatric
Urology


Technical Advisory Committee (TAC)



Engr. Shujat Ali
Bio Medical Engineer/
Technical Officer



Muhammad Riaz Shahid
Incharge Pay Roll Section



Dr. Syed Ali Abu Zar
AMS (Coordination)



Dr. Faeza Hasnain
Chief Pharmacist



Dr. Muhammad Zeeshan Khan
Assistant Professor Paediatric
Medicine



Dr. Mumtaz Hussain
Assistant Professor Paediatric
Orthopedic Surgery



Dr. Muhammad Bilal Mirza
Associate Prof. of Pead Surgery



Dr. Zia ur Rehman
Associate Prof. of Peads
Neurology



Dr. Nazish Saqlan
Associate Prof. of Pathology
Department


Dr. Sarfraz Ahmad
Associate Prof. Peadiatric
PM&R


Dr. Farah Haroon
Associate Prof. of Neonatology


Dr. Uzma Kazmi
Professor of Paediatric
Cardiology


Dr. Syed Muhammad Javed Iqbal
Prof. of Paediatric Medicine Unit-II


Dr. Junaid Rashid
Prof. of Paediatric Medicine / Chairman
Department of Paediatric Medicine

The Children's Hospital & University of Child Health Sciences Lahore

Technical Evaluation

Package No. / Bid No. PC/2020/42/CH & UCHS

Name of the Equipment:	Peadtric Uretroscope -Renoscope 4.5-7.5
Required Quantity:	2-Nos.

Part - I

KNOCKED DOWN CRITERIA - (BID EVALUATION)

Evaluated by (TAC) Committee


Sr. #	Evaluation Parameters	Allmed Solution	Verizone
01	Complete Package / Tender (Mandatory)	Yes	Yes
02	Original Receipt of Tender (Mandatory)	Yes	Yes
03	Affidavit from Bidder	Yes	Yes
04	Bid Security CDR / Bank Guarantee (Mandatory)	Yes	Yes
05	Bid Validity (Mandatory)	Yes	Yes
06	Delivery Period (Mandatory)	Yes	Yes
Remarks:		Responsive	Responsive

PART - II

KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	Allmed Solution	Verizone
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	Yes	Yes
02	Technical & Engineering capability	Yes	Yes
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal	Yes	Yes
04	Bidder Past Performance for Quoted Product in Teaching Hospitals	Yes	Yes
05	Availability of relevant Tools and Testing / Calibration Equipment	Yes	Yes
06	Warranty	Yes	Yes
Remarks:		Responsive	Responsive




 M. Nadeem
 12-3-2021

PART - III			
KNOCKED DOWN CRITERIA - (PRODCUT EVALUATION)			
Evaluated by (TAC) Committee			
42	Required Item Name	Allmed Solution	Verizone
	Brand	Karlz Storz	R.Wolf
Sr. #	Quoted Model		
01	Country of Manufacturer	Germany	Germany
02	Manufacturing site	Germany	Germany
03	Compliance with defined Quality Standards	Compilance	Compliance
04	Specifications Compliance Feature Wise	Compliance	Compliance
05	Technical Eligibility of Product	Accepted	Accepted
06	Technical Eligibility of Firm	Accepted	Accepted
Technical Status		Responsive	Responsive

Enduser (Concerned Department/ Co-opted Member)



Engr. Irfan Haider
Bio Medical Engineer/
Co-opted Member


Dr. Ghulam Mujtaba Zafar
Associate Prof. Paediatric Urology

Technical Advisory Committee (TAC)


Engr. Shujat Ali
Bio Medical Engineer/
Technical Officer


Muhammad Riaz Shahid
Incharge Pay Roll Section



Dr. Syed Ali Abu Zar
AMS (Coordination)


Dr. Faeza Hasnain
Chief Pharmacist


Dr. Muhammad Zeeshan Khan
Assistant Professor Paediatric
Medicine


Dr. Mumtaz Hussain
Assistant Professor Paediatric
Orthopedic Surgery


Dr. Muhammad Bilal Mirza
Associate Prof. of Pead Surgery


Dr. Zia ur Rehman
Associate Prof. of Peads Neurology



Dr. Nazish Saqlian
Associate Prof. of
Pathology Department

Dr. Sarfraz Ahmad
Associate Prof. Paediatric
PM&R


Dr. Farah Haroon
Assocaite Prof. of Neonatology


Dr. Uzma Kazmi
Professor of Paediatric Cardiology


Dr. Syed Muhammad Javed Iqbal
Prof. of Paediatric Medicine Unit-II


Dr. Junaid Rashid
Prof. of Paediatric Medicine / Chairman
Department of Paediatric Medicine

The Children's Hospital & University of Child Health Sciences Lahore

Technical Evaluation

Package No. / Bid No. PC/2020/43/CH & UCHS		
Name of the Equipment:		Peadtric Uretroscope -Renoscope 6-9.5
Required Quantity:		2-Nos.

Part - I

KNOCKED DOWN CRITERIA - (BID EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	Allmed Solution	Verizone
01	Complete Package / Tender (Mandatory)	Yes	Yes
02	Original Receipt of Tender (Mandatory)	Yes	Yes
03	Affidavit from Bidder	Yes	Yes
04	Bid Security CDR / Bank Guarantee (Mandatory)	Yes	Yes -
05	Bid Validity (Mandatory)	Yes	Yes
06	Delivery Period (Mandatory)	Yes	Yes
Remarks:		Responsive	Responsive

PART - II

KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	Allmed Solution	Verizone
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	Yes	Yes
02	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal	Yes	Yes
04	Bidder Past Performance for Quoted Product in Teaching Hospitals	Yes	Yes
05	Availability of relevant Tools and Testing / Calibration Equipment	Yes	Yes
06	Warranty	Yes	Yes
Remarks:		Responsive	Responsive



 M. H. E. F. A.

 M. Ahsan

 12-03-2024

PART - III			
KNOCKED DOWN CRITERIA - (PRODCUT EVALUATION)			
Evaluated by (TAC) Committee			
43	Required Item Name	Allmed Solution	Verizone
	Brand	Karlz Storz	R.Wolf
Sr. #	Quoted Model		
01	Country of Manufacturer	Germany	Germany
02	Manufacturing site	Germany	Germany
03	Compliance with defined Quality Standards	Compliance	Compliance
04	Specifications Compliance Feature Wise	Compliance	Compliance
05	Technical Eligibility of Product	Accepted	Accepted
06	Technical Eligibility of Firm	Accepted	Accepted
Technical Status		Responsive	Responsive

Enduser (Concerned Department/ Co-opted Member)

M. Naveed
12-3-2024

Dr. Ghulam Mujtaba Zafar
Associate Prof. Paediatric Urology

Technical Advisory Committee (TAC)

Irfan Haider
Eng. Irfan Haider
Bio Medical Engineer/
Co-opted Member

Muhammad Riaz Shahid
Eng. Muhammad Riaz
Bio Medical Engineer/
Technical Officer

Muhammad Riaz Shahid
Incharge Pay Roll Section

Syed Ali Abu Zar
Dr. Syed Ali Abu Zar
AMS (Coordination)

Faeza Hasnain
Dr. Faeza Hasnain
Chief Pharmacist

M. Zeeshan Khan
Dr. Muhammad Zeeshan Khan
Assistant Professor Paediatric
Medicine

Mumtaz Husain
Dr. Mumtaz Husain
Assistant Professor Paediatric
Orthopedic Surgery

Muhammad Bilal Mirza
Dr. Muhammad Bilal Mirza
Associate Prof. of Pead
Surgery

Zia ur Rehman
Dr. Zia ur Rehman
Associate Prof. of Peads
Neurology

Hazish Saqlan
Dr. Hazish Saqlan
Associate Prof. of Pathology
Department

Sarfraz Ahmad
Dr. Sarfraz Ahmad
Associate Prof. Peadiatric
PM&R

Farah Haroon
Dr. Farah Haroon
Associate Prof. of Neonatology

Uzma Kazmi
Dr. Uzma Kazmi
Professor of Paediatric
Cardiology

Syed Muhammad Javed Iqbal
Dr. Syed Muhammad Javed Iqbal
Prof. of Paediatric Medicine Unit-II

Junaid Rashid
Dr. Junaid Rashid
Prof. of Paediatric Medicine / Chairman
Department of Paediatric Medicine

The Children's Hospital & University of Child Health Sciences Lahore

Technical Evaluation

Package No. / Bid No. PC/2020/44/CH & UCHS

Name of the Equipment:

Flexible Digital Uretroscope 6-0 FR (URS)

Required Quantity:

5

Part - I

KNOCKED DOWN CRITERIA - (BID EVALUATION)

Evaluated by (TAC) Committee


Sr. #	Evaluation Parameters	Ontech Co.
01	Complete Package / Tender (Mandatory)	Yes
02	Original Receipt of Tender (Mandatory)	Yes
03	Affidavit from Bidder	Yes
04	Bid Security CDR / Bank Guarantee (Mandatory)	Yes
05	Bid Validity (Mandatory)	Yes
06	Delivery Period (Mandatory)	Yes
Remarks:		Responsive

PART - II

KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)

Evaluated by (TAC) Committee


Sr. #	Evaluation Parameters	Ontech Co.
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	No
02	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal	No
04	Bidder Past Performance for Quoted Product in Teaching Hospitals	No
05	Availability of relevant Tools and Testing / Calibration Equipment	Yes
06	Warranty	Yes
Remarks:		Non-Responsive




 M. Arshad
 12-3-2024
 2 JH

PART - III		
KNOCKED DOWN CRITERIA - (PRODCUT EVALUATION)		
Evaluated by (TAC) Committee		
44	Required Item Name	Ontech Co.
	Flexible Digital Uretroscope G O FR (URS) - 5	
	Brand	Hugemed
Sr. #	Quoted Model	Hugemed
01	Country of Manufacturer	China
02	Manufacturing site	China
03	Compliance with defined Quality Standards	
04	Specifications Compliance Feature Wise	
05	Technical Eligibility of Product	
06	Technical Eligibility of Firm	
Technical Status		Non-Responsive

Enduser (Concerned Department/ Co-opted Member)



Eng. Irfan Haider
Bio Medical Engineer/
Co-opted Member


12.3.18/21
Dr. Ghulam Mujtaba Zafar
Associate Prof. Paediatric
Urology


Technical Advisory Committee (TAC)


Eng. Shaukat Ali
Bio Medical Engineer/
Technical Officer


Muhammad Riaz Shahid
Incharge Pay Roll Section



Dr. Syed Ali Abu Zar
AMS (Coordination)


Dr. Faeza Hasnain
Chief Pharmacist


Dr. Muhammad Zeeshan Khan
Assistant Professor Paediatric
Medicine


Dr. Mumtaz Hussain
Assistant Professor Paediatric
Orthopedic Surgery


Dr. Muhammad Bilal Mirza
Associate Prof. of Pead Surgery



Dr. Zia ur Rehman
Associate Prof. of Peads
Neurology



Dr. Harish Saqlan
Associate Prof. of Pathology
Department

Dr. Sarfraz Ahmad
Associate Prof. Peadiatric
PM&R


Dr. Farah Haroon
Associate Prof. of Neonatology


Dr. Uzma Kazmi
Professor of Paediatric Cardiology


Dr. Syed Muhammad Javed Iqbal
Prof. of Peadiatric Medicine Unit-II


Dr. Junaid Rashid
Prof. of Paediatric Medicine / Chairman
Department of Paediatric Medicine

The Children's Hospital & University of Child Health Sciences Lahore

Technical Evaluation

Package No. / Bid No. PC/2020/45/CH & UCHS	
Name of the Equipment:	Pediatric Laparoscope Set / System (4K System) having 1. Tower 2. Instrument Set
Required Quantity:	1

Part - I

KNOCKED DOWN CRITERIA - (BID EVALUATION)

Evaluated by (TAC) Committee


Sr. #	Evaluation Parameters	Vertex Medical (Pvt) ltd	Verizon	Allmed Solutions
01	Complete Package / Tender (Mandatory)	Yes	Yes	Yes
02	Original Receipt of Tender (Mandatory)	Yes	Yes	Yes
03	Affidavit from Bidder	Yes	Yes	Yes
04	Bid Security CDR / Bank Guarantee (Mandatory)	Yes	Yes	Yes
05	Bid Validity (Mandatory)	Yes	Yes	Yes
06	Delivery Period (Mandatory)	Yes	Yes	Yes
Remarks:		Responsive	Responsive	Responsive

PART - II

KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	Vertex Medical (Pvt) ltd	Verizon	Allmed Solutions
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	Yes	Yes	Yes
02	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes	Yes
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal	Yes	Yes	Yes
04	Bidder Past Performance for Quoted Product in Teaching Hospitals	No	Yes	Yes
05	Availability of relevant Tools and Testing / Calibration Equipment	Yes	Yes	Yes
06	Warranty	Yes	Yes	Yes
Remarks:		Non Responsive	Responsive	Responsive



 M. T. R. 6
 2023-3-24

PART - III								
KNOCKED DOWN CRITERIA - (PRODCUT EVALUATION)								
Evaluated by (TAC) Committee								
45	Required Item Name		Vertex Medical (Pvt) ltd	Verizon	Allmed Solutions			
	Pediatic Laparoscope Set / System (4K System) having 1. Tower 2. Instrument Set - 1							
	Brand					B.Braun	R. Wolf	Karl Storz
	Camera Head					PV485	logic 4k85525942	TH120
	Monitor					PV628	LMD-X310NB	TM343
	Endoflator					PG150	22350111	UI500S1
Sr. #	Quoted Model	Asculap AG	R. Wolf	26003AA				
01	Country of Manufacturer	Germany	Germany	Germany				
02	Manufacturing site	Germany	Germany	Germany				
03	Compliance with defined Quality Standards	Compliance	Compliance	Compliance				
04	Specifications Compliance Feature Wise	Non Compliance	Compliance	Compliance				
05	Technical Eligibility of Product	Not Accepted	Accepted	Accepted				
06	Technical Eligibility of Firm	Not Accepted	Accepted	Accepted				
Technical Status		Non-Responsive	Responsive	Responsive				

Enduser (Concerned Department/ Co-opted Member)

Engr. Irfan Haider
Bio Medical Engineer/
Co-opted Member

Dr. Ghulam Mujtaba Zafar
Associate Prof. Paediatric
Urology

Technical Advisory Committee (TAC)

Engr. Saqib Ali
Bio Medical Engineer/
Technical Officer

Muhammad Riaz Shahid
Incharge Pay Roll Section

Dr. Syed Ali Abu Zar
AMS (Coordination)

Dr. Faeza Hasnain
Chief Pharmacist

Dr. Muhammad Zeeshan Khan
Assistant Professor Paediatric
Medicine

Dr. Mumtaz Hussain
Assistant Professor
Paediatric Orthopedic
Surgery

Dr. Muhammad Bilal Mirza
Associate Prof. of Pead Surgery

Dr. Zia ur Rehman
Associate Prof. of Peads Neurology

Dr. Nazish Saqlian
Associate Prof. of Pathology
Department

Dr. Sarfraz Ahmad
Associate Prof. Peadiatric
PM&R

Dr. Farah Haroon
Assocaite Prof. of Neonatology

Dr. Uzma Kazmi
Professor of Paediatric Cardiology

Dr. Syed Muhammad Javed Iqbal
Prof. of Peadiatric Medicine Unit-II

Dr. Junaid Rashid
Prof. of Paediatric Medicine / Chairman
Department of Paediatric Medicine

The Children's Hospital & University of Child Health Sciences Lahore

Technical Evaluation

Package No. / Bid No. PC/2020/46/CH & UCHS	
Name of the Equipment:	Advance Bipolar Diathermy
Required Quantity:	1

Part - I

KNOCKED DOWN CRITERIA - (BID EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	Vertex Medical (Pvt) Ltd	Radiant Medical (Pvt) Ltd	Mediland Pakistan (Pvt) Ltd
01	Complete Package / Tender (Mandatory)	Yes	Yes	Yes
02	Original Receipt of Tender (Mandatory)	Yes	Yes	Yes
03	Affidavit from Bidder	Yes	Yes	Yes
04	Bid Security CDR / Bank Guarantee (Mandatory)	Yes	Yes	Yes
05	Bid Validity (Mandatory)	Yes	Yes	Yes
06	Delivery Period (Mandatory)	Yes	Yes	Yes
Remarks:		Responsive	Responsive	Responsive

PART - II

KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)


Evaluated by (TAC) Committee


Sr. #	Evaluation Parameters	Vertex Medical (Pvt) Ltd	Radiant Medical (Pvt) Ltd	Mediland Pakistan (Pvt) Ltd
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	Yes	Yes	Yes
02	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes	Yes
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal	Yes	Yes	Yes
04	Bidder Past Performance for Quoted Product in Teaching Hospitals	Yes	NO	No
05	Availability of relevant Tools and Testing / Calibration Equipment	Yes	Yes	Yes
06	Warranty	Yes	Yes	Yes
Remarks:		Reponsive	Non-Reponsive	Non-Reponsive

M. Adnan
12.5.2024


PART - III				
KNOCKED DOWN CRITERIA - (PRODCUT EVALUATION)				
Evaluated by (TAC) Committee				
46	Required Item Name	Vertex Medical (Pvt) Ltd	Radiant Medical (Pvt) Ltd	Mediland Pakistan (Pvt) Ltd
	Advance Bipolar Diathermy -1			
	Brand	Bova	Lamidey Noury	KLS Martin
Sr. #	Quoted Model	ARC-350	Optima	Maxium
01	Country of Manufacturer	Germany	France	Germany
02	Manufacturing site	Germany	France	Germany
03	Compliance with defined Quality Standards	Compliance		
04	Specifications Compliance Feature Wise	Compliance		
05	Technical Eligibility of Product	Accepted		
06	Technical Eligibility of Firm	Accepted		
Technical Status		Responsive	Non-Reponsive	Non-Reponsive

Enduser (Concerned Department/ Co-opted Member)



Engr. Arfan Halder
Bio Medical Engineer/
Co-opted Member


Dr. Ghulam Mujtaba Zafar
Associate Prof. Paediatric
Urology


Technical Advisory Committee (TAC)


Engr. Srujeet Ali
Bio Medical Engineer/
Technical Officer


Muhammad Riaz Shahid
Incharge Pay Roll Section



Dr. Syed Ali Abu Zar
AMS (Coordination)


Dr. Faeza Hasnain
Chief Pharmacist


Dr. Muhammad Zeeshan Khan
Assistant Professor Paediatric
Medicine


Dr. Mumtaz Hussain
Assistant Professor Paediatric
Orthopedic Surgery


Dr. Muhammad Bilal Mirza
Associate Prof. of Pead Surgery



Dr. Zia ur Rehman
Associate Prof. of Peads
Neurology



Dr. Nazish Saqlan
Associate Prof. of Pathology
Department

Dr. Sarfraz Ahmad
Associate Prof. Peadiatric
PM&R


Dr. Farah Haroon
Assocaite Prof. of Neonatology


Dr. Uzma Kazmi
Professor of Paediatric Cardiology


Dr. Syed Muhammad Javed Iqbal
Prof. of Peadiatric Medicine Unit-II


Dr. Junaid Rashid
Prof. of Paediatric Medicine / Chairman
Department of Paediatric Medicine

The Children's Hospital & University of Child Health Sciences Lahore

Technical Evaluation

Package No. / Bid No. PC/2020/47/CH & UCHS	
Name of the Equipment:	Mobile Ultrasound Machine with Doppler
Required Quantity:	1

Part - I

KNOCKED DOWN CRITERIA - (BID EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	Vertex Medical (Pvt) Ltd	Medical Equipment & Systems
01	Complete Package / Tender (Mandatory)	Yes	Yes
02	Original Receipt of Tender (Mandatory)	Yes	Yes
03	Affidavit from Bidder	Yes	Yes
04	Bid Security CDR / Bank Guarantee (Mandatory)	Yes	Yes
05	Bid Validity (Mandatory)	Yes	Yes
06	Delivery Period (Mandatory)	Yes	Yes
Remarks:		Responsive	Responsive

PART - II

KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	Vertex Medical (Pvt) Ltd	Medical Equipment & Systems
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	Yes	Yes
02	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal	Yes	Yes
04	Bidder Past Performance for Quoted Product in Teaching Hospitals	Yes	Yes
05	Availability of relevant Tools and Testing / Calibration Equipment	Yes	Yes
06	Warranty	Yes	Yes
Remarks:		Responsive	Responsive

4

Z lb

M. Z. E. F


d

for M. A. Arani
12.3.2024
2 yf

PART - III			
KNOCKED DOWN CRITERIA - (PRODCUT EVALUATION)			
Evaluated by (TAC) Committee			
47	Required Item Name	Vertex Medical (Pvt) Ltd	Medical Equipment & Systems
	Mobile Ultrasound Machine with Doppler - 1		
	Brand	Siemens	Cannon
Sr. #	Quoted Model	Acuson NX3	Aplio Flex
01	Country of Manufacturer	Germany	Japan
02	Manufacturing site	Korea	Japan
03	Compliance with defined Quality Standards	Compliance	Compliance
04	Specifications Compliance Feature Wise	Compliance	Compliance
05	Technical Eligibility of Product	Accepted	Accepted
06	Technical Eligibility of Firm	Accepted	Accepted
Technical Status		Responsive	Responsive

Enduser (Concerned Department/ Co-opted Member)


Engg. Irfan Haider
Bio Medical Engineer/
Co-opted Member


12-3-2024
Dr. Ghulam Mujtaba Zafar
Associate Prof. Paediatric Urology


Technical Advisory Committee (TAC)


Engg. Shujaat Ali
Bio Medical Engineer/
Technical Officer


Muhammad Riaz Shahid
Incharge Pay Roll Section



Dr. Syed Ali Abu Zar
AMS (Coordination)


Dr. Faeza Hasnain
Chief Pharmacist


Dr. Muhammad Zeeshan Khan
Assistant Professor Paediatric Medicine


Dr. Mumtaz Hussain
Assistant Professor Paediatric
Orthopedic Surgery



Dr. Muhammad Bilal Mirza
Associate Prof. of Pead
Surgery



Dr. Zia ur Rehman
Associate Prof. of Peads Neurology



Dr. Nazish Saqlan
Associate Prof. of
Pathology Department

Dr. Sarfraz Ahmad
Associate Prof. Peadiatric
PM&R


Dr. Farah Haroon
Assocaite Prof. of Neonatology


Dr. Uzma Kazmi
Professor of Paediatric
Cardiology


Dr. Syed Muhammad Javed Iqbal
Prof. of Peadiatric Medicine Unit-II


Dr. Junaid Rashid
Prof. of Paediatric Medicine / Chairman
Department of Paediatric Medicine

The Children's Hospital & University of Child Health Sciences Lahore

Technical Evaluation

Package No. / Bid No. PC/2020/49/CH & UCHS	49
Name of the Equipment:	Extra Corporial Shock Wave Lithotripor (ESWL) Intergrated Type
Required Quantity:	1

Part - I

KNOCKED DOWN CRITERIA - (BID EVALUATION)

Evaluated by (TAC) Committee


Sr. #	Evaluation Parameters	Allmed Solutions	Hospital Supply Co
01	Complete Package / Tender (Mandatory)	Yes	Yes
02	Original Receipt of Tender (Mandatory)	Yes	Yes
03	Affidavit from Bidder	Yes	Yes
04	Bid Security CDR / Bank Guarantee (Mandatory)	Yes	Yes
05	Bid Validity (Mandatory)	Yes	Yes
06	Delivery Period (Mandatory)	Yes	Yes
Remarks:		Responsive	Responsive

PART - II

KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)

Evaluated by (TAC) Committee

Sr. #	Evaluation Parameters	Allmed Solutions	Hospital Supply Co
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	Yes	Yes
02	Technical & Engineering capability (As defined for the specific Bid in specifications)	Yes	Yes
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal	Yes	Yes
04	Bidder Past Performance for Quoted Product in Teaching Hospitals	Yes	No
05	Availability of relevant Tools and Testing / Calibration Equipment	Yes	Yes
06	Warranty	Yes	Yes
Remarks:		Responsive	Non-Responsive




 M. N. Khan
 12.3.2021


PART - III			
KNOCKED DOWN CRITERIA - (PRODCUT EVALUATION)			
Evaluated by (TAC) Committee			
49	Required Item Name	Allmed Solutions	Hospital Supply Co
	Extra Corporal Shock Wave Lithotripor (ESWL) Intergrated Type - 1		
	Brand		
Sr. #	Quoted Model	SLXFD01	Delta-III Pro (FPD)
01	Country of Manufacturer	Switzerland	Germany
02	Manufacturing site	Switzerland	Germany
03	Compliance with defined Quality Standards	Compliance	Compliance
04	Specifications Compliance Feature Wise	Compliance	Non-Compliance
05	Technical Eligibility of Product	Accepted	Non-Accepted
06	Technical Eligibility of Firm	Accepted	Accepted
Technical Status		Responsvie	Non-Responsvie

Enduser (Concerned Department/ Co-opted Member)



 Eng. Irfan Haider
 Bio Medical Engineer/
 Co-opted Member


 Dr. Ghulam Mujtaba Zafar
 Associate Prof. Paediatric Urology


Technical Advisory Committee (TAC)


 Eng. Shujaat Ali
 Bio Medical Engineer/
 Technical Officer


Muhammad Riaz Shahid
 Incharge Pay Roll Section



 Dr. Syed Ali Abu Zar
 AMS (Coordination)


Dr. Faeza Hasnain
 Chief Pharmacist


 Dr. Muhammad Zeeshan Khan
 Assistant Professor Paediatric Medicine


Dr. Mumtaz Hussain
 Assistant Professor Paediatric
 Orthopedic Surgery



 Dr. Muhammad Bilal Mirza
 Associate Prof. of Pead Surgery



 Dr. Zia ur Rahman
 Associate Prof. of Peads Neurology



 Dr. Nazish Saqlan
 Associate Prof. of Pathology
 Department

Dr. Sarfraz Ahmad
 Associate Prof. Peadiatric
 - PM&R


 Dr. Farah Haroon
 Assocaite Prof. of Neonatology


 Dr. Uzma Kazmi
 Professor of Paediatric Cardiology


 Dr. Syed Muhammad Javed Iqbal
 Prof. of Peadiatric Medicine Unit-II


 Dr. Junaid Rashid
 Prof. of Paediatric Medicine / Chairman
 Department of Paediatric Medicine

The Children's Hospital		
University of Child Health Sciences, Lahore		
Technical Evaluation		
Package No. / Bid No:	Bid No. P.C./2020/Sr. No. 50 /CH&UCHS	
Name of the Equipment:	Computer System + LED + UPS (With Complete Accessories)	
Required Quantity:	9	
Part - I		
KNOCKED DOWN CRITERIA - (BID EVALUATION)		
<i>Evaluated by Purchase Committee</i>		
Sr. #	Evaluation Parameters	PTCL
01	Complete Package / Tender (Mandatory)	Yes
02	Original Receipt of Tender (Mandatory)	Yes
03	Affidavit from Bidder	Yes
04	Bid Security / CDR (Mandatory)	Yes
05	Bid Validity (Mandatory)	Yes
06	Delivery Period (Mandatory)	Yes
Remarks:		Responsive

PART - II


KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)		
<i>Evaluated by Technical Committee</i>		
Sr. #	Evaluation Parameters	PTCL
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	N/A
02	Technical & Engineering capability (As defined for the specific Bid in specifications)	N/A
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal	N/A
04	Bidder Past Performance for Quoted Item	Yes
05	Availability of relevant Tools and Testing / Calibration Equipment	N/A
06	Warranty	Yes
Remarks:		Responsive


m. [Signature] P A S H [Signature] [Signature] [Signature] [Signature]

PART - III

KNOCKED DOWN CRITERIA - (PRODUCT EVALUATION)	
Evaluated by Technical Committee	
Required Item Name	
Brand	
Sr. #	Quoted Model
	PTCL
	Dell
	Optiplex Tower Plus 7010
01	Country of Manufacturer
02	Country of Origin of Product
03	Compliance with defined Quality Standards
04	Specifications Compliance Feature Wise
05	Technical Eligibility of Product
06	Technical Eligibility of Firm
Technical Status	
Responsive	


Enduser (Concerned Department/ Co-opted Member


Mazhar Iqbal Noor
Computer Programmer / IT
Incharge



Dr. Ghulam Mujtaba Zafar
Associate Prof. Paediatric
Urology

Dr. Imran Saeed
Prof. of Paediatric ENT


Technical Advisory Committee (TAC)


Engr. Shujaat Ali
Bio Medical Engineer/
Technical Officer


Muhammad Riaz Shahid
Incharge Pay Roll Section



Dr. Syed Ali Abu Zar
AMS (Coordination)


Dr. Faeza Hasnain
Chief Pharmacist


Dr. Muhammad Zeeshan Khan
Assistant Professor Paediatric
Medicine


Dr. Mumtaz Hussain
Assistant Professor Paediatric
Orthopedic Surgery



Dr. Muhammad Bilal Mirza
Associate Prof. of Paed Surgery


Dr. Zia ur Rahman
Associate Prof. of Paeds
Neurology



Dr. Nazish Saqlan
Associate Prof. of Pathology
Department

Dr. Sarfraz Ahmad
Associate Prof. Paediatric
PM&R


Dr. Farah Haroon
Associate Prof. of Neonatology


Dr. Uzma Kazmi
Professor of Paediatric Cardiology


Dr. Syed Muhammad Javed Iqbal
Prof. of Paediatric Medicine Unit-II


Dr. Junaid Rashid
Prof. of Paediatric Medicine / Chairman
Department of Paediatric Medicine

The Children's Hospital	
University of Child Health Sciences, Lahore	
Technical Evaluation	
Package No. / Bid No:	Bid No. P.C./2020/ 51 /CH&UCHS
Name of the Equipment:	LaserPrinter (Black & White)
Required Quantity:	5
Part - I	
KNOCKED DOWN CRITERIA - (BID EVALUATION)	
<i>Evaluated by Purchase Committee</i>	
Sr. #	Apex Traders
	Evaluation Parameters
01	Complete Package / Tender (Mandatory) Yes
02	Original Receipt of Tender (Mandatory) Yes
03	Affidavit from Bidder Yes
04	Bid Security / CDR (Mandatory) Yes
05	Bid Validity (Mandatory) Yes
06	Delivery Period (Mandatory) Yes
Remarks:	
Responsive	

PART - II

KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)	
<i>Evaluated by Technical Committee</i>	
Sr. #	Apex Traders
	Evaluation Parameters
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer N/A
02	Technical & Engineering capability (As defined for the specific Bid in specifications) N/A
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal N/A
04	Bidder Past Performance for Quoted Item Yes
05	Availability of relevant Tools and Testing / Calibration Equipment N/A
06	Warranty Yes
Remarks:	
Responsive	



 M.A. R. S. J.S. M. T. P. M. Khan
 12.3.2024

Package No. / Bid No: Bid No. P.C./2020/51 /CH&UCHS

PART - III

KNOCKED DOWN CRITERIA - (PRODCUT EVALUATION)	
Evaluated by Technical Committee	
Required Item Name	
Brand	
Sr. #	Quoted Model
01	Country of Manufacturer
02	Country of Origin of Product
03	Compliance with defined Quality Standards
04	Specifications Compliance Feature Wise
05	Technical Eligibility of Product
06	Technical Eligibility of Firm
Technical Status	

Apex Traders

HP

HP107

China

USA

Yes


Yes


Yes

Yes

Responsive


Enduser (Concerned Department/ Co-opted Member)


Mazhar Iqbal Noor
Computer Programmer / IT
Incharge



Dr. Ghulam Mujtaba Zafar
Associate Prof. Paediatric
Urology

Dr. Imran Saeed
Prof. of Paediatric ENT


Technical Advisory Committee (TAC)


Engr. Shujaat
Bio Medical Engineer/
Technical Officer


Muhammad Riaz Shahid
Incharge Pay Roll Section



Dr. Syed Ali Abu Zar
AMS (Coordination)


Dr. Faeza Hasnain
Chief Pharmacist


Dr. Muhammad Zeeshan Khan
Assistant Professor Paediatric
Medicine


Dr. Mumtaz Hussain
Assistant Professor Paediatric Orthopedic
Surgery



Dr. Muhammad Bilal Mirza
Associate Prof. of Pead
Surgery


Dr. Zia ur Rehman
Associate Prof. of Peads
Neurology



Dr. Nazish Saqlan
Associate Prof. of
Pathology Department

Dr. Sarfraz Ahmad
Associate Prof. Paediatric
PM&R


Dr. Farah Haroon
Associate Prof. of Neonatology


Dr. Uzma Kazmi
Professor of Paediatric Cardiology


Dr. Syed Muhammad Javed Iqbal
Prof. of Paediatric Medicine Unit-II


Dr. Junaid Rashid
Prof. of Paediatric Medicine / Chairman
Department of Paediatric Medicine

The Children's Hospital		
University of Child Health Sciences, Lahore		
Technical Evaluation		
Package No. / Bid No:	Bid No. P.C./2020/Sr. No. 52 /CH&UCHS	
Name of the Equipment:	Multimedia Projector	
Required Quantity:	2	
Part - I		
KNOCKED DOWN CRITERIA - (BID EVALUATION)		
<i>Evaluated by Purchase Committee</i>		
Sr. #	Evaluation Parameters	Apes Traders
01	Complete Package / Tender (Mandatory)	Yes
02	Original Receipt of Tender (Mandatory)	Yes
03	Affidavit from Bidder	Yes
04	Bid Security / CDR (Mandatory)	Yes
05	Bid Validity (Mandatory)	Yes
06	Delivery Period (Mandatory)	Yes
Remarks:		Responsive

PART - II

KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)		
<i>Evaluated by Technical Committee</i>		
Sr. #	Evaluation Parameters	Apes Traders
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	N/A
02	Technical & Engineering capability (As defined for the specific Bid in specifications)	N/A
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal	N/A
04	Bidder Past Performance for Quoted item	Yes
05	Availability of relevant Tools and Testing / Calibration Equipment	N/A
06	Warranty	Yes
Remarks:		Responsive

A series of handwritten signatures and initials in black ink, including a large signature on the left, several smaller initials in the middle, and a signature on the right with the date '12.5.2024' written below it.


Package No. / Bid No: Bld No. P.C./2020/52/CH&UCHS


PART - III

KNOCKED DOWN CRITERIA - (PRODUCT EVALUATION)

<i>Evaluation by Technical Committee</i>		Apex Traders
Required Item Name		
Brand		Accr
Sr. #	Quoted Model	X1226AWH
01	Country of Manufacturer	China
02	Country of Origin of Product	Taiwan
03	Compliance with defined Quality Standards	Yes
04	Specifications Compliance Feature Wise	Yes
05	Technical Eligibility of Product	Yes
06	Technical Eligibility of Firm	Yes
Technical Status		Responsive


Enduser (Concerned Department/ Co-opted Member)


Mazhar Iqbal Noor
Computer Programmer / IT
Incharge



Dr. Ghulam Mujtaba Zafar
Associate Prof. Paediatric
Urology

Dr. Imran Saeed
Prof. of Paediatric ENT


Technical Advisory Committee (TAC)


Engr. Shujaat Ali
Bio-Medical Engineer/
Technical Officer


Muhammad Riaz Shahid
Incharge Pay Roll Section



Dr. Syed Ali Abu Zar
AMS (Coordination)


Dr. Faeza Hasnain
Chief Pharmacist


Dr. Muhammad Zeeshan Khan
Assistant Professor Paediatric
Medicine


Dr. Mumtaz Hussain
Assistant Professor Paediatric
Orthopedic Surgery



Dr. Muhammad Bilal Mirza
Associate Prof. of Paed Surgery



Dr. Zia ur Rehman
Associate Prof. of Paed
Neurology


Dr. Naeem Saqlan
Associate Prof. of Pathology
Department

Dr. Sarfraz Ahmad
Associate Prof. Paediatric
PM&R


Dr. Farah Haroon
Associate Prof. of Neonatology


Dr. Uzma Kazmi
Professor of Paediatric Cardiology


Dr. Syed Muhammad Javed Iqbal
Prof. of Paediatric Medicine Unit-II


Dr. Junaid Rashid
Prof. of Paediatric Medicine / Chairman
Department of Paediatric Medicine

The Children's Hospital	
University of Child Health Sciences, Lahore	
Technical Evaluation	
Package No. / Bid No:	Bid No. P.C./2020/Sr. No. 53 /CH&UCHS
Name of the Equipment:	LaserPrinter with Scanner & Photocopier (Black & White)
Required Quantity:	4
Part - I	
KNOCKED DOWN CRITERIA - (BID EVALUATION)	
<i>Evaluated by Purchase Committee</i>	
Sr. #	Apex Traders
01	Complete Package / Tender (Mandatory)
02	Original Receipt of Tender (Mandatory)
03	Affidavit from Bidder
04	Bid Security / CDR (Mandatory)
05	Bid Validity (Mandatory)
06	Delivery Period (Mandatory)
Remarks:	
Responsive	

PART - II

KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)	
<i>Evaluated by Technical Committee</i>	
Sr. #	Apex Traders
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer
02	Technical & Engineering capability (As defined for the specific Bid in specifications)
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal
04	Bidder Past Performance for Quoted Item
05	Availability of relevant Tools and Testing / Calibration Equipment
06	Warranty
Remarks:	
Responsive	





 M. W. Q. A. S. J. T. H. M. Ahmed
 12-3-2024

PART - III

Evaluated by Technical Committee		Apex Traders
Required Item Name		
Brand		Pantum
Sr. #	Quoted Model	M6500
01	Country of Manufacturer	China
02	Country of Origin of Product	China
03	Compliance with defined Quality Standards	Yes
04	Specifications Compliance Feature Wise	Yes
05	Technical Eligibility of Product	Yes
06	Technical Eligibility of Firm	Yes
Technical Status		Responsive


Enduser (Concerned Department/ Co-opted Member)


Mazhar Iqbal Noor
Computer Programmer / IT Incharge



Dr. Ghulam Mujtaba Zafar
Associate Prof. Paediatric Urology

Dr. Imran Saeed
Prof. of Paediatric ENT


Technical Advisory Committee (TAC)


Engr. Shujaat
Bio Medical Engineer/
Technician Officer


Muhammad Riaz Shahid
Incharge Pay Roll Section



Dr. Syed Ali Abu Zar
AMS (Coordination)


Dr. Faeza Hasnain
Chief Pharmacist


Dr. Muhammad Zeeshan Khan
Assistant Professor Paediatric
Medicine


Dr. Mumtaz Hussain
Assistant Professor Paediatric
Orthopedic Surgery



Dr. Muhammad Bilal Mirza
Associate Prof. of Paed Surgery



Dr. Zia ur Rehman
Associate Prof. of Paeds Neurology



Dr. Nazish Saqlan
Associate Prof. of Pathology
Department

Dr. Sarfraz Ahmad
Associate Prof. Paediatric PM&R


Dr. Farah Haroon
Associate Prof. of Neonatology


Dr. Uzma Kazmi
Professor of Paediatric Cardiology


Dr. Syed Muhammad Javed Iqbal
Prof. of Paediatric Medicine Unit-II


Dr. Junaid Rashid
Prof. of Paediatric Medicine / Chairman
Department of Paediatric Medicine

The Children's Hospital

University of Child Health Sciences, Lahore

Technical Evaluation

Package No. / Bid No:	Bid No. P.C./2020/Sr. No. 54 /CH&UCHS		
Name of the Equipment:	Internet Networking		
Required Quantity:	Package (as per list)		

Part - I

KNOCKED DOWN CRITERIA - (BID EVALUATION)			
Evaluated by Purchase Committee			
Sr. #	Evaluation Parameters	PTCL	Apex Traders
01	Complete Package / Tender (Mandatory)	Yes	Yes
02	Original Receipt of Tender (Mandatory)	Yes	Yes
03	Affidavit from Bidder	Yes	Yes
04	Bid Security / CDR (Mandatory)	Yes	Yes
05	Bid Validity (Mandatory)	Yes	Yes
06	Delivery Period (Mandatory)	Yes	Yes
Remarks:		Responsive	Responsive

PART - II


KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)			
Evaluated by Technical Committee			
Sr. #	Evaluation Parameters	PTCL	Apex Traders
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer	N/A	N/A
02	Technical & Engineering capability (As defined for the specific Bid in specifications)	N/A	N/A
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal	N/A	N/A
04	Bidder Past Performance for Quoted item	Yes	Yes
05	Availability of relevant Tools and Testing / Calibration Equipment	N/A	N/A
06	Warranty	Yes	Yes
Remarks:		Responsive	Responsive


M. Zulqarnain
 12-3-2024

PART - III

KNOCKED DOWN CRITERIA - (PRODCUT EVALUATION)			
Evaluated by Technical Committee		PTCL	Apex Traders
Required Item Name		Internet Networking	Internet Networking
Brand		As per List	As per List
Sr. #	Quoted Model		
01	Country of Manufacturer	China	China
02	Country of Origin of Product	China	China
03	Compliance with defined Quality Standards	Yes	Yes
04	Specifications Compliance Feature Wise	Yes	Yes
05	Technical Eligibility of Product	Yes	Yes
06	Technical Eligibility of Firm	Yes	Yes
Technical Status		Responsive	Responsive


Enduser (Concerned Department/ Co-opted Member)


Mazhar Iqbal Noor
Computer Programmer / IT Incharge



Dr. Ghulam Mujtaba Zafar
Associate Prof. Paediatric
Urology

Dr. Imran Saeed
Prof. of Paediatric ENT

Technical Advisory Committee (TAC)


Engr. Shujaat Ali
Bio Medical Engineer/
Technical Officer


Muhammad Riaz Shahid
Incharge Pay Roll Section



Dr. Syed Ali Abu Zar
AMS (Coordination)


Dr. Faeza Hasnain
Chief Pharmacist


Dr. Muhammad Zeeshan Khan
Assistant Professor Paediatric Medicine


Dr. Mumtaz Hussain
Assistant Professor
Paediatric Orthopedic
Surgery



Dr. Muhammad Bilal Mirza
Associate Prof. of Pead Surgery


Dr. Zia ur Rehman
Associate Prof. of Peads
Neurology


Dr. Nazish Saqlan
Associate Prof. of Pathology
Department

Dr. Sarfraz Ahmad
Associate Prof. Paediatric PM&R


Dr. Farah Haroon
Assocate Prof. of Neonatology


Dr. Uzma Kazmi
Professor of Paediatric Cardiology


Dr. Syed Muhammad Javed Iqbal
Prof. of Paediatric Medicine Unit-II


Dr. Junaid Rashid
Prof. of Paediatric Medicine / Chairman
Department of Paediatric Medicine

The Children's Hospital
University of Child Health Sciences, Lahore

Technical Evaluation

Package No. / Bid No:	Bid No. P.C./2020/55/CH&UCHS
Name of the Equipment:	Digital Camera High Defination
Required Quantity:	1
Part - I	
KNOCKED DOWN CRITERIA - (BID EVALUATION)	
<i>Evaluated by Purchase Committee</i>	
Sr. #	Apex Traders
01	Complete Package / Tender (Mandatory) Yes
02	Original Receipt of Tender (Mandatory) Yes
03	Affidavit from Bidder Yes
04	Bid Security / CDR (Mandatory) Yes
05	Bid Validity (Mandatory) Yes
06	Delivery Period (Mandatory) Yes
Remarks:	
Responsive	

PART - II


KNOCKED DOWN CRITERIA - (VENDOR EVALUATION)	
<i>Evaluated by Technical Committee</i>	
Sr. #	Apex Traders
01	Exclusive Authorization / Sole Agent Certificate by the Manufacturer N/A
02	Technical & Engineering capability (As defined for the specific Bid in specifications) N/A
03	Up-to-Date After Sales Service Certificate from Manufacturer / Principal N/A
04	Bidder Past Performance for Quoted item Yes
05	Availability of relevant Tools and Testing / Calibration Equipment N/A
06	Warranty Yes
Remarks:	
Responsive	


Handwritten signatures and initials at the bottom of the page, including a date stamp: 12-3.

KNOCKED DOWN CRITERIA - (PRODCUT EVALUATION)

Evaluated by Technical Committee		Apex Traders
Required Item Name		
Brand		Nikon
Sr. #	Quoted Model	Z50
01	Country of Manufacturer	Japan
02	Country of Origin of Product	Japan
03	Compliance with defined Quality Standards	Yes
04	Specifications Compliance Feature Wise	Yes
05	Technical Eligibility of Product	Yes
06	Technical Eligibility of Firm	Yes
Technical Status		Responsive


Enduser (Concerned Department/ Co-opted Member)


Mazhar Iqbal Noor
Computer Programmer / IT
Incharge



Dr. Ghulam Mujtaba Zafar
Associate Prof. Paediatric
Urology

Dr. Imran Saeed
Prof. of Paediatric ENT


Technical Advisory Committee (TAC)


Engr. Ghujar Ali
Bio Medical Engineer/
Technical Officer


Muhammad Riaz Shahid
Incharge Pay Roll Section



Dr. Syed Ali Abu Zar
AMS (Coordination)


Dr. Faeza Hasnain
Chief Pharmacist


Dr. Muhammad Zeeshan Khan
Assistant Professor Paediatric
Medicine


Dr. Mumtaz Hussain
Assistant Professor Paediatric
Orthopedic Surgery



Dr. Muhammad Bilal Mirza
Associate Prof. of Pead Surgery


Dr. Zia ur Rehman
Associate Prof. of Peads
Neurology



Dr. Nazish Saqlan
Associate Prof. of Pathology
Department

Dr. Sarfraz Ahmad
Associate Prof. Paediatric PM&R


Dr. Farah Haroon
Assocaite Prof. of Neonatology


Dr. Uzma Kazmi
Professor of Paediatric Cardiology


Dr. Syed Muhammad Javed Iqbal
Prof. of Peadiatric Medicine Unit-II


Dr. Junaid Rashid
Prof. of Paediatric Medicine / Chairman
Department of Paediatric Medicine