



**The Children's Hospital
University of Child Health Sciences**

Ferozepur Road, Lahore Phone # (92) (42) 99230901-23 Fax # 99231560



No. P.C./

16670

/CH&UCHS

Dated 03-03-2023

To

✓
Mr. Abdul Wahab

I/C Programmer (I.T.)


Health Department, Civil Secretariat,
Lahore.

Subject: -

**REQUEST FOR UPLOADING THE MINUTES OF GRIEVANCE COMMITTEE MEETING
REGARDING THE PURCHASE OF FRAMEWORK CONTRACT FOR COCHLEAR
IMPLANT (HEARING DEVICES) ON HEALTH DEPARTMENT WEBSITE FOR THE
YEAR 2022-2023**

Please find enclosed herewith Minutes of the **Grievance Committee Meeting regarding The Purchase of Framework Contract for Cochlear Implant (Hearing Devices)** at The Children's Hospital, University of Child Health Sciences, Lahore for the year 2022-2023, need to be upload in the Health Department website.

Your positive response will be highly appreciated.


Prof. Dr. Muhammad Saleem
(FCPS, FACS (USA), MME-HPE)
Professor of Paediatric Surgery
Medical Director

MINUTES OF GRIEVANCES COMMITTEE MEETING HELD ON 28-02-2023 AT 12:00 NOON IN THE CONFERENCE ROOM AT THE CHILDREN'S HOSPITAL, UNIVERSITY OF CHILD HEALTH SCIENCES, LAHORE REGARDING THE FRAMEWORK CONTRACT FOR PURCHASE OF COCHLEAR IMPLANT (HEARING DEVICES) FOR THE YEAR 2022-2023

With reference to letter No. PC/14889-93/CH&ICH dated 23-02-2023, a Grievance Committee meeting was held on 28.02.2023 at 12:00 Noon in the conference room of the admin block of The Children's Hospital, Lahore. For addressal of the grievance of the firms submitted under bid reference "Framework contract for purchase of **Cochlear Implant (Hearing Devices)**" for the year **2022-2023** at The Children's Hospital, University of Medical Sciences, Lahore.

Following members and Firms Represent attended the meeting:-

Sr. No.	Name	Designation
01	Dr. Samina Zaman	Prof. of Histopathology / Chairperson
02	Dr. Muhammad Khalid Masood	Prof. of Paeds Medicine / Head of Department Medical Unit-III / Member
03	Dr. Edrees Anwar Sheikh	AMS (Admin) / Member
04	Dr. Zia Ur Rehman	Associate Prof. of Paeds Neurology / Member
05	Engr. Abdul Jalil	Biomedical Engineer Jinnah Hospital, Lahore / Member
06	Mr. Ehsan Danish	M/s MED-EL Pakistan.
07	Mr. Sarfaraz Ahmed	M/s Audio Medics.

Meeting was started with the recitation of Holy Quran and formal introduction of all participant to chaired Dr. Samina Zaman Prof. of Histopathology / Chairperson. It was briefed to the chair that two firms were participated in the said tender and both were declared responsive by the TAC committee. It was further briefed to committee that both firms submitted their grievance as per PPRA rule and meeting of grievances committee has been called to address the grievance of aggrieved firms.

1.	Framework Contract of Cochlear Implant (Hearing Devices) Reference Bid No. 3788/CH&UCHS Dated 19-1-2023 (For the year 2022-2023)	M/s Audio Medics V/S M/s Med-el Pakistan	That Committee should recheck the free sales certificate and reference documents that quoted product are being used in country of manufacturer (Austria)	The committee rechecked the bid submitted by the firm and found that free sale certificate of the quoted product /Model is attached furthermore the undertaking from manufacturer relating the use of quoted model in Austria and currently in production line is also available. It was further advised to M/s Med-el representative to present the reference of institutes /Surgeons who have been implants the quoted model in Austria. The firm submitted the reference as per directed. Decision: The Grievances Re-addressal Committee rejected the Grievance submitted by M/s Audio Medics and sustains the Decision of TAC and M/s Med-el Pakistan remained as responsive.
----	---	---	--	---



MINUTES OF GRIEVANCES COMMITTEE MEETING HELD ON 28-02-2023 AT 12:00 NOON IN THE CONFERENCE ROOM AT THE CHILDREN'S HOSPITAL, UNIVERSITY OF CHILD HEALTH SCIENCES, LAHORE REGARDING THE FRAMEWORK CONTRACT FOR PURCHASE OF COCHLEAR IMPLANT (HEARING DEVICES) FOR THE YEAR 2022-2023

2.		M/s Medel Pakistan V/S M/s Audio Medics	That the quoted model Ci522 by M/s Audio Medics is not MRI compatible with 1.5 T or 3 T without magnet Removal	<p>Representative of M/s Med-el Pakistan submitted the documents against the quoted model of M/s Audio medics i.e Cochlear Ci522 is not compatible with 1.5 T or 3 T MRI</p> <p>On which the Representative of M/s Audio Medic Submitted the FDA certificate issued to the manufacturer by authority in December 2022 In which MRI Kit has been approved by the FDA for the said implants and allowed the recipient to undergo MRI scan at 1.5 T with the magnetic removal and any surgical procedure.</p> <p>Decision:</p> <p>The Grievances Re-adderssal Committee rejected the Grievance submitted by M/s Med-el Pakistan and sustain the Decision of TAC and M/s Audio Medics remained as responsive</p>
----	--	---	--	---



Abdul Jalil
Biomedical Engineer Jinnah Hospital,
Lahore



Dr. Zia Ur Rehman
Associate Prof. of Peads Neurology



Dr. Edrees Anwar Sheikh
AMS (Admin)



Dr. Muhammad Khalid Masood
Prof. of Paeds Medicine / Head of
Department Medical Unit-III



Dr. Samina Zaman
Prof. of Histopathology

It is submitted for information & approval please



Corrigendum

MINUTES OF GRIEVANCES COMMITTEE MEETING HELD ON 28-02-2023 AT 12:00 NOON IN THE CONFERENCE ROOM AT THE CHILDREN'S HOSPITAL, UNIVERSITY OF CHILD HEALTH SCIENCES, LAHORE REGARDING THE FRAMEWORK CONTRACT FOR PURCHASE OF COCHLEAR IMPLANT (HEARING DEVICES) FOR THE YEAR 2022-2023

With reference to letter **No.PC/14889-93/CH&ICH** dated **23-02-2023**, a meeting was held on 28.02.2023 at 12:00 Noon. in the conference room of the admin block of The Children's Hospital, Lahore. For addresal of the grievance of the firms submitted under bid reference "Framework contract for purchase of **Cochlear Implant (Hearing Devices)**" for the year **2022-2023** at The Children's Hospital, University of Medical Sciences, Lahore.

Published	Read As
Representative of M/s Med-el Pakistan submitted the documents against the quoted model of M/s Audio medics i.e Cochlear Ci522 is not compatible with 1.5 T or 3 T MRI On which the Representative of M/s Audio Medic Submitted the FDA certificate issued to the manufacturer by authority in December 2022 In which MRI Kit has been approved by the FDA for the said implants and allowed the recipient to undergo MRI scan at 1.5 T with the magnetic removal and any surgical procedure.	Representative of M/s Med-el Pakistan submitted the documents against the quoted model of M/s Audio medics i.e Cochlear Ci522 is not compatible with 1.5 T or 3 T MRI On which the Representative of M/s Audio Medic Submitted the FDA certificate issued to the manufacturer by authority in December 2022 In which MRI Kit has been approved by the FDA for the said implants and allowed the recipient to undergo MRI scan at 1.5 T with out the magnetic removal or any surgical procedure.
There was a typographic mistake word out was omitted. The over all decision remain same as pervious both firms are Responsive .	


Abdul Jalil


Biomedical Engineer Jinnah
Hospital, Lahore




Dr. Muhammad Khalid Masood
Prof. of Paeds Medicine / Head
of Department Medical Unit-III



Dr. Zia Ur Rehman
Associate Prof. of Paeds
Neurology



Dr. Samina Zaman
Prof. of Histopathology


Dr. Edrees Anwar Sheikh
AMS (Admin)