



**The Children's Hospital,
University of Child Health Sciences,**
Ferozpur Road, Lahore Phone # (92) (42) 99230901-23 Fax # 99231560



No. P.C./

SS111

/CH&UCHS

Dated 07-10- 2022

To


Mr. Abdul Wahab
I/C Programmer (I.T.)
Health Department, Civil Secretariat,
Lahore.

Subject: -

**REQUEST FOR UPLOADING THE MINUTES OF PRE-BID MEETING
REGARDING THE PURCHASE OF DENTAL
EQUIPMENT/MACHINERY FOR THE YEAR 2022-2023.ON HEALTH
DEPARTMENT WEBSITE**

Please find enclosed herewith Minutes of the Pre-Bid Meeting regarding the Purchase of Dental Equipment/Machinery in The Children's Hospital, University of Child Health Sciences, Lahore for the year 2022-2023, need to be upload in the Health Department website.

Your positive response will be highly appreciated.


Prof. Dr. Muhammad Saleem
(FICS, FACS (USA), MME-HPE)
Professor of Paediatric Surgery
Medical Director

MINUTES OF THE PRE-BID MEETING HELD ON 29-09-2022 AT 11:30AM IN THE CONFERENCE ROOM OF THE ADMIN BLOCK REGARDING THE PURCHASE OF DENTAL EQUIPMENT / MACHINERY FOR THE YEAR 2022-23

Bid regarding the purchase of Dental Equipments/ Machinery was advertised vide bid notice No. PC/50468/1-5/CH&UCHS dated 13-09-2022. Tender published in newspaper on 16-09-2022 under IPL No. 9287 as well as on PPRA website on 19-09-2022 and ready for sale. As per advertised tender Pre-Bid Meeting held on 29-09-2022 at 11:30am vide letter No. PC/52899 -52903/CH&UCHS dated 26-09-2022 (As per schedule already given in the advertisement).

Meeting started with the name of Allah, and following companies representatives attended the meeting

1. M/s Total technologies
2. M/s Forecare
3. M/s Advance System
4. M/s Glow Pak International
5. M/s Combined Engineers

1. CBCT (3D DIGITAL DENTAL PANORAMIC AND CEPHALOMETRIC X-RAY UNIT)

Approved PVMS of all medical equipment were advertised. Fair chance was provided to all attendee regarding clarification / quires if any. All quires raised by firm's representatives were addressed on spot by the End-users and biomedical engineers. Few changes were made for further clarifications.

Advertised Specifications	Clarification/ Justification
2D Sensor: CCD	2D Sensor: CCD / DC / CMOSE
Software : Implant planning, Reporting, Crown and implant library, Nerve locator	Software: All essential software required to Dental, Maxillofacial, Orthodontics and ENT e.g. Ceph, OPG, facial reconstruction, Prosthesis planning Middle Ear, Implant planning, reporting, nerve locator e.t.c.
Pure Sine wave Online UPS to take load of X-ray, PC and DICOM Printer with 10-15 backup time	Pure Sine wave Online UPS to take load of X-ray, PC and DICOM Printer with 10-15 <i>minutes</i> backup time
UPS	UPS should be imported but may be supplied locally
Printer Specifications	Printer should be dual tray and support all standard sizes and may be supplied locally

2. DENTAL X-RAY MACHINE

Power on light signal, x-ray emission light signal Pure Sine wave Online UPS to take load of X-ray, PC and DICOM Printer with 10-15 backup time	Power on light signal, X-Ray emission light signal Pure Sine wave Online UPS to take load of X-ray, PC and DICOM Printer with 10-15 Backup time
Focal point 0.4mm-0.0mm or better with filtration system	Focal point 0.4mm- 0.0mm or better with filtration system

3. DIGITAL FILM PRINTER

Specifications	Printer should be dual tray and support all standard sizes and may be supplied locally. Both printers Laser Dicom, Inkjet.
----------------	--

MINUTES OF THE PRE-BID MEETING HELD ON 29-09-2022 AT 11:30AM IN THE CONFERENCE ROOM OF THE ADMIN BLOCK REGARDING THE PURCHASE OF DENTAL EQUIPMENT / MACHINERY FOR THE YEAR 2022-23

4. DIGITAL IMAGING SYSTEM (PHOTOSTIMULABLE PHOSPHOR)

Plate Sizes and Quantities	Size	Quantity
	0	4
	1	4
	2	8
	3	4
	4	4
UPS	Pure Sine wave Online UPS to take load of X-ray, PC and DICOM Printer with 10-15 minutes backup time. UPS should be imported but may be supplied locally	
Printer Specifications	Ink Jet printer with 1000 Plastic Sheets, should support all standard sizes and may be supplied locally	

5. DIGITAL IMAGING SYSTEM (INTRA ORAL SENSOR)

UPS	Pure Sine wave Online UPS to take load of X-ray, PC and DICOM Printer with 10-15 minutes backup time. UPS should be imported but may be supplied locally
Printer Specifications	Ink Jet printer with Plastic Sheets, should support all standard sizes and may be supplied locally


CIVIL WORKS


Complete renovation of machine room and Patients waiting area having dimensions (20x20) & (20x40) respectively as per requirement / standards including safety standards.


Note: All firms representative were duly agreed on the clarifications/ changes made in the specifications

Pre Bid Committee Members/ End-users


Engr. Man Haider
Bio Medical Engineer/ Purchase
Officer


Engr. Shujaat Ali
Bio Medical Engineer/ Technical
Officer


Dr. Aysha Ashraf
Associate Prof. of Orthodontics


Dr. Saeed Ahmed
Associate Prof. of Maxillo
Surgery / Head of Dental
Department


Dr. Edrees Anywar
AMS (Admin)

It is submitted for information & approval please.

Medical Director