



**MINUTES OF THE PRE BID MEETING FOR THE PROCUREMENT OF FRAMEWORK CONTRACT OF COCHLEAR IMPLANT (HEARING DEVICES) UNDER CHIEF MINISTER SPECIAL INITIATIVE FOR TRANSPLANT PROGRAM (CMSITP) FY 2025-26 HELD ON 05-08-2025 AT 11:30 A.M. IN THE SYNDICATE ROOM OF THE ADMIN BLOCK THE CHILDREN'S HOSPITAL, UNIVERSITY OF CHILD HEALTH SCIENCES, LAHORE.**

1. Bid regarding purchase of Framework Contract of Cochlear Implant (Hearing Devices) advertised vide bid notice No/ COCHLEAR-25/P-250897653/CH&UCHS dated 01-08-2025 on FOR basis with quantity **350** units as per approved PVMS specifications dully vetted by end users as well as biomedical engineer.
2. Tender Published on E-Pads as well as on PPRA website dated 01-08-2025.
3. A pre-bid meeting held on 05-08-2025 at 11:30 am (As per schedule already given in the Bid invitation)
4. Following Firms participated in the meeting.
  - a. M/s Med-EL Pakistan
  - b. M/s Audio Medics
  - c. M/s PK Medi Engineering
5. The following observations/clarifications rose by the representatives of the firms regarding advertised specifications and Bidding Documents.

Sr. #	Observation raised by the Firm	Clarification/Decision
1.	All Firms Suggested that the warranty period mentioned in the specification was 5 years it should be 3 years.	Warranty period mentioned in specification will remain same i.e. Comprehensive warranty of external unit including accessories 5 years. Comprehensive warranty of internal unit of 10 years.
2.	Firm suggested that in cochlear Implant specification Internal Unit point no 08 i.e. MRI compatible 1.5T or better should also be removable.	<b>Not Accepted.</b>
3.	Cochlear Implant specification Accessories One Laptop for mapping as per manufacturer recommendation should be place in Audiology Department.	Two Laptops latest and well-known model for mapping as per manufacturer recommendation one for ENT department & one for Audiology Department should be placed.
4.	Cochlear Implant specification Accessories.	Surgical kit for Cochlear Implant (Qty 05) also included in specifications (Accessories).
5.	M/s PK Medi Engineering requested to consider international references for Past performance.	<b>Not Accepted.</b>
6.	Cochlear Implant specification (Training and Attachments) One international attachment of Doctors. Firms requested that mentioned number of persons for training and attachment	Cochlear Implant specification (Training and Attachments). Two doctors (one from ENT and one from Audiology) are recommended for training and attachment.

  
Engr. Muzamil Janjua  
Bio Medical Engineer/Member

  
Dr. Warda Aziz  
Audiologist /Member

  
Dr. Imran Saeed  
Prof. Paediatric ENT  
/Chairman

It is submitted for information and approval please.

  
Medical Director